

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## HIV

Function of the test	Technology	Performance evaluation protocols available for the following specimen types <sup>b</sup>
Aid to diagnosis	Rapid diagnostic tests <sup>c</sup>	Serum/plasma, capillary whole blood, oral fluid, urine
	Enzyme immunoassays	Serum/plasma
	Nucleic acid tests (qualitative)	Whole blood, Dried blood spot
Monitoring	Flow cytometer for enumeration of lymphocyte subset including CD4+ T cells, or a technology that can be used at or near the patient (quantitative or semi-quantitative)	Whole blood
	Nucleic acid tests for measuring viral load (quantitative or semi-quantitative)	Plasma, Dried blood spot, Plasma separation card

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## Hepatitis C virus

<b>Function of the test</b>	<b>Technology</b>	<b>Performance evaluation protocols available for the following specimen types<sup>b</sup></b>
Aid to diagnosis	Rapid diagnostic tests	Serum/plasma
	Enzyme immunoassays	Serum/plasma
Aid to diagnosis and monitoring	Nucleic acid tests (quantitative or qualitative)	Plasma, Dried blood spot

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## Hepatitis B virus

<b>Function of the test</b>	<b>Technology</b>	<b>Performance evaluation protocols available for the following specimen types<sup>b</sup></b>
Aid to diagnosis and monitoring	Rapid diagnostic tests	Serum/plasma
	Enzyme immunoassays	Serum/plasma
Monitoring	Quantitative nucleic acid tests	Plasma

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## Malaria parasites

Function of the test	Technology	Performance evaluation protocols available for the following specimen types <sup>b</sup>
Aid to diagnosis and monitoring	Rapid diagnostic tests	Serum/plasma

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## Human papilloma virus

Function of the test	Technology	Performance evaluation protocols available for the following specimen types <sup>b</sup>
Screening	Nucleic acid tests (DNA or mRNA)	Cervical swab

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## Glucose-6-phosphate dehydrogenase (G6PD) enzyme

Function of the test	Technology	Performance evaluation protocols available for the following specimen types <sup>b</sup>
Enzyme deficiency detection	Technologies/formats to be used at or near the patient (quantitative or semi-quantitative or qualitative)	Whole blood

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## Toxigenic *Vibrio cholerae*

<b>Function of the test</b>	<b>Technology</b>	<b>Performance evaluation protocols available for the following specimen types<sup>b</sup></b>
Surveillance	Rapid diagnostic tests	Stool

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## ***Treponema pallidum (Syphilis)***

<b>Function of the test</b>	<b>Technology</b>	<b>Performance evaluation protocols available for the following specimen types<sup>b</sup></b>
Screening and aid to diagnosis	Rapid diagnostic tests <sup>c</sup>	Serum/plasma

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>

## Mycobacterium tuberculosis complex and resistance to first and/or second line anti-tuberculosis drugs

Function of the test	Technology	Performance evaluation protocols available for the following specimen types <sup>b</sup>
Aid to diagnosis	Qualitative nucleic acid tests	Sputum

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing

**Table 1: Products currently eligible for WHO's performance evaluation as a prerequisite to apply for WHO's prequalification assessment<sup>a</sup>**

## SARS-CoV-2

<b>Function of the test</b>	<b>Technology</b>	<b>Performance evaluation protocols available for the following specimen types<sup>b</sup></b>
Aid to diagnosis	Rapid diagnostic tests	Nasopharyngeal swabs, nasal swabs
	Qualitative nucleic acid tests	Analytical evaluation only, not dependent on specimen type

<sup>a</sup> For multiplex tests, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>b</sup> If no claimed specimen types are included in the table above, the manufacturer is requested to contact WHO in advance of submitting an EOI.

<sup>c</sup> Including rapid diagnostic tests intended for self-testing