

WHO Emergency Use Assessment SARS-CoV-2 IVDs PUBLIC REPORT

Product: STANDARD Q COVID-19 Ag Test

Manufacturer: SD Biosensor, Inc.

EUL Number: EUL 0563-117-00

Outcome: Accepted

The EUL process is intended to expedite the availability of in vitro diagnostics needed in public health emergency situations and to assist interested UN procurement agencies and the Member States in determining the acceptability of using specific products in the context of a Public Health Emergency of International Concern (PHEIC), based on an essential set of available quality, safety and performance data. The EUL procedure includes the following:

- Quality Management Systems Review and Plan for Post-Market Surveillance: desk-top review of the manufacturer's Quality Management System documentation and specific manufacturing documents;
- Product Dossier Review: assessment of the documentary evidence of safety and performance.

STANDARD Q COVID-19 Ag Test, product codes 09COV30D, 09COV31D, 09COV32D, 09COV33D and 10COVC11 (COVID-19 Ag Control swab), CE-mark regulatory version, manufactured by SD Biosensor, Inc., C 4th and 5th, 16 Deogyeong-daero, 1556 beon-gil Suwon-si, Gyeonggi-do, 16690, Republic of Korea, was listed on 22 September 2020.

Report amendments and product changes

This public report has since been amended. Amendments may have arisen because of changes to the EUL product for which WHO has been notified and has undertaken a review. Amendments to the report are summarized in the table below.

Version	Summary of amendment	Date of report amendment
2.0	Addition of a warning that STANDARD COVID-19 Ag Control (product code:10COVC10) was not assessed with STANDARD Q COVID-19 Ag Test and is not part of EUL.	6 November 2020
3.0	A warning correction was added to the version to read as " <i>STANDARD COVID-19 Ag Control (product code:10COVC10) was assessed and found unacceptable.</i> "	18 November 2020
4.0	Change the outside packaging box in the public report to align with the approved IFU.	15 January 2021
5.0	1. Changed the shape of the plastic container cover for the nozzle cap (Blister - flat).	29 June 2022

	<ol style="list-style-type: none"> 2. Addition of a new buffer tube rack. 3. Addition of a new manufacturing site. 4. Addition of a new warehousing site. 5. Addition of new suppliers for components. 6. Change the Site plan of BioNote. 7. Introduction of new quality control swabs. 8. Addition of a specimen type (Nasal swab). 9. Change of the product IFU and addition of new IFUs. 10. Change of the product package and addition of new packages. 	
6.0	<ol style="list-style-type: none"> 1. Change of Labelling of Nasopharyngeal swab. 2. Change of Labelling of Nasal Swab. 3. Change of Labelling of Packing materials (Nozzle cap and extraction buffer pouch). 	20 June 2024

Intended use

According to the claim of intended use from SD Biosensor, Inc., "*STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific antigens of SARS-CoV-2 present in human nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.*"

Specimen types that were validated

Nasopharyngeal and nasal swab specimens (please pay attention to the product codes).

Test kit contents:

Component	25 tests (product code 09COV30D)	25 tests (product code 09COV31D)	25 tests (product code 09COV32D)	25 tests (product code 09COV33D)	10 tests/kit COVID-19 Ag Control swab (product code 10COVC11)
Test device (individually in a foil pouch with desiccant)	25	25	25	25	N/A
Extraction buffer tube	25	25	25	25	N/A
Nozzle cap	25	25	25	25	N/A
Sterile swab	25 (Nasopharyngeal)	25 (Nasal)	25 (Nasopharyngeal)	25 (Nasal)	N/A
Negative and Positive control swabs	N/A	N/A	1 of each	1 of each	10 of each
Instructions for use	1	1	1	1	1

Items required but not provided

- Personal Protective Equipment per local recommendations (i.e. gown/lab coat, face mask, face shield/eye goggles and gloves)
- Timer
- Biohazard container
- STANDARD F Analyzer

Storage

2-30°C.

Shelf-life upon manufacture

24 months for STANDARD Q COVID-19 Ag Test (real-time stability studies are ongoing).

30 months for STANDARD COVID-19 Ag Control Swab.

Warnings/limitations:

STANDARD COVID-19 Ag Control (product code:10COVC10) was assessed and found unacceptable.

Product dossier assessment

SD Biosensor, Inc. submitted a product dossier for the STANDARD Q COVID-19 Ag Test for detecting severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) as per the "*Instructions for Submission Requirements: In vitro diagnostics (IVDs) Detecting SARS-CoV-2 Nucleic Acid and rapid diagnostics tests detecting SARS-CoV-2 antigens (PQDx_0347 version 4)*". The information (data and documentation) submitted in the product dossier was reviewed by WHO staff and external technical experts (assessors) appointed by WHO.

Post listing commitment for EUL

As a commitment to listing, the manufacturer is required to;

1. Provide the real-time stability studies report by 31 July 2022. The commitment is under review.
2. Submit the final transport and real-time stability studies report of the new quality control swab as part of the next EUL renewal application.
3. Submit nasal swab storage stability studies at -20 °C for 4 months and -70 °C for 6 months as part of the next EUL renewal application.
4. Submit the results of the accelerated stability testing up to week 20 as part of the next EUL renewal application.
5. Amend the IFU's font size to make it more user-friendly in the following IFU review. Submit the amended version of the IFU as part of the next EUL renewal application.

The risk-benefit assessment is acceptable.

Quality Management Systems Review

To establish eligibility for WHO procurement, SD Biosensor, Inc. was asked to provide up-to-date information about the status of its quality management system.

Based on the review of the submitted quality management system documentation by WHO staff, it was established that sufficient information was provided by SD Biosensor, Inc. to fulfil the requirements described in the "*Instructions for Submission Requirements: In vitro diagnostics (IVDs) Detecting SARS-CoV-2 Nucleic Acid and rapid diagnostics tests detecting SARS-CoV-2 antigens (PQDx_347 version 4)*".

Quality management documentation assessment is acceptable.

Plan for Post-Market Surveillance

Post-market surveillance, including monitoring all customer feedback, detecting and acting on adverse events, product problems, non-conforming goods and processes is a critical component of minimizing potential harm of an IVD listed for emergency use.

The following post-EUL activities are required to maintain the EUL listing status:

1. Notification to WHO of any planned changes to an EUL product, in accordance with "*WHO procedure for changes to a WHO prequalified in vitro diagnostic*" (document number *PQDx_121*); and
2. Post-market surveillance activities, in accordance with "*Guidance for post-market surveillance and market surveillance of medical devices, including in vitro diagnostics*" (ISBN 978-92-4-001531-9).

SD Biosensor, Inc. is also required to submit an annual report that details sales data and all categories of complaints in a summarized form. There are certain categories of complaints and changes to the product that must be notified immediately to WHO, as per the above-mentioned documents.

The manufacturer has committed to ensuring that post-emergency use listing safety, quality and performance monitoring activities are in place in accordance with WHO guidance "*WHO guidance on post-market surveillance of in vitro diagnostics*".¹

Scope and duration of procurement eligibility

STANDARD Q COVID-19 Ag Test, product codes 09COV30D, 09COV31D, 09COV32D, 09COV33D and 10COVC11 (COVID-19 Ag Control swab), manufactured by SD Biosensor, Inc. is considered to be eligible for WHO procurement until further notice. The assay may be used to detect the Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2) antigens. This listing does not infer that the product meets WHO prequalification requirements and does not mean that the product is listed as WHO-prequalified.

As part of the ongoing requirements for listing as eligible for WHO procurement, SD Biosensor, Inc. must engage in post-market surveillance activities to ensure that the product continues to meet safety, quality and performance requirements. SD Biosensor, Inc. is required to notify WHO of any complaints, including adverse events related to the use of the product, within 7 days and any changes to the product.

WHO reserves the right to rescind eligibility for WHO procurement if additional information on the safety, quality, and performance during post-market surveillance activities and if new data becomes available to WHO that changes the risk-benefit balance.

¹ Available on the web page

<https://www.who.int/publications/i/item/guidance-for-post-market-surveillance-and-market-surveillance-of-medical-devices-including-in-vitro-diagnostics>

Labelling

1.0 Labels

2.0 Instructions for Use (IFU)

1.0 Product labels

1.1 Outside box labels

1.1.1 Product code 09COV30D

전면



1.1.2 Product code 09COV31D

STANDARD Q
COVID-19 Ag (Nasal) 25T



1.1.3 Product code 09COV32D




1.1.4 Product code 09COV33D

STANDARD Q
COVID-19 Ag (Nasal) 25T




1.1.5 Product code 10COVC11 (COVID-19 Ag Control swab)

STANDARD
COVID-19 Ag Control swab 10T




1.2 Device Pouch label




COVID-19 Ag Test

(2열 포장기 roll type)_120mm x 70mm




1.3 Nasopharyngeal swab labels




20mm



1.4 Nasal swab labels





189MM

175MM

6mm

A blue triangle pointing upwards, indicating where to peel off the label.



Specimen Collection Swab



Sterile only if peel-pouch is intact.
This product is a disposable medical device

ss: Room 201 of Building 14th and Building 17th, Hengyi
Yuanhu Road, Zhangbei Industrial Park, Longcheng Street,
ang district, Shenzhen, Guangdong, China
medicoswah.com

0413 FDA Registered
Technology Co., Ltd

EC

W
Er
B

ellkang Ltd(www.CE-marking.eu)
Enterprise Hub,NW Business Complex,1
eraghmore Rd.Derry.BT488SE.N.Ireland.UK

Product code: MFS-97000KC

YYYY.MM.DD

1000x MM RE

FFFF.MM.DD

LOT xoooooo

6mm

7.6MM

30MM

6mm



OPEN

Disposable sampling swab

EC REP Luxus Lebenswelt GmbH
Kochstr. 1, 47877, Willich, Germany

CE 0197

Type: G-014
STERILE EO
FDA Registered



1. Please read the IFU carefully before use.
2. The product is sterile.
3. Don't use if package is damaged.

190mm



Jiangsu HanHeng Medical Technology Co., Ltd.
16-E4, #1 North Qingyang Road, Tanning District, Changzhou,
213017 Jiangsu P.R. China

EXECUTOR: EXTREME MACHINERY MANUFACTURE IMP&EXP CO., LIMITED
Flat A, Floor 15, Manly Commercial Building, 16 Soy Street,
Mong Kok, Kowloon, Hong Kong



LOT EXP

20mm

(2) (3) (4) (5) (6) (7) (8) (9) (10) (11) (12)
(13) (14) (15) (16) (17) (18) (19) (20) (21) (22) (23) (24)




1.5 Controls pouch label

Printing part
(black)

STANDARD COVID-19 Ag Negative Control swab


195mm x 50



STANDARD COVID-19 Ag Positive Control swab

195mm x 50

 Printing part
(black)



1.6 Extraction tube label

Extraction Buffer Tube

STANDARD Q COVID-19 Ag Test

LOT No. : XXXXXXXXXX
MFG DATE : YYYY.MM.DD.
EXP DATE : YYYY.MM.DD.
Quantity : 25pcs

SD BIOSENSOR



ML27RT1ENR2
Issue date: 2021.01

1.7 Nozzle label

80mm x 12mm

Nozzle cap

Code

MSH10-2

Qty 25 pcs

LOT

XXXXXX-XXXXX



YYYY.MM.DD

SD BIOSENSOR



L47RUE092

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

0

1

2.0 Instructions for use²


² English versions of the IFU were the ones that the WHO assessed. It is the responsibility of the manufacturer to ensure correct translation into other languages.

REF Q-NCOV-01G
Cat No. 09COV30D**COVID-19 Ag Test**

STANDARD™ Q COVID-19 Ag Test

PLEASE READ INSTRUCTIONS CAREFULLY BEFORE YOU PERFORM THE TEST


KIT CONTENTS



PREPARATION AND TEST PROCEDURE

■ PREPARATION


- Carefully read instructions for using the STANDARD Q COVID-19 Ag Test.
- Check the expiry date at the back of the foil pouch. Do not use the kit, if expiry date has passed.
- Check the test device and the desiccant pack in the foil pouch.



■ COLLECTION OF SPECIMEN

[Nasopharyngeal swab]

- To collect a nasopharyngeal swab specimen, insert a sterile swab into the nose of the patient, reaching the surface of the posterior nasopharynx.
- Using gentle rotation, push the swab until resistance is met at the level of the turbinates.
- Rotate the swab 3-4 times against the surface of the nasopharynx.
- Remove the swab from the nostril carefully.
- Insert the swab into an extraction buffer tube. While squeezing the buffer tube, stir the swab more than 5 times.



6. Remove the swab while squeezing the sides of the tube to extract the liquid from the swab.

7. Press the nozzle cap tightly onto the tube.

8. Specimen should be tested as soon as possible after collection.

9. Specimens may be stored at room temperature (15–25°C) for up to 1 hours or at 2–8°C/36–46°F for up to 4 hours prior to testing.

- If the specimen storage condition is out of instructions as below, do not use.
- The Nasopharyngeal swab is stored in extraction buffer for more than 4 hours at 55°C or 1 hour at 205°C.
 - Freezing the Nasopharyngeal swab specimen stored in extraction buffer. Freezing/thawing the specimen stored in UTM exceed 3 cycles.
 - The Nasopharyngeal swab is stored in UTM for more than 12 hours at 55°C or 8 hours at 205°C.

[Specimens in transport media]

- Using a micropipette, collect the 350µl of specimen from the collection cup or VTM. Mix the specimen with an extraction buffer.
- Press the nozzle cap tightly onto the tube.



Minimal dilution of the specimen is recommended, as dilution may result in decreased test sensitivity.

■ ANALYSIS OF SPECIMEN

- Apply 3 drops of extracted specimen to the specimen well of the test device.
- Read the test result in 15-30 minutes.

SPECIMEN COLLECTION AND PREPARATION

■ Transport medium

Recommended Storage Condition		
Virus Transport Medium(VTM)	2°C to 8°C	25°C
Copan UTM® Universal Transport Media	12 hours	8 hours
BD® Universal Viral Transport	12 hours	8 hours
STANDARD™ Transport Medium	12 hours	8 hours

When using viral transport medium (VTM), it is important to ensure that the VTM containing the specimen is warmed to room temperature. Cold specimens will not flow correctly and can lead to erroneous or invalid results. Several minutes will be required to bring a cold specimen to room temperature.

PERFORMANCE CHARACTERISTICS

■ Clinical evaluation

The prospective diagnostic evaluation of STANDARD Q COVID-19 Ag Test with a total number of enrolled individuals of 1659 was conducted by FIND with collaborators in Germany and Brazil.
A total of 153 positive specimens from Germany and Brazil were tested using the STANDARD Q COVID-19 Ag Test. These specimens consisted of nasopharyngeal swabs from symptomatic patients. The specificity of STANDARD Q COVID-19 Ag Test was tested using 1506 negative specimens. The sensitivity and specificity of the STANDARD Q COVID-19 Ag Test was compared to the specific RT-PCR method. The pooled sensitivity was 84.97% (130/153, 95% CI 78.36–90.23%) and the pooled specificity was 98.94% (1490/1506, 95% CI 98.88%–99.39%). Performance data was calculated from a study of patients within 24 days of onset of symptoms.

Table 1. STANDARD Q COVID-19 Ag Test result by FIND.


Country	Brazil	Germany	Overall
Sensitivity (Ct ≤ 25)	95.92% (47/49, 95% CI 86.02–99.50%)	100% (21/21, 95% CI 83.89–100%)	97.14% (68/70, 95% CI 90.06–99.65%)
Sensitivity (Ct ≤ 33)	91.92% (91/99, 95% CI 86.70–96.45%)	87.80% (36/41, 95% CI 73.80–95.92%)	90.71% (127/140, 95% CI 84.64–94.96%)
Sensitivity (if ≤ from the symptom onset days ≤ 3)	95% (19/20, 95% CI 75.13–99.87%)	85.71% (18/21, 95% CI 63.66–96.59%)	90.24% (37/41, 95% CI 76.87–97.28%)
Sensitivity (from the symptom onset days ≤ 7)	90.72% (88/97, 95% CI 83.12–95.67%)	80% (28/35, 95% CI 63.06–91.56%)	87.88% (116/132, 95% CI 81.06–92.91%)
Clinical Sensitivity	88.68% (94/106, 95% CI 81.06–94.01%)	76.60% (36/47, 95% CI 61.97–87.70%)	84.97% (130/153, 95% CI 78.36–90.23%)
Clinical Specificity	97.62% (287/294, 95% CI 95.16–99.04%)	99.26% (1203/1212, 95% CI 98.60–99.66%)	98.94% (1490/1506, 95% CI 98.28–99.39%)

ANALYTICAL PERFORMANCE

1. Limit of Detection (LoD): The SARS-CoV-2 positive specimen was prepared by spiking Inactivated SARS-CoV-2 (2019-nCoV) NCCP 43326/2020/Korea strain to SARS-CoV-2 negative nasopharyngeal swab confirmed with PCR. LoD is determined as 3.12×10^{12} TCID₅₀/ml for direct Nasopharyngeal swab, 5 $\times 10^{12}$ TCID₅₀/ml for Nasopharyngeal swab stored in VTM by testing serially diluted the mock positive specimen.

2. Cross-Reactivity & Microbial interference: There was no cross-reaction and interference with the potential cross-reacting microorganisms listed below except SARS-CoV.

Potential cross reacting substance	Strain	Concentration of potentially cross reacting substance
SARS-coronavirus	Urban	3.5 µg/ml
MERS-Coronavirus	Florida/USA-2_Saudi Arabia_2014	4 $\times 10^4$ TCID ₅₀ /ml
	229E	1 $\times 10^4$ TCID ₅₀ /ml
	OC43	1 $\times 10^4$ TCID ₅₀ /ml
	NL63	1 $\times 10^4$ TCID ₅₀ /ml



15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

Not valid
After 30 mins.
No test result

15-30 mins

EN


REF Q-NCOV-01G

Cat No. 09COV31D


COVID-19 Ag Test

STANDARD™ Q COVID-19 Ag Test

PLEASE READ INSTRUCTIONS CAREFULLY BEFORE YOU PERFORM THE TEST

KIT CONTENTS**SPECIMEN COLLECTION AND PREPARATION****Specimen preparation**

[Nasal swab]




1. Tie patient's head back slightly.


2. While rotating the swab, insert swab less than one inch (about 2cm) into nostril until resistance is met at turbinate.

3. Rotate the swab 4 times against nasal wall.

4. Repeat in other nostril using the same swab.




Only one swab is used per test. Rub both nostrils with the same swab. Using proper pressure, rub the swab against all of the inside nasal walls of your nostril. Make at least 4 big circles along the nasal wall. Do not just spin the swab.



- 5. Insert the swab into an extraction buffer tube. While squeezing the buffer tube, stir the swab more than 5 times.
- 6. Remove the swab while squeezing the sides of the tube to extract the liquid from the swab.
- 7. Press the nozzle cap tightly onto the tube.
- 8. Specimen should be tested as soon as possible after collection.
- 9. Specimens may be stored at room temperature (15 – 25°C) or 2-8°C/ 36-46°F for up to 4 hours prior to testing.


- Without the tube squeezing process, improper results may occur due to the large amount of buffer absorption by the swab.
- If the specimen storage condition is out of instructions as below, do not use.
- The nasal swab is stored in extraction buffer for more than 4 hours at 53°C or 205°F.
- Freezing and thawing of nasal swab is more than 1 cycle.

PREPARATION AND TEST PROCEDURE**Preparation**

1. Carefully read instructions for using STANDARD Q COVID-19 Ag Test.

2. Check the expiry date at the back of the foil pouch. Do not use the kit, if expiry date has passed.

3. Check the test device and the desiccant pack in the foil pouch.

Test Procedure

1. Apply 4 drops of extracted specimen to the specimen well of the test device.

2. Read the test result in 15-30 minutes.

PERFORMANCE CHARACTERISTICS**Clinical evaluation**

Clinical performance of the STANDARD Q COVID-19 Ag Test was evaluated using nasal swab samples from 696 subjects in a prospective study at a clinical center in Germany. The study cohort included adults at high risk for SARS-CoV-2 infection according to clinical suspicion. 311 subjects underwent nasal sampling performed by healthcare professionals and 385 subjects followed instructions to obtain a nasal swab sample by themselves. Self-collection was performed under the supervision of healthcare workers without interference or assistance. Test procedures and result reading were always performed by healthcare professionals. RT-PCR tests (Roche cobas® SARS-CoV-2 and Cobas® SARS-CoV-2 e-gene assay) using combined nasopharyngeal/oropharyngeal swab samples were used as the comparator methods. Nasal sampling always preceded the combined NP/OP sampling.

Test sensitivity & specificity

The following tables summarize the patient and performance characteristics of the STANDARD Q COVID-19 Ag Test. The relative sensitivity was 89.6 % (Ct value ≤ 30; 95 % CI: 79.7 % - 95.7 %) for professionally collected samples, and 89.1 % (Ct value ≤ 30; 95 % CI: 78.8 % - 95.5 %) for self-collected samples. For patients for whom days post symptom onset was known, and was 0-5 days, the relative sensitivity in comparison to RT-PCR was 86.7 % (95 % CI: 75.4 % - 94.1 %) for professionally collected nasal samples and 88.9 % (95 % CI: 77.4 % - 95.8 %) for self-collected nasal samples. The relative specificity in comparison to RT-PCR was 99.1 % (95 % CI: 96.9 % - 99.9 %) for professionally collected nasal samples. For patients for whom days post symptom onset was known, and was 0-5 days, the relative specificity in comparison to RT-PCR was 99.1 % (95 % CI: 97.9 % - 99.7 %), respectively.

- Place the test device on a flat surface.
- Dispense the specimen at 90 degree angle to allow for free falling drops and avoid bubbles.
- Do not read test results after 30 minutes. It may give false results.

INTERPRETATION OF TEST RESULTS

Test result	Example	Description
Negative		1. A purple colored band will appear in the top section of the result window to show that the test is working properly. This band is control line (C).
Positive		2. A purple colored band will appear in the lower section of the result window. This band is test line of SARS-CoV-2 antigen (T).
Invalid		3. Even if the control line is faint, or the test line isn't uniform, the test should be considered to be performed properly and the test result should be interpreted as a positive result.

hazard waste. Laboratory chemical and biohazard wastes must be handled and disposed in accordance with all local, state, and national regulations.

11. Desiccat in foil pouch is to absorb moisture and keep humidity from affecting products. If the moisture indicating desiccat bands change from yellow to green, the test device in the pouch should be discarded.

12. This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

- Warning: H317 May cause an allergic skin reaction.
- H412 Harmful to aquatic life with long lasting effects.
- H319 Causes serious eye irritation.
- Prevention: P261 Avoid breathing dust/fume/gas/mist/vapours/spray.
- P273 Avoid release to the environment.
- P280 Wear eye protection/face protection.
- Response: P333 + P313 If skin irritation or rash occurs: Get medical advice/attention. P337 + P313 If eye irritation persists: Get medical advice/attention.
- P362 + P364 Take off contaminated clothing and wash it before reuse.
- For customers in the European Economic Area: Contains SVHC: octyl-nonylphenol ethoxylates.
- For use as part of an IVD method and under controlled conditions only – acc. to Art. 56.3 and 3.23 REACH regulation.

LIMITATION OF TEST

- 1. The test procedure, precautions and interpretation of results for this test must be followed strictly when testing.
- 2. The test should be used for the detection of SARS-CoV-2 antigen in human nasal swab specimens only.
- 3. This test can not be used for quantifying SARS-CoV-2 concentration.
- 4. Failure to follow the test procedure and interpretation of test results may adversely affect test performance and/or produce invalid results.

- 5. The test result must always be evaluated with other data available to the physician.
- 6. A negative result may occur if the concentration of antigen in a specimen is below the detection limit of the test or if the specimen was collected or transported improperly, therefore a negative test result does not eliminate the possibility of SARS-CoV-2 infection, and should be confirmed by viral culture or molecular assay.

- 7. Positive test results do not rule out co-infections with other pathogens.
- 8. The test is not a basis for confirming or excluding diseases only.

EXPLANATION AND SUMMARY

- **Introduction**
Coronavirus is a strand-positive RNA virus with an envelope of about 80 to 120 nm in diameter. Its genetic material is the largest of all RNA viruses and is an important pathogen of many domestic animals, pets, and human diseases. It can cause a variety of acute and chronic diseases. Common signs of a person infected with a coronavirus include respiratory symptoms, fever, cough, shortness of breath, and dyspnea. In more severe cases, infection can cause pneumonia, severe acute respiratory syndrome, kidney failure, and even death. The 2019 new coronavirus, "SARS-CoV-2 (COVID-19)", was discovered because of Wuhan Viral Pneumonia cases in 2019, and was named by the World Health Organization on January 12, 2020. It is known that it can cause colds and Middle East Respiratory Syndrome (MERS) and many other diseases such as Severe Acute Respiratory Syndrome (SARS). These kits are helpful for the auxiliary diagnosis of coronavirus infection. The test results are for clinical reference only and are not used as a basis for confirming or excluding diseases only.
- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.
- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

- **Intended use**
STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleocapsid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

STANDARD Q COVID-19 Ag Test
 STANDARD™ Q COVID-19 Ag Test

LEIA AS INSTRUÇÕES COM ATENÇÃO ANTES DE REALIZAR O TESTE

SD BIOSENSOR

INTERPRETAÇÃO DOS RESULTADOS DO TESTE

Resultado do teste	Exemplo	Descrição
Negativo		1. Uma banda roxa aparecerá na seção superior da janela de resultados para mostrar que o teste está funcionando adequadamente. Essa banda é a linha de controle (C).
Positivo		2. Uma banda roxa aparecerá na seção inferior da janela de resultados. Essa banda é a linha de teste para o antígeno SARS-CoV-2 (T).
Inválido		3. Mesmo se a linha de controle for fraca ou a linha de teste não for visível, o teste deverá ser considerado realizado adequadamente e o resultado deverá ser interpretado como positivo.

* A presença de qualquer linha, não importa quanto fraca, é considerada um resultado positivo.

* Os resultados positivos devem ser considerados em conjunto com o histórico clínico e outros dados disponíveis.

CONTEÚDO DO KIT**RECOLHA E PREPARAÇÃO DA AMOSTRA****■ Preparação da amostra**

[Zaragato nasal]



- Incline o cabeça do paciente levemente para trás.
- Enquanto gira a zaragato, insira na pupila menos duma polegada (cerca de 2 cm) até encontrar resistência no nariz.
- Gire a zaragato 4 vezes contra a parede nasal.
- Repita o procedimento na outra narina usando a mesma zaragato.

Apenas uma zaragato é usada para teste. Esfregue as duas narinas com a mesma zaragato. Com uma pressão adequada, esfregue a zaragato em todas as paredes nasais internas de sua narina. Faça pelo menos 4 movimentos circulares em volta da parede nasal. Não basta simplesmente girar a zaragato.

■ Indicação de uso
O coronavírus é um vírus de RNA de sentido positivo e cadeia simples com um envelope de cerca de 80 a 120 nm de diâmetro. Seu material genético é o maior de todos os vírus de RNA e é um patógeno importante de muitos animais domésticos e selvagens, bem como humanos. É causador de doenças agudas e crônicas. Os síntomas comuns de uma pessoa infectada por um coronavírus incluem sintomas respiratórios, febre, tosse, falta de ar e dispneia. Nos casos mais graves, a infecção pode causar pneumonite, síndrome respiratória aguda grave, insuficiência renal e até mesmo a morte. O novo coronavírus de 2019, ou "SARS-CoV-2 (COVID-19)", foi descoberto por causa dos casos de pneumonia viral ocorridos em Wuhan em 2019 e foi assim denominado pela Organização Mundial da Saúde em 12 de janeiro de 2020, quando se confirmou que esse vírus podia causar resfriados, síndrome respiratória do Oriente Médio (MERS) e doenças mais graves, como a Síndrome Respiratória Aguda Grave (SARS). Estes kits são úteis para auxiliar no diagnóstico de infecção por coronavírus. Os resultados do teste são apenas para referência clínica e não podem ser utilizados isoladamente como base para confirmação ou exclusão de casos.

■ Princípio do teste
O dispositivo de teste STANDARD Q COVID-19 tem duas linhas pré-revestidas, uma de controle "C" e uma de teste "T" na superfície da membrana de nitrocelulose. Nem a linha de controle nem a linha de teste ficam visíveis na janela de resultados antes da aplicação de quaisquer amostras. Anticorpo monoclonal anti-SARS-CoV-2 presente nas amostras de nariz humana e o anticorpo anti-IgG de galinha monoclonal de camundongo revestem a região da linha de controle. Anticorpos monoclonais anti-SARS-CoV-2 na amostra interage com o anticorpo monoclonal anti-SARS-CoV-2 conjugado com partículas coloridas, formando o complexo antígeno-anticorpo-partícula colorida. Esse complexo migra ao longo da membrana, por ação capilar, até a linha de teste, onde será capturado pelo anticorpo monoclonal anti-SARS-CoV-2 de camundongo. Uma linha de teste colorida ficará visível na janela de resultados se houver antígenos contra SARS-CoV-2 na amostra. A intensidade da linha de teste colorida varia dependendo da quantidade de antígeno de SARS-CoV-2 presente na amostra. Se o anticorpo do SARS-CoV-2 estiver na amostra, então nenhuma cor aparecerá na linha de teste. A linha de controle é usada para controlar o procedimento e deve aparecer sempre se o procedimento de teste para realizada da forma correta e os reagentes de teste da linha de controle estiverem funcionando.

■ Conteúdo do kit (09COV31D)
① Dispositivo de teste (individualmente numa bolsa de alumínio com dessecante) x 25
② Tubo buffer de extração x 25 ③ Tampa do bico x 25 ④ Zaragato estéril x 25
⑤ Instruções de uso x 1

MATERIAIS NECESSÁRIOS, MAS NÃO FORNECIDOS

- Equipamento de proteção individual de acordo com as recomendações locais (ou seja, avental/jaleco, máscara facial, proteção facial/olhos de proteção e luvas)
- Temporizador
- Recipiente para produtos de risco biológico
- Coloque o dispositivo de teste e o pacote de dessecante dentro da bolsa de alumínio
- Leia as instruções atentamente antes de usar o STANDARD Q COVID-19 Ag Test.
- Verifique a data de validade na parte traseira da bolsa de alumínio. Se o prazo de validade já expirou, não utilize o kit.
- Verifique o dispositivo de teste e o pacote de dessecante dentro da bolsa de alumínio.
- Procedimento de teste**

4 gotas
1. Aplique 4 gotas da amostra extraída no poço de amostras do dispositivo de teste.
2. Leia o resultado do teste em 15-30 minutos.

■ Advertências e precauções

- Deixe o conteúdo do teste e as amostras atinjam a temperatura ambiente antes de iniciar o teste.
- Não reutilize o kit de teste.
- Não use o kit se a bula estiver danificada ou se o selo estiver danificado.
- Não use o tubo buffer de extração de outro lote.
- Não fume, não beba e não coma enquanto estiver manipulando a amostra.
- Use equipamento de proteção individual, como luvas e aventais de laboratório quando estiver manipulando os reagentes do kit. Lave bem as mãos depois de terminar os testes.
- Lime completamente os respingos usando um desinfetante apropriado.
- Manuseie todas as amostras como se elas contivessem agentes infeciosos.
- Observe as precauções estabelecidas contra riscos microbiológicos durante os procedimentos de teste.
- Desarrolle todas as amostras e todos os materiais utilizados para realização do teste, como resíduos biológicos perigosos. Os resíduos químicos de desinfecção e resíduos biológicos perigosos devem ser manuseados e descartados de acordo com os regulamentos municipais, estaduais e federais.
- O dessecante na bolsa de alumínio serve para absorver a umidade e evitar que o amido para verificá-lo.
- Se os grânulos de dessecante mudarem de cor de amarelo para verificá-los.
- Coloque o dispositivo de teste sobre uma superfície plana.
- Dispense a amostra a um ângulo de 90 graus para permitir que as gotas caiam livremente e evitar a formação de bolhas.
- Não leia os resultados do teste depois de 30 minutos. Ele pode fornecer resultados falsos.

Aviso:
H317 Pode causar uma reação alérgica na pele.
H412 Nocivo para organismos aquáticos com efeitos duradouros. H319 Causa grave irritação dos olhos.
Prevenção:
P261 Evite aspirar poeiras/fumos/gases/névoas/vapores/aerosóis. P273 Evite a liberação para o ambiente.
P280 Utilize proteção ocular/proteção facial.

BIBLIOGRAPHY

- Clinical management of severe acute respiratory infection when novel coronavirus(nCoV) infection is suspected. Interim guidance. WHO.2020
- Diagnostic detection of Wuhan Coronavirus 2019 by real-time RT-PCR.2020
- Diagnosis and treatment of pneumonia caused by new coronavirus (trial version 4) National Health Commission. 2020

STANDARD Q COVID-19 Ag Test
 STANDARD™ Q COVID-19 Ag Test

LEIA AS INSTRUÇÕES COM ATENÇÃO ANTES DE REALIZAR O TESTE

SD BIOSENSOR

CONTEÚDO DO KIT**RECOLHA E PREPARAÇÃO DA AMOSTRA****■ Preparação da amostra**

[Zaragato nasal]



- Incline o cabeça do paciente levemente para trás.
- Enquanto gira a zaragato, insira na pupila menos duma polegada (cerca de 2 cm) até encontrar resistência no nariz.
- Gire a zaragato 4 vezes contra a parede nasal.
- Repita o procedimento na outra narina usando a mesma zaragato.

Apenas uma zaragato é usada para teste. Esfregue as duas narinas com a mesma zaragato. Com uma pressão adequada, esfregue a zaragato em todas as paredes nasais internas de sua narina. Faça pelo menos 4 movimentos circulares em volta da parede nasal. Não basta simplesmente girar a zaragato.

■ Indicação de uso
O coronavírus é um vírus de RNA de sentido positivo e cadeia simples com um envelope de cerca de 80 a 120 nm de diâmetro. Seu material genético é o maior de todos os vírus de RNA e é um patógeno importante de muitos animais domésticos e selvagens, bem como humanos. É causador de doenças agudas e crônicas. Os síntomas comuns de uma pessoa infectada por um coronavírus incluem sintomas respiratórios, febre, tosse, falta de ar e dispneia. Nos casos mais graves, a infecção pode causar pneumonite, síndrome respiratória aguda grave, insuficiência renal e até mesmo a morte. O novo coronavírus de 2019, ou "SARS-CoV-2 (COVID-19)", foi descoberto por causa dos casos de pneumonia viral ocorridos em Wuhan em 2019 e foi assim denominado pela Organização Mundial da Saúde em 12 de janeiro de 2020, quando se confirmou que esse vírus podia causar resfriados, síndrome respiratória do Oriente Médio (MERS) e doenças mais graves, como a Síndrome Respiratória Aguda Grave (SARS). Estes kits são úteis para auxiliar no diagnóstico de infecção por coronavírus. Os resultados do teste são apenas para referência clínica e não podem ser utilizados isoladamente como base para confirmação ou exclusão de casos.

■ Princípio do teste
O dispositivo de teste STANDARD Q COVID-19 tem duas linhas pré-revestidas, uma de controle "C" e uma de teste "T" na superfície da membrana de nitrocelulose. Nem a linha de controle nem a linha de teste ficam visíveis na janela de resultados antes da aplicação de quaisquer amostras. Anticorpo monoclonal anti-SARS-CoV-2 presente nas amostras de nariz humana e o anticorpo anti-IgG de galinha monoclonal de camundongo revestem a região da linha de controle. Anticorpos monoclonais anti-SARS-CoV-2 na amostra interage com o anticorpo monoclonal anti-SARS-CoV-2 conjugado com partículas coloridas, formando o complexo antígeno-anticorpo-partícula colorida. Esse complexo migra ao longo da membrana, por ação capilar, até a linha de teste, onde será capturado pelo anticorpo monoclonal anti-SARS-CoV-2 de camundongo. Uma linha de teste colorida ficará visível na janela de resultados se houver antígenos contra SARS-CoV-2 na amostra. A intensidade da linha de teste colorida varia dependendo da quantidade de antígeno de SARS-CoV-2 presente na amostra. Se o anticorpo do SARS-CoV-2 estiver na amostra, então nenhuma cor aparecerá na linha de teste. A linha de controle é usada para controlar o procedimento e deve aparecer sempre se o procedimento de teste para realizada da forma correta e os reagentes de teste da linha de controle estiverem funcionando.

■ Conteúdo do kit (09COV31D)
① Dispositivo de teste (individualmente numa bolsa de alumínio com dessecante) x 25
② Tubo buffer de extração x 25 ③ Tampa do bico x 25 ④ Zaragato estéril x 25
⑤ Instruções de uso x 1

MATERIAIS NECESSÁRIOS, MAS NÃO FORNECIDOS

- Equipamento de proteção individual de acordo com as recomendações locais (ou seja, avental/jaleco, máscara facial, proteção facial/olhos de proteção e luvas)
- Temporizador
- Recipiente para produtos de risco biológico
- Coloque o dispositivo de teste e o pacote de dessecante dentro da bolsa de alumínio
- Leia as instruções atentamente antes de usar o STANDARD Q COVID-19 Ag Test.
- Verifique a data de validade na parte traseira da bolsa de alumínio. Se o prazo de validade já expirou, não utilize o kit.
- Verifique o dispositivo de teste e o pacote de dessecante dentro da bolsa de alumínio.
- Procedimento de teste**

4 gotas
1. Aplique 4 gotas da amostra extraída no poço de amostras do dispositivo de teste.
2. Leia o resultado do teste em 15-30 minutos.

■ Advertências e precauções

- Deixe o conteúdo do teste e as amostras atinjam a temperatura ambiente antes de iniciar o teste.
- Não reutilize o kit de teste.
- Não use o kit se a bula estiver danificada ou se o selo estiver danificado.
- Não use o tubo buffer de extração de outro lote.
- Não fume, não beba e não coma enquanto estiver manipulando a amostra.
- Use equipamento de proteção individual, como luvas e aventais de laboratório quando estiver manipulando os reagentes do kit. Lave bem as mãos depois de terminar os testes.
- Lime completamente os respingos usando um desinfetante apropriado.
- Manuseie todas as amostras como se elas contivessem agentes infeciosos.
- Observe as precauções estabelecidas contra riscos microbiológicos durante os procedimentos de teste.
- Desarrolle todas as amostras e todos os materiais utilizados para realização do teste, como resíduos biológicos perigosos. Os resíduos químicos de desinfecção e resíduos biológicos perigosos devem ser manuseados e descartados de acordo com os regulamentos municipais, estaduais e federais.
- O dessecante na bolsa de alumínio serve para absorver a umidade e evitar que o amido para verificá-los.
- Coloque o dispositivo de teste sobre uma superfície plana.
- Dispense a amostra a um ângulo de 90 graus para permitir que as gotas caiam livremente e evitar a formação de bolhas.
- Não leia os resultados do teste depois de 30 minutos. Ele pode fornecer resultados falsos.

Aviso:
H317 Pode causar uma reação alérgica na pele.
H412 Nocivo para organismos aquáticos com efeitos duradouros. H319 Causa grave irritação dos olhos.
Prevenção:
P261 Evite aspirar poeiras/fumos/gases/névoas/vapores/aerosóis. P273 Evite a liberação para o ambiente.
P280 Utilize proteção ocular/proteção facial.

STANDARD Q COVID-19 Ag Test
 STANDARD™ Q COVID-19 Ag Test

LEIA AS INSTRUÇÕES COM ATENÇÃO ANTES DE REALIZAR O TESTE

SD BIOSENSOR

CONTEÚDO DO KIT**RECOLHA E PREPARAÇÃO DA AMOSTRA****■ Preparação da amostra**

[Zaragato nasal]



- Incline o cabeça do paciente levemente para trás.
- Enquanto gira a zaragato, insira na pupila menos duma polegada (cerca de 2 cm) até encontrar resistência no nariz.
- Gire a zaragato 4 vezes contra a parede nasal.
- Repita o procedimento na outra narina usando a mesma zaragato.

Apenas uma zaragato é usada para teste. Esfregue as duas narinas com a mesma zaragato. Com uma pressão adequada, esfregue a zaragato em todas as paredes nasais internas de sua narina. Faça pelo menos 4 movimentos circulares em volta da parede nasal. Não basta simplesmente girar a zaragato.

■ Indicação de uso
O coronavírus é um vírus de RNA de sentido positivo e cadeia simples com um envelope de cerca de 80 a 120 nm de diâmetro. Seu material genético é o maior de todos os vírus de RNA e é um patógeno importante de muitos animais domésticos e selvagens, bem como humanos. É causador de doenças agudas e crônicas. Os síntomas comuns de uma pessoa infectada por um coronavírus incluem sintomas respiratórios, febre, tosse, falta de ar e dispneia. Nos casos mais graves, a infecção pode causar pneumonite, síndrome respiratória aguda grave, insuficiência renal e até mesmo a morte. O novo coronavírus de 2019, ou "SARS-CoV-2 (COVID-19)", foi descoberto por causa dos casos de pneumonia viral ocorridos em Wuhan em 2019 e foi assim denominado pela Organização Mundial da Saúde em 12 de janeiro de 2020, quando se confirmou que esse vírus podia causar resfriados, síndrome respiratória do Oriente Médio (MERS) e doenças mais graves, como a Síndrome Respiratória Aguda Grave (SARS). Estes kits são úteis para auxiliar no diagnóstico de infecção por coronavírus. Os resultados do teste são apenas para referência clínica e não podem ser utilizados isoladamente como base para confirmação ou exclusão de casos.

■ Princípio do teste
O dispositivo de teste STANDARD Q COVID-19 tem duas linhas pré-revestidas, uma de controle "C" e uma de teste "T" na superfície da membrana de nitrocelulose. Nem a linha de controle nem a linha de teste ficam visíveis na janela de resultados antes da aplicação de quaisquer amostras. Anticorpo monoclonal anti-SARS-CoV-2 presente nas amostras de nariz humana e o anticorpo anti-IgG de galinha monoclonal de camundongo revestem a região da linha de controle. Anticorpos monoclonais anti-SARS-CoV-2 na amostra interage com o anticorpo monoclonal anti-SARS-CoV-2 conjugado com partículas coloridas, formando o complexo antígeno-anticorpo-partícula colorida. Esse complexo migra ao longo da membrana, por ação capilar, até a linha de teste, onde será capturado pelo anticorpo monoclonal anti-SARS-CoV-2 de camundongo. Uma linha de teste colorida ficará visível na janela de resultados se houver antígenos contra SARS-CoV-2 na amostra. A intensidade da linha de teste colorida varia dependendo da quantidade de antígeno de SARS-CoV-2 presente na amostra. Se o anticorpo do SARS-CoV-2 estiver na amostra, então nenhuma cor aparecerá na linha de teste. A linha de controle é usada para controlar o procedimento e deve aparecer sempre se o procedimento de teste para realizada da forma correta e os reagentes de teste da linha de controle estiverem funcionando.

■ Conteúdo do kit (09COV31D)
① Dispositivo de teste (individualmente numa bolsa de

sont présents dans l'échantillon, aucun couleur n'apparaît sur la ligne de test. La ligne de contrôle est utilisée pour le contrôle procédural et doit toujours apparaître si la procédure de test est correctement effectuée et que les réactifs de test de la ligne de contrôle fonctionnent.

■ Matériaux reçus mais non fournis

- Équipements de protection individuelle conformément aux recommandations locales (ex: blouse de laboratoire, masque facial, masque/lunettes de protection et gants)
- Minuterie
- Conteneur Biohazard

KIT DE STOCKAGE ET STABILITÉ

Conserver le kit à une température entre 2 et 30 °C (36 - 86 °F), à l'abri de la lumière directe du soleil. Les matériaux du kit sont stables jusqu'à leur date d'expiration indiquée à l'extérieur de la boîte. Ne pas congeler le kit.

AVERTISSEMENTS ET PRÉCAUTIONS

- Laissez les matériaux du kit et les échantillons revenir à température ambiante avant de lancer le test.
- Ne pas réutiliser le kit de test.
- Ne pas utiliser le kit de test si le sachet est endommagé ou ouvert.
- Ne pas utiliser le tube avec le tampon d'extraction d'un autre lot.
- Ne pas fumer, boire ou manger pendant la manipulation des échantillons.
- Ne pas effectuer une manipulation individuelle, comme des gants et une blouse de laboratoire lors de la manipulation des kits de réactifs. Se laver les mains soigneusement après les tests.
- Nettoyer soigneusement les déversements en utilisant un désinfectant approprié.
- Manipuler tous les échantillons avec les mêmes précautions que s'ils contenaient des agents infectieux.
- Respecter les précautions établies contre les dangers microbiologiques pendant toute la procédure.
- Éliminer tous les échantillons et les matériaux utilisés pour effectuer le test en les considérant comme des déchets dangereux. Manipuler et éliminer les déchets dans les échantillons biologiques conformément aux réglementations locales, régionales et nationales.
- Le dessicant dans le sachet en aluminium absorbe l'humidité pour empêcher d'affecter les produits. Si les perles de dessiccant indiquent l'humidité changent de jaune à vert, le dispositif de test dans la pochette doit être jeté.
- Le kit comprend des composants classés comme suit, conformément à la réglementation (EC) No 1272/2008:
 - Avertissement : H317 Peut provoquer une réaction allergique cutanée.
 - H412 Nocif pour les organismes aquatiques, entraîne des effets néfastes à long terme.
 - H319 Ce détergent de synthèse peut irriter fortement les yeux.
- Prévention : P261 Eviter de respirer les poussières/fumées/gaz/brouillards/vapeurs/aérosols.
- P273 Éviter le rejet dans l'environnement.
- P280 Porter un équipement de protection des yeux/visage

P333 + P313 En cas d'irritation ou d'éruption cutanée : consulter un médecin. P337 + P313 Si irritation persiste : consulter un médecin.

P362 + P364 Enlever les vêtements contaminés et les laver avant réutilisation.

Concernant les clients de l'Espace économique européen : Contenu des substances extrêmement préoccupantes (SVHC) : acétyle/monoxyde de carbone.

Pour une utilisation dans le cadre d'un diagnostic *in vitro* et uniquement dans des conditions contrôlées : conformément aux articles 56.3 et 3.23 du règlement "Enregistrement, évaluation, autorisation et restriction des substances chimiques".

LIMITES DU TEST

1. La procédure de test, les précautions et l'interprétation des résultats pour ce kit de test doivent être strictement suivies lors du test.

2. Ce test doit être utilisé pour détecter l'antigène du SARS-CoV-2 uniquement dans les échantillons d'écouvillons nasopharyngés humains. Les autres types d'échantillons n'ont pas été validés.

3. Ce test ne peut pas être utilisé pour classifier la concentration de l'antigène SARS-CoV-2.

4. Le résultat de la procédure de test et de l'interprétation des résultats de test peut affecter négativement les performances du test et/ou produire des résultats invalides.

5. Le résultat de test doit toujours être évalué avec d'autres données disponibles pour le médecin.

6. Il est possible d'obtenir un résultat négatif si la concentration d'anticorps ou d'anticorps dans un échantillon est inférieure au seuil de détection du test ou si l'échantillon n'a pas été correctement prélevé ou transporté. Par conséquent, un résultat négatif n'implique pas la possibilité d'être infecté par le SARS-CoV-2, et l'infection doit être confirmée par une culture virale ou un test moléculaire.

7. Les résultats positifs n'excluent pas la possibilité d'une co-infection avec d'autres pathogènes.


8. Lors de l'utilisation d'un milieu de transport viral (VTM), la sensibilité peut être réduite en raison de la dilution.

9. Seuls les moyens de transports Copan UTM, BD UTM, et STANDARD™ ont été validés pour cet essai.

■ COLETA DO ESPECÍME

[Swab nasofaringe]

- Para recolher uma amostra nasofaringea com o uso da zaragata, insira a zaragata estéril na narina do paciente, atingindo a superfície da nasofaringe posterior.
- Gire a zaragata delicadamente, empurrando-a até encontrar resistência na altura do osso turbinado (concha nasal).
- Estique a zaragata na superfície da nasofaringe 3 a 4 vezes com movimentos circulares.
- Remova a zaragata da narina cuidadosamente.



3-4 vezes



Cat No. 09COV33D

STANDARD Q COVID-19 Ag Test

STANDARD™ Q COVID-19 Ag Test


PLEASE READ INSTRUCTIONS CAREFULLY BEFORE YOU PERFORM THE TEST

© SD BIOSENSOR

KIT CONTENTS**SPECIMEN COLLECTION AND PREPARATION****■ Specimen preparation [Nasal swab]**

1. Tilt patient's head back slightly.
2. While rotating the swab, insert swab less than one inch (about 2cm) into nostril until resistance is met at turbinate.
3. Rotate the swab 4 times against nasal wall.
4. Repeat in other nostril using the same swab.

Only one swab is used per test. Rub both nostrils with the same swab. Using proper pressure, rub the swab against all of the inside nasal walls of your nostril. Make at least 4 big circles along the nasal wall. Do not just spin the swab.




5. Insert the swab into an extraction buffer tube. While squeezing the buffer tube, stir the swab more than 5 times.
6. Remove the swab while squeezing the sides of the tube to extract the liquid from the swab.
7. Press the nozzle cap tightly onto the tube.

Specimens should be tested as soon as possible after collection.

Specimens may be stored at room temperature (15 – 25°C) or 2-8°C/ 36-46°F for up to 4 hours prior to testing.

- Without the tube squeezing process, improper results may occur due to the large amount of buffer absorption by the swab.
- If the specimen storage condition is out of instructions as below, do not use.
- The nasal swab is stored in extraction buffer for more than 4 hours at 53°C or 205°F.
- Freezing and thawing of Nasal swab is more than 1 cycle.

PREPARATION AND TEST PROCEDURE**■ Preparation**

1. Carefully read instructions for using STANDARD Q COVID-19 Ag Test.
2. Check the expiry date at the back of the foil pouch. Do not use the kit, if expiry date has passed.
3. Check the test device and the desiccant pack in the foil pouch.

■ Test Procedure

1. Apply 4 drops of extracted specimen to the specimen well of the test device.
2. Read the result in 15-30 minutes.

- Place the test device on a flat surface.
- Dispense the specimen at 90 degree angle to allow for free falling drops and avoid bubbles.
- Do not read test results after 30 minutes. It may give false results.

INTERPRETATION OF CONTROL TEST RESULT


INTERPRETATION OF TEST RESULTS

Test result	Example	Description
Negative		depending upon the amount of SARS-CoV-2 antigen present in the specimen. If SARS-CoV-2 antigens are not present in the specimen, then no color appears in the test line. The control line is used for procedural control, and should always appear if the test procedure is performed properly and the test reagents of the control line are working.
Positive		1. A purple colored band will appear in the top section of the result window to show that the test is working properly. This band is control line (C).
Invalid		2. A purple colored band will appear in the lower section of the result window. This band is test line of SARS-CoV-2 antigen (T).

1. Even if the control line is faint, or the test line isn't uniform, the test should be considered to be performed properly and the test result should be interpreted as a positive result.

* The presence of any line no matter how faint the result is considered positive.


* Positive results should be considered in conjunction with the clinical history and other data available.

CONTROL PREPARATION AND TEST PROCEDURE**- Positive/Negative control****■ Preparation**

1. Put the positive or negative control swab into an extraction buffer tube.
2. While squeezing the buffer tube, stir the swab more than 5 times.

3. Remove the swab while squeezing the sides of the tube to extract the liquid from the swab.

4. Press the nozzle cap tightly onto the tube.

■ Test Procedure

1. Apply 4 drops of extraction buffer to the specimen well.
2. Read the result in 15-30 minutes.

- Place the test device on a flat surface.
- Dispense the specimen at 90 degree angle to allow for free falling drops and avoid bubbles.
- Do not read test results after 30 minutes. It may give false results.

INTERPRETATION OF CONTROL TEST RESULT

STANDARD COVID-19 Ag Control Positive swab: Positive	Result	Interpretation	Follow up
Test (T) Line Positive	PASS	-	
Test (T) Line Negative	FAIL	Retest	
No Control (C) Line	Invalid	Retest	

STANDARD COVID-19 Ag Negative Control swab: Negative	Result	Interpretation	Follow up
Test (T) Line Negative	PASS	-	
Test (T) Line Positive	FAIL	Retest	
No Control (C) Line	Invalid	Retest	

INTERPRETATION AND SUMMARY

■ Introduction

Coronavirus is a single-stranded positive-sense RNA virus with an envelope of about 80 to 120 nm in diameter. Its material genetic is the largest of all RNA viruses and is an important pathogen of many domestic animals, pets, and human diseases. It can cause a variety of acute and chronic diseases. Common signs of a person infected with a coronavirus include respiratory symptoms, fever, cough, shortness of breath, and dyspnea. In more severe cases, infection can cause pneumonia, severe acute respiratory syndrome, kidney failure, and even death. The 2019 new coronavirus or SARS-CoV-2 (COVID-19), was discovered because of the first case reported in Wuhan, China on December 31, 2019. World Health Organization on January 12, 2020, confirming that it can cause colds and more serious diseases such as Severe Acute Respiratory Syndrome (SARS). These kits are helpful for the auxiliary diagnosis of coronavirus infection. The test results are for clinical reference only and cannot be used as a basis for confirming or excluding cases alone.

■ Intended use

STANDARD Q COVID-19 Ag Test is a rapid chromatographic immunoassay for the qualitative detection of specific nucleic acid protein antigen from SARS-CoV-2 present in human nasal or nasopharyngeal specimens. This product is intended for healthcare professionals at the clinical setup and point of care sites, as an aid to early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms of SARS-CoV-2 infection. It provides only an initial screening test result. This product is strictly for medical professional use only and not intended for personal use. The administration of the test and the interpretation of the results should be done by a trained health professional. The result of this test should not be the sole basis for the diagnosis; confirmatory testing is required.

■ Test principle

STANDARD Q COVID-19 Ag test device has two pre-coated lines, "C" Control line, "T" Test line on the surface of the nitrocellulose membrane. Both the control line and test line in the result window are not visible before applying any specimens. Monoclonal anti-SARS-CoV-2 antibody conjugated with color particles are used as detectors for SARS-CoV-2 antigen device. During the test, SARS-CoV-2 antigen in the specimen interact with monoclonal anti-SARS-CoV-2 antibody conjugated with color particles making antigen-antibody color particle complex. This complex migrates on the membrane via capillary action until the test line, where it will be captured by the mouse monoclonal anti-SARS-CoV-2 antibody. A colored line will be visible in the result window if SARS-CoV-2 antigens are present in the specimen. The intensity of colored test line will vary

PERFORMANCE CHARACTERISTICS**■ Clinical evaluation**

The following tables summarize the patient and performance characteristics of the STANDARD Q COVID-19 Ag Test. The relative sensitivity was 89.6% (Ct value < 30; 95% CI: 79.1% - 95.7%) for professionally collected samples, and 89.1% (Ct value < 30; 95% CI: 78.8% - 95.5%) for self-collected samples. For patients for whom days post symptom onset were known, and was 0-5 days, the relative sensitivity in comparison to RT-PCR was 86.7% (95% CI: 75.4% - 94.1%) for professionally collected healthcare workers without interference or assistance. Test procedures and result reading were always performed by healthcare professionals. RT-PCR tests (Roche cobas® SARS-CoV-2 and Abbott SARS-CoV-2 E-gene assay) using combined nasopharyngeal/oropharyngeal swab samples were used as the comparator methods. Nasal samples always preceded the combined NP/OP sampling.

■ Test sensitivity & specificity

The following tables summarize the patient and performance characteristics of the STANDARD Q COVID-19 Ag Test. The relative sensitivity was 89.6% (Ct value < 30; 95% CI: 79.1% - 95.7%) for professionally collected samples, and 89.1% (Ct value < 30; 95% CI: 78.8% - 95.5%) for self-collected samples. For patients for whom days post symptom onset were known, and was 0-5 days, the relative sensitivity in comparison to RT-PCR was 86.7% (95% CI: 75.4% - 94.1%) for professionally collected healthcare workers without interference or assistance. Test procedures and result reading were always performed by healthcare professionals. RT-PCR tests (Roche cobas® SARS-CoV-2 and Abbott SARS-CoV-2 E-gene assay) using combined nasopharyngeal/oropharyngeal swab samples were used as the comparator methods. Individuals were evaluated using the STANDARD Q COVID-19 Ag Test. The relative sensitivity and relative specificity were 82.7% (95% CI: 75.6% - 88.4%) and 99.1% (95% CI: 97.9% - 99.7%), respectively.

a) For samples run on cobas® Target 2 (E gene) Ct values were used.

b) Relative specificity, % (95% CI), N

c) All Ct values

d) Professional collection

e) Self-collection

f) Combined OP/NP

g) Relative sensitivity, % (95% CI), N

h) All Ct values

i) Professional collection

j) Self-collection

k) Combined OP/NP

l) Relative specificity, % (95% CI), N

m) All Ct values

n) Professional collection

o) Self-collection

p) Combined OP/NP

q) Relative sensitivity, % (95% CI), N

r) All Ct values

s) Professional collection

t) Self-collection

u) Combined OP/NP

v) Relative specificity, % (95% CI), N

w) All Ct values

x) Professional collection

y) Self-collection

z) Combined OP/NP

aa) Relative sensitivity, % (95% CI), N

bb) All Ct values

cc) Professional collection

dd) Self-collection

ee) Combined OP/NP

ff) Relative specificity, % (95% CI), N

gg) All Ct values

hh) Professional collection

ii) Self-collection

jj) Combined OP/NP

kk) Relative specificity, % (95% CI), N

ll) All Ct values

mm) Professional collection

nn) Self-collection

oo) Combined OP/NP

pp) Relative specificity, % (95% CI), N

qq) All Ct values

rr) Professional collection

ss) Self-collection

tt) Combined OP/NP

uu) Relative specificity, % (95% CI), N

vv) All Ct values

ww) Professional collection

xx) Self-collection

yy) Combined OP/NP

zz) Relative specificity, % (95% CI), N

aa) All Ct values

bb) Professional collection

cc) Self-collection

dd) Combined OP

Réponse :
P333 + P313 En cas d'irritation ou d'éruption cutanée : consulter un médecin.
P337 + P313 Si l'irritation persiste : consulter un médecin.
P362 + P364 Enlever les vêtements et les laver avant réutilisation.
Concernant les clients de l'Espace économique européen : Contient des substances extrêmement préoccupantes (SVHC) : acétate d'octyle/nonylphénol. Pour une utilisation dans le cadre d'un diagnostic in vitro et uniquement dans des conditions contrôlées - conformément aux articles 56.3 et 3.23 du règlement "Régistrement, évaluation, autorisation et restriction des substances chimiques".

LIMITES DU TEST

- Les procédures de test, les précautions et l'interprétation des résultats pour ce kit de test doivent être rigoureusement respectées lors du test.
- Ce test doit être utilisé pour détecter les antigènes SARS-CoV-2 uniquement dans les échantillons dérivés humains nasaux. Les autres types d'échantillons n'ont pas été validés.
- Ce test ne peut pas être utilisé pour quantifier la concentration d'antigènes SARS-CoV-2.
- Le non-respect de la procédure de test et de l'interprétation des résultats de test peuvent nuire à la performance du test et/ou entraîner des résultats de test invalides.
- Les résultats de test doivent toujours être évalués avec d'autres données mises à disposition du médecin.
- Il est possible d'obtenir un résultat négatif si la concentration d'antigènes dans l'échantillon n'est pas suffisante au cours de la détection du test et si l'échantillon n'a pas été correctement prélevé ou transporté. Par conséquent, un résultat négatif n'implique pas la possibilité d'être infecté par le SARS-CoV-2, et l'infection devra être confirmée par une culture virale ou un test moléculaire.
- Les résultats positifs n'excluent pas la possibilité d'une co-infection par d'autres pathogènes.

PT

FE-Q-NCOV-01G

Cat No. 09COV33D

STANDARD Q COVID-19 Ag Test

STANDARD™ Q COVID-19 Ag Test

LEIA AS INSTRUÇÕES COM ATENÇÃO ANTES DE REALIZAR O TESTE

SD BIOSENSOR

CONTEÚDO DO KIT




RECOLHA E PREPARAÇÃO DA AMOSTRA

■ Preparação da amostra


[Zaragatoa nasal]



- Incline a cabeça do paciente ligeiramente para trás.
- Enquanto gira a zaragatoa, insira-a na narina menos dura plegada (cerca de 2 cm) até encontrar resistência no narino.
- Gire a zaragatoa 4 vezes contra a parede nasal.
- Repita o procedimento na outra narina usando a mesma zaragatoa.



- Após uma zaragatoa é usado para teste. Enfraqueça as duas narinas com a mesma zaragatoa. Com uma pressão adequada, esfregue a zaragatoa em todas as paredes nasais internas de sua narina. Faça pelo menos 4 movimentos circulares em volta da parede nasal. Não basta simplesmente girar a zaragatoa.



- Insira a zaragatoa num tubo buffer de extração. Enquanto abre o tubo de extração, mexa a zaragatoa mais de 5 vezes.

- Remova a zaragatoa enquanto aperta os lados do tubo para extraír o líquido da zaragatoa.

- Pressione a tampa do bico firmemente no tubo.

- A amostra deve ser testada o mais rápido possível após a recolha.

- As amostras podem ser armazenadas em temperatura (15 - 25°C) ou a 2-8°C/36-46°F por até 4 horas antes do teste.

- Se o processo de compressão do tubo, podem ocorrer resultados inadequados devido à grande quantidade de absorção do buffer pela zaragatoa.

- Se as condições de armazenamento da amostra estiverem fora das instruções abaixo, não use.

- A zaragatoa nasal é armazenada em buffer de extração por mais de 4 horas a 5±3°C ou 20±5°C.

- O congelamento e descongelamento da zaragatoa nasal demora mais de 1 ciclo.

PREPARAÇÃO E PROCEDIMENTO DE TESTE

■ Preparação



- Leia as instruções atentamente antes de usar o STANDARD Q COVID-19 Ag Test.

- Verifique a data de validade na parte traseira da bolsa de alumínio. Se o prazo de validade já expirou, não utilize o kit.

BIBLIOGRAPHY

- Clinical management of severe acute respiratory infection when novel coronavirus(nCoV) infection is suspected. Interim guidance. WHO.2020
- Diagnostic detection of Wuhan Coronavirus 2019 by real-time RT-PCR.2020
- Diagnosis and treatment of pneumonia caused by new coronavirus (trial version 4) National Health Commission. 2020

3. Verifique o dispositivo de teste e o pacote de dessecante dentro da bolsa de alumínio.

■ Procedimento de teste



- Aplique 4 gotas da amostra extraída no poço de amostras do dispositivo de teste.

- O resultado do teste em 15-30 minutos.

- Coloque o dispositivo de teste sobre uma superfície plana.
- Dispense a amostra a um ângulo de 90° para permitir que as gotas caíam livremente e evitar a formação de bolhas.
- Não use os resultados do teste depois de 30 minutos. Ele pode fornecer resultados falsos.

INTERPRETAÇÃO DOS RESULTADOS DO TESTE

Resultado do teste	Exemplo	Descrição
Negativo		
Positivo		
Inválido		

1. Uma banda roxa aparecerá na seção superior da janela de resultados para mostrar que o teste está funcionando adequadamente. Essa banda é a linha de controle (C).

2. Uma banda roxa aparecerá na seção inferior da janela de resultados. Essa banda é a linha de teste do antígeno SARS-CoV-2 (T).

3. Mesmo se a linha de controle for fraca ou a linha de teste não for uniforme, o teste deverá ser considerado realizado adequadamente e o resultado deverá ser interpretado como positivo.

MATERIAIS NECESSÁRIOS, MAS NÃO FORNECIDOS

- Equipamento de proteção individual de acordo com as recomendações locais (ou seja, avental/palete, máscara facial, proteção facial/olhos de proteção e luvas).
- Temporizador
- Recipiente para produtos de risco biológico

ARMAZENAMENTO E ESTABILIDADE DO KIT

Armazene o kit a 2-30°C / 36-86°F longe da luz solar direta. Os materiais do kit são estáveis até a data de validade impressa na caixa. Não congele o kit.

ADVERTÊNCIAS E PRECAUÇÕES

- Deixe que o conteúdo do kit e as amostras atinjam a temperatura ambiente antes de realizar o teste.
- Não utilize o kit de teste.
- Não use o kit de teste se a bula estiver danificada ou se o selo estiver quebrado.
- Não utilize o kit de extração de corte liso.
- Use equipamento de proteção individual, como luvas e aventais de laboratório quando estiver manuseando os reagentes do kit. Lave bem os mãos depois de terminar os testes.
- Lime completamente os respingos usando um desinfetante apropriado.
- Manuseie todas as amostras e os materiais utilizados para realização do teste como resíduos biológicos perigosos. Os resíduos químicos de laboratório e de equipamentos de proteção individual devem ser eliminados de acordo com as normas ambientais estabelecidas.
- O descontento da bala de alumínio serve para absorver a umidade e evitar que ela afete os produtos. Se os grânulos de dessecante mudarem de cor para verde indicando umidade, o dispositivo de teste na bala deverá ser descartado.
- Este kit contém componentes classificados de acordo com o Regulamento (CE) № 1272/2008:

- Aviso: H317 Pode causar uma reação alérgica na pele.
- H412 Noctivo para organismos aquáticos com efeitos duradouros. H319 Causa irriteção dos olhos.
- Prevenção: P261 Evite aspirar/levar fumos/gases/névoas/vapores/aerosóis. P273 Evite a liberação para o ambiente.
- P280 Utilize proteção ocular/proteção facial.


- Resposta: P333 + P313 Caso ocorra irritação ou erupção na pele: Procure orientação/cuidados médicos.

- P337 + P313 Caso a irritação nos olhos persista: Procure orientação/cuidados médicos. P362 + P364 Retire a roupa contaminada e a lave antes de usá-la novamente.

- Para clientes no Espaço Económico Europeu: Contém SVHC: octil/nonyfenol etoxilado. Para uso como parte de um método IVD e somente sob condições controladas - de acordo com o art. 56.3 e 3.23 Norma REACH.

PREPARAÇÃO DO CONTROLE E PROCEDIMENTO DE TESTE

■ Controle positivo/negativo




1. Coloque a zaragatoa de controle positivo ou negativo num tubo buffer de extração. Mexa a zaragatoa mais de 5 vezes.

2. Remova a zaragatoa enquanto aperta os lados do tubo para extraír o líquido da zaragatoa.

3. Pressione a tampa do bico firmemente no tubo.

■ Procedimento de teste



1. Aplique 4 gotas da amostra extraída no poço de amostras do dispositivo de teste.

2. O resultado do teste em 15-30 minutos.

- Coloque o dispositivo de teste sobre uma superfície plana.

- Dispense a amostra a um ângulo de 90° para permitir que as gotas caíam livremente e evitar a formação de bolhas.

- Não use os resultados do teste depois de 30 minutos. Ele pode fornecer resultados falsos.

INTERPRETAÇÃO DO RESULTADO DO TESTE DE CONTROLE

Zaragatoa de controle positivo para STANDARD COVID-19 Ag: Positivo

Resultado Interpretation Acompanhamento

Resultado	Interpretation	Acompanhamento
Linha de teste (T) Positivo	Passe	-
Linha de teste (T) Negativo	Falha	Testar novamente

Zaragatoa de controle negativo para STANDARD COVID-19 Ag: Negativo

Resultado Interpretation Acompanhamento

Resultado	Interpretation	Acompanhamento
Linha de teste (T) Negativo	Passe	-
Linha de teste (T) Positivo	Falha	Testar novamente
Sem linha de controle (C)	Inválido	Testar novamente

EXPLICAÇÃO E RESUMO

■ Introdução

- O coronavírus é um vírus de RNA de sentido positivo e cadeia simples com um envelope de cerca de 80 a 120 nm de diâmetro. Seu material genético é o maior de todos os vírus de envelopamento de ambientes. Antes de 2019, havia poucas informações sobre o coronavírus. O vírus causou surtos de doenças agudas e crônicas. Os sintomas comuns de uma pessoa infectada por um coronavírus incluem sintomas respiratórios, febre, tosse, falta de ar e dispneia. Nos casos mais graves, a infecção pode causar pneumonia, síndrome respiratória aguda grave, insuficiência renal e até mesmo a morte. O novo coronavírus de 2019, ou "SARS-CoV-2 (COVID-19)", foi descoberto por causa dos casos de pneumonia viral ocorridos em Wuhan em 2019 e foi assim denominado pela Organização Mundial da Saúde em 12 de janeiro de 2020, quando se confirmou que esse vírus pode causar resfriados, síndrome respiratória do Oriente Médio (MERS) e doenças mais graves, como a síndrome Respiratória Aguda Grave (SARS). Estes kits são úteis para auxiliar no diagnóstico de infecção por coronavírus. Os resultados do teste são úteis para referência clínica e não podem ser utilizados isoladamente como base para confirmação ou exclusão de casos.

■ Indicações de uso

- O Teste STANDARD Q COVID-19 Ag é um imunoensaio cromotrofico rápido para a detecção qualitativa de anticorpos de nucleocapside específicos para SARS-CoV-2 presentes nas secreções nasais humanas ou nasofaringe. Este produto é destinado a profissionais de saúde, em instalações clínicas e locais de atendimento, como um auxílio para o diagnóstico precoce de infecção por SARS-CoV-2 em pacientes com sintomas clínicos de infecção por SARS-CoV-2. Fornece

apenas o resultado de uma triagem inicial. Este produto é estritamente apenas para uso profissional médico e não destina ao uso pessoal. A administração do teste e a interpretação dos resultados devem ser realizados por um profissional da saúde competente.

■ Princípio do teste

O dispositivo de teste STANDARD Q COVID-19 Ag tem duas linhas pré-revestidas, uma de controle ("C", linha de teste "T") na superfície da membrana de nitrocelulose. Nem a linha de controle nem a linha de teste ficam visíveis na janela de resultados antes da aplicação de quaisquer amostras. Anticorpo monoclonal anti-SARS-CoV-2 de camundongo reveste a região da linha de teste e o anticorpo monoclonal anti-Zaragatoa esterilizado com partículas coloridas são usados como detectores para o dispositivo antígeno.

Anticorpo monoclonal anti-SARS-CoV-2 de camundongo conjugado com partículas coloridas são usados como detectores para o dispositivo antígeno.

Quando o complexo antígeno-anticorpo é adicionado ao dispositivo de teste, o anticorpo monoclonal anti-SARS-CoV-2 se une ao anticorpo monoclonal anti-Zaragatoa esterilizado.

Quando o complexo antígeno-anticorpo é adicionado ao dispositivo de teste, o anticorpo monoclonal anti-SARS-CoV-2 se une ao anticorpo monoclonal anti-Zaragatoa esterilizado.

Quando o complexo antígeno-anticorpo é adicionado ao dispositivo de teste, o anticorpo monoclonal anti-SARS-CoV-2 se une ao anticorpo monoclonal anti-Zaragatoa esterilizado.

STANDARD COVID-19 Ag Control swab

STANDARD™ COVID-19 Ag Control swab

PLEASE READ INSTRUCTIONS CAREFULLY BEFORE YOU
PERFORM THE TEST



EXPLANATION AND SUMMARY

■ Intended use


STANDARD COVID-19 Ag Control swab is intended for use as an external quality control material to monitor the performance of STANDARD COVID-19 Ag products. It is important to perform control tests with positive and negative control materials to assure STANDARD COVID-19 Ag testing system is working properly.

■ Principle

This product was designed for use with STANDARD COVID-19 Ag products for purpose of monitoring assay performance and maintaining quality assurance.

KIT CONTENTS

STANDARD COVID-19 Ag Control swab is contained 10 Positive control swabs and 10 Negative control swabs.



MATERIALS REQUIRED BUT NOT PROVIDED

- STANDARD COVID-19 Ag Products

PREPARATION AND CONTROL TEST PROCEDURE

This product should be treated the same as patient specimens and run in accordance with instruction accompanying the instrument and kit being used. It is recommended that positive and negative controls be run:


- Once for each new lot,
- Once for each untrained operator,
- Once for each new shipments of test kits,
- As required by test procedures in this instructions and in accordance with local, state and federal regulations of accreditation requirements.

■ Preparation

1. Check the expiry date at the label of a STANDARD COVID-19 Ag Control swab package and Front of the foil pouch of a STANDARD COVID-19 Ag Control swab. Use another lot, if it has expired.




2. Insert the positive or negative control swab into an extraction buffer tube which is in the STANDARD COVID-19 Ag products. Stir the swab at least five times.



3. Remove the swab while squeezing the sides of the tube to extract the liquid from swab. Dispose of the used swab in accordance with your biohazard waste disposal protocol.



4. Press the nozzle cap tightly onto the tube.
5. Testing and interpreting in accordance with the instruction accompanying the STANDARD COVID-19 Ag products.



■ Test procedure

1. Test should be run following the instructions for use provided by the test kit for unknown specimens.
2. Results should be determined in the same manner as used for unknown specimens when tested using the test kits.
3. Results may vary among methodologies, among products, and among different lots of the same kit.

INTERPRETATION OF TEST RESULTS

Test should be interpreted following the instructions for use provided by the test kit for unknown specimens.

STORAGE AND STABILITY

Store the STANDARD COVID-19 Ag Control swab at 2-30°C/36-86°F. Kit materials are stable until expiration date printed on the outer box.

WARNINGS AND PRECAUTIONS

1. If there is evidence of microbial contamination in the reconstituted control, discard the control.
2. Wear protective clothing and gloves when handling specimens or reagents.
3. Clean any spillage by immediately and thoroughly wiping up with a suitable disinfectant such as 1% sodium hypochlorite solution.
4. Handle and dispose of all specimens, controls and materials used in testing as though they contain infectious agents.
5. Dispose of any discarded materials in accordance with the requirements of your local waste management authorities. In the event of damage to packaging, contact the local distributor of SD Biosensor, Inc.

LIMITATIONS

1. This product is provided for quality assurance purposes and must not be used for calibration or as primary reference preparations in any test procedure.
2. Adverse storage conditions or use of outdated reagents may produce erroneous results.
3. This product should not be used past the expiration date.
4. Alterations in physical appearance may indicate instability or deterioration of this product. If there is evidence of microbial contamination in this product, discard it.



Manufactured by SD Biosensor, Inc.

Head office : C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA

Manufacturing site : 74, Osongsangmyeong 4-ro, Osong-eup, Heungsdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA



Authorized Representative

MT Promedt Consulting GmbH

Altenhofstrasse 80 66386 St. Ingbert Germany
Phone : +49 6894 581020, Fax : +49 6894 581021

Please contact us for any complaints/inquiries/suggestions via email (sales@sdbiosensor.com), website (www.sdbiosensor.com).

L24COVC11ENR2
Issue date: 2022.05



Reference number



In vitro Diagnostics



Consult Instructions for Use



Contains Sufficient for rnt Tests



To indicate the temperature limitations in which the transport package has to be kept and handled.



Do not re-use



Use by



Batch code



Manufacturer



Date of manufacture



Indicate that you should keep the product dry from sunlight



Fulfils the requirements of Directive 98/79/EC on in vitro diagnostic medical devices