



E003:

Refrigerators and freezers

PQS code:	E003/097
Type of appliance:	Combined icelined refrigerator/waterpacks freezer
Manufacturer's reference:	HBCD-90
Manufactured in:	China; People's Republic of
Company:	Qingdao Haier Biomedical Co., Ltd
Address:	No.280 Fengyuan Road High-tech Zone Qingdao(266111), China; People's Republic of
Telephone:	+86-532-88935593
Email:	wenming.wang@haierbiomedical.com
Web address:	http://www.haiermedical.com

Specifications

Climate zone:	Hot	Min rated ambient temp:	+5°C
Refrigerant:	R600a	Energy source:	Electric Mains 220-240V 50/60Hz
Appliance tested at:	+43°C	Ext dimensions (HxLxD)	872*1128*717mm
Performance at:	+43°C	Fuel and cycle type:	Electric - compression

REFRIGERATOR

FREEZER

Freeze protection: Grade A

Vaccine storage capacity: (Litres)	30 L		
Gross volume (Litres):	42 L	Gross volume:	32 L
Holdover time (hours):	63hrs 48mins	Waterpack freezing capacity:	4.0 KG/24h
Energy consumption, stable running (kWh/24 hours):	0.7 KWh/24 h		16 KG
		Waterpack storage capacity:	
Energy consumption, cool down test (kWh/24 hours):	0.66 KWH/24h	Energy consumption during freezing:	0.97 KWH/24h

Comments:

Accessories: "Keys 2
Handle attachment 1
Refiller 1
Corrosion inhibitors 1
Defroster spade 1"

Spare parts (ref, price): Included in the UNICEF Supply Division spare parts kit:
1 No. Dry filter - Part No. 0074180001
2 No. overload protector - Part No. 0274300008B
1 No. Switch - Part No. 0075050003
1 No. Thermo sensor - Part No. 0274000353A
1 No. Thermo sensor - Part No. 0274000358
2 No. Sticker - Part No. 0270501015C
1 No. Thermostat - Part No. 0274000352A
1 No. Compressor - Part No. 0274000418
1 No. Fan - Part No. 0074000172
2 No. Indicator - Part No. 0075050018
2 No. Displayer - Part No. 0274000357

Other parts are available as spares on request.

Shipping volume: 0.99 m3 **Shipping weight:** 135 kg**Price year** 2021 **Incoterms** FCA**Price** 1-100units USD 1650 >100units USD 1570

Quality standard: - ISO 9001:2008 - **Specification ref.**

Verification report: WB-18-448 **Verification laboratory:** CHEARI

Current PQS status: *pre-qualified: 16 Oct 2018* **Validity until:** *May, 2024*

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.