

**E008: Auto-disable syringe for fixed dose immunization**

**PQS code:** E008/040  
**Description:** AD Syringe 0.5ml  
**Manufacturer's reference:** AD Syringe 0.5ml - Revital CADY  
**Manufactured in:** Kenya  
**Company:** Revital Health Care (EPZ) Limited  
**Address:** LR No. 5025/1239, Takaungu  
80100 Mombasa  
Kenya  
**Telephone:** +254 72241 2900  
**Email:** [info@revitalhcare.com](mailto:info@revitalhcare.com)  
**Web address:** [www.revitalhealthcare.com](http://www.revitalhealthcare.com)

**Specifications**

<b>Graduations:</b>	0.5ml	<b>Quality standard:</b>	ISO 13485,
<b>Syringe material(s):</b>	Polypropylene	<b>Markings:</b>	CE 0123
<b>Fixed needle size:</b>	23G x 1" (0.60 x 25 mm)	<b>Pieces per carton:</b>	3600
<b>AD mechanism:</b>	Plunger Disable	<b>Volume per carton (m3):</b>	0.134/0.112
<b>AD location:</b>	Upon completion of injection	<b>Weight per carton (kg):</b>	6 kg
<b>Number of components:</b>	2 pieces	<b>Minimum order:</b>	860 000
<b>Other needle options:</b>	Yes	<b>Incoterms:</b>	FCA
<b>Primary packaging:</b>	Blister pack	<b>Prices per unit:</b>	<i>Per unit: 0.06 €</i>
<b>Year base price:</b>	2008		

**Comments:** Pieces per carton 4,000/3600  
Shipper cartons of 8600 boxes of 100 pieces  
Note: this syringe is to be used with caution by health care workers as per instructions of use.

**Current PQS status:** *pre-qualified ; 08 Oct 2009*

**Valid until :** *May 2024*

**Note:** If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased