

**E003:** Refrigerators and freezers

PQS code: E003/002

Type of appliance: Vaccine/waterpacks freezer

Manufacturer's reference: HBD 116

Manufactured in: China; People's Republic of
Company: Qingdao Haier Biomedical Co., Ltd

Address: No.280 Fengyuan Road

High-tech Zone Qingdao(266111),

2111gua0(200111),

China; People's Republic of

*Telephone:* +86-532-88935593

Email: wenming.wang@haierbiomedical.com

Web address: http://www.haiermedical.com

**Specifications** 

Climate zone: Hot Min rated ambient temp: N/A

Refrigerant: R134a Energy source: Electric Mains 220-240V 50/60Hz

Appliance tested at: +43°C Ext dimensions (HxLxD) 82 x 67 x 63 cm

Performance at: +43°C Fuel and cycle type: Electric - compression

REFRIGERATOR FREEZER Freeze protection: N/A

Vaccine storage capacity: no

(Litres)

Gross volume (Litres): none Gross volume: 121 L

Holdover time (hours): 2 hours 29 mn Waterpack freezing capacity: 12 kg/24 h

Energy consumption, stable 0.38 KWh / 24 h

running (kWh/24 hours):

 $$136\times0.6\ L$$  Waterpack storage capacity:

Energy consumption, cool Energy consumption during 3.77 kWh/24h

down test (kWh/24 hours): freezing:

Comments:

Accessories: Instruction manual; Keys; Tray

Spare parts (ref, price): Included in the UNICEF Supply Division spare parts kit:

1 No. Compressor - Part No. 0070701331 1 No. Thermostat - Part No. 0074091494E 1 No. Dry filter - Part No. 0074180002A 1 No. Fan - Part No. 0075030020

1 No. Fan - Part No. 0075030020 1 No. Switch - Part No. 0075050003

1 No. Thermo sensor - Part No. 0074091494C

Other parts are available as spares on request.

Shipping volume:0.5 m3Shipping weight:79 kgPrice year2021IncotermsEXW

**Price** 1-100units USD 550 >100units USD 530

Quality standard: - ISO 13485 - ISO 14001 - Specification ref.

 Verification report:
 WB-09-145
 Verification laboratory:
 CHEARI

Current PQS status: pre-qualified: 24 Sep 2009 Validity until: May, 2024

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.