

**E003:** Refrigerators and freezers

PQS code: E003/131

Type of appliance: Vaccine/waterpacks freezer

Manufacturer's reference: HBD265

Manufactured in:China; People's Republic ofCompany:Qingdao Haier Biomedical Co., Ltd

Address: No.280 Fengyuan Road

High-tech Zone Qingdao(266111),

China; People's Republic of

1647 x 717 x 940 mm

*Telephone:* +86-532-88935593

Email: wenming.wang@haierbiomedical.com

Web address: http://www.haiermedical.com

Ext dimensions (HxLxD)

**Specifications** 

Climate zone: Hot Min rated ambient temp: 0°C

Refrigerant: R600a Energy source: Electric Mains 220-240V 50/60Hz

Appliance tested at:

Performance at: +43°C Fuel and cycle type: Electric - compression

REFRIGERATOR FREEZER Freeze protection: N/A

Vaccine storage capacity:

(Litres)

Gross volume (Litres): Gross volume: 265L

Holdover time (hours): 11hours 25min Waterpack freezing capacity:

Energy consumption, stable 1.42 KWh/24 h

running (kWh/24 hours):

Waterpack storage capacity:

Energy consumption, cool Energy consumption during 1.40 KWh/24 h

down test (kWh/24 hours): freezing:

211 L

"1. Display: Solar LED Temperature Display

2. Low power consumption"

Comments:

Accessories: 1. Instructions 2. Key 3. Handle component 4. Basket 5. Alcohol thermometer

Spare parts (ref, price): "1 No. Dry filter Part-0074180001

2 No. overload protector 0274300008B

1 No. Switch 0075050003

1 No. Thermo sensor-0274000357A

1 No. Thermostat with sensor 0274000352A 1 No. Compressor-0060705994A

1 No. Compressor-006070591 No. Fan 00740001722 No. Indicator 0075050018

2 No. Displayer 0274000357 1 No. Sticker-0270501015B"

**Shipping volume:** 1.4cbm **Shipping weight:** 160kg

Price year 2023 Incoterms

Price

Quality standard: - ISO 9001:2008 - Specification ref.

Verification report: WB-22-0203 Verification laboratory: CHEARI

Current PQS status: pre-qualified: 23 Mar 2023 Validity until: May, 2024