

**E008: Auto-disable syringe for fixed dose immunization**

**PQS code:** E008/065  
**Description:** AD Syringe 0.1 mL  
**Manufacturer's reference:** Oneject II AD Syringe 0.1 mL  
**Manufactured in:** Indonesia  
**Company:** PT Oneject Indonesia  
**Address:** Sentul Industrial Estate  
Jl.Olympic Raya Kav B9  
Bogor 16810,  
Indonesia

**Telephone:** +62 21 879 17 422

**Email:** [marketing@oneject.co.id](mailto:marketing@oneject.co.id)

**Web address:** [www.oneject.co.id](http://www.oneject.co.id)

**Specifications**

<b>Graduations:</b>	0.1ml	<b>Quality standard:</b>	ISO 9001, ISO 13485,
<b>Syringe material(s):</b>	Polypropylene	<b>Markings:</b>	CE mark
<b>Fixed needle size:</b>	27G x 3/8" (0.40 x 10 mm)	<b>Pieces per carton:</b>	1600
<b>AD mechanism:</b>	Locked / broken plunger	<b>Volume per carton (m3):</b>	0.09 m3/carton
<b>AD location:</b>	Upon completion of injection	<b>Weight per carton (kg):</b>	10.6 Kg
<b>Number of components:</b>		<b>Minimum order:</b>	1600
<b>Other needle options:</b>	Yes	<b>Incoterms:</b>	FOB
<b>Primary packaging:</b>	Blister pack	<b>Prices per unit:</b>	<i>Upon request to manufacturer</i>
<b>Year base price:</b>	2015		

**Comments:**

**Current PQS status:** *pre-qualified ; 12 Sep 2015*

**Valid until :** *May 2024*

**Note:** If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased