

**E003:****Refrigerators and freezers**

<b>PQS code:</b>	<b>E003/133</b>
<b>Type of appliance:</b>	Icelined refrigerator
<b>Manufacturer's reference:</b>	I425H120
<b>Manufactured in:</b>	India
<b>Company:</b>	Western Refrigeration Private Limited
<b>Address:</b>	7/B, Pannalal Silk Mill Compound, L.B.S. Marg, Bhandup (West), Mumbai 400078. India
<b>Telephone:</b>	+91 22 - 67161200
<b>Email:</b>	<a href="mailto:paresh@westernequipments.com">paresh@westernequipments.com</a>
<b>Web address:</b>	<a href="https://www.westernequipments.com/">https://www.westernequipments.com/</a>

**Specifications**

<b>Climate zone:</b>	Multi-zone	<b>Min rated ambient temp:</b>	+10°C
<b>Refrigerant:</b>	#R290	<b>Energy source:</b>	
<b>Appliance tested at:</b>		<b>Ext dimensions (HxLxD)</b>	1320mm (W) X 805mm (D) X 970mm (H)
<b>Performance at:</b>	<b>+43°C</b>	<b>Fuel and cycle type:</b>	Electric - compression

**REFRIGERATOR****FREEZER****Freeze protection: Grade A**

<b>Vaccine storage capacity: (Litres)</b>	192 L
<b>Gross volume (Litres):</b>	253 L
<b>Holdover time (hours):</b>	
<b>Energy consumption, stable running (kWh/24 hours):</b>	1.50 kWh/24 h
<b>Energy consumption, cool down test (kWh/24 hours):</b>	

<b>Gross volume:</b>	
<b>Waterpack freezing capacity:</b>	
<b>Waterpack storage capacity:</b>	
<b>Energy consumption during freezing:</b>	

**Comments:****Accessories:** keys, 30 day Temperature data logger, Voltage Stabilizer and Instruction Manual**Spare parts (ref, price):** 1 No Compressor, 1 No Condensor Fan, 1 No Evaporator Fan, 1 No Digital Controller, 1 No Battery for Digital controller, 1 No Transformer for Digital controller, 1 No 2 Pole ON/ OFF switch, 1 No 30 day temperature data logger, 1 Door Switch, 1 No Application Sensor, 1 Power cord, 1 Buzzer, 1 No Drier, 1 No Stabilizer and Other parts are available as spares on request.**Shipping volume:** 1.24 m3 **Shipping weight:** 135Kg ± 5**Price year:** 2023**Incoterms****Price:** 1 to 99 Units - 1550 USD

100 to 200 Units- 1420 USD

**Quality standard:** - ISO 9001:2015 -**Specification ref.****Verification report:** 4790194057-01-OTHER-S1**Verification laboratory:** UL India Private Limited**Current PQS status:** pre-qualified: 07 Aug 2023**Validity until:** May, 2024

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.