



E003:

Refrigerators and freezers

PQS code:	E003/088
Type of appliance:	Icelined refrigerator
Manufacturer's reference:	HBC-150
Manufactured in:	China; People's Republic of
Company:	Qingdao Haier Biomedical Co., Ltd
Address:	No.280 Fengyuan Road High-tech Zone Qingdao(266111), China; People's Republic of
Telephone:	+86-532-88935593
Email:	wenming.wang@haierbiomedical.com
Web address:	http://www.haiermedical.com

Specifications

Climate zone:		Min rated ambient temp:	+5°C
Refrigerant:	R600a	Energy source:	Electric Mains 220-240V 50/60Hz
Appliance tested at:	+43°C	Ext dimensions (HxLxD)	872*1128*717mm
Performance at:	+43°C	Fuel and cycle type:	Electric - compression

REFRIGERATOR

FREEZER

Freeze protection: Grade A

Vaccine storage capacity: (Litres)	122 Litres
Gross volume (Litres):	150 Litres
Holdover time (hours):	60 hrs 50 mins
Energy consumption, stable running (kWh/24 hours):	0.54kWh/24h
Energy consumption, cool down test (kWh/24 hours):	0.62 kWh/24h

Gross volume:	
Waterpack freezing capacity:	
Waterpack storage capacity:	
Energy consumption during freezing:	

Comments:	1. Display: Solar LED Temperature Display 2. 3 baskets 3. Low power consumption 4. Optional: Remote temperature monitoring device		
Accessories:	Accessory : 1. Instructions 2. Key 3. Handle component 4. Basket 5. Data logger		
Spare parts (ref, price):	Included in the UNICEF Supply Division spare parts kit: 1 No. Dry filter - Part No. 0074180001 2 No. overload protector - Part No. 0274300008B 1 No. Switch - Part No. 0075050003 1 No. Thermo sensor - Part No. 0274000353A 1 No. Thermo sensor - Part No. 0274000358 1 No. Thermostat with sensor - Part No. 0274000352A 1 No. Compressor - Part No. 0274000418 1 No. Fan - Part No. 0074000172 2 No. Indicator - Part No. 0075050018 2 No. Displayer - Part No. 0274000357 1 No. Sticker - Part No. 0270501015A Other parts are available as spares on request.		
Shipping volume:	0.99m3	Shipping weight:	140kg
Price year	2021	Incoterms	
Price	1-100units USD 1070	>100units USD 1020	
Quality standard:	- ISO 9001:2008 - ISO 13485 - ISO 14001 -	Specification ref.	
Verification report:		Verification laboratory:	CHEARI
Current PQS status:	<i>pre-qualified: 09 Jun 2017</i>	Validity until:	<i>May, 2024</i>

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.