

**E008: Auto-disable syringe for fixed dose immunization**

<b>PQS code:</b>	<b>E008/091</b>
<b>Description:</b>	0.1 mL AD syringe
<b>Manufacturer's reference:</b>	0.1 mL AD syringe
<b>Manufactured in:</b>	China; People's Republic of
<b>Company:</b>	Zhejiang Lingyang Medical Apparatus Co., Ltd
<b>Address:</b>	Bai Shuiyang 317031 Linhai Peoples's Republic of China
<b>Telephone:</b>	+86-13676623978
<b>Email:</b>	<a href="mailto:lingyangmedical@163.com">lingyangmedical@163.com</a>
<b>Web address:</b>	<a href="http://www.ly-medical.com">www.ly-medical.com</a>

**Specifications**

<b>Graduations:</b>	0.1ml	<b>Quality standard:</b>	ISO 13485,
<b>Syringe material(s):</b>	Polypropylene	<b>Markings:</b>	-----
<b>Fixed needle size:</b>	26G x 3/8" (0.45 x 10 mm)	<b>Pieces per carton:</b>	4000
<b>AD mechanism:</b>		<b>Volume per carton (m3):</b>	0.138
<b>AD location:</b>	Upon completion of injection	<b>Weight per carton (kg):</b>	16.6
<b>Number of components:</b>	3 pieces	<b>Minimum order:</b>	4,000
<b>Other needle options:</b>	Yes	<b>Incoterms:</b>	FOB
<b>Primary packaging:</b>	Blister pack	<b>Prices per unit:</b>	<i>Price available on request</i>
<b>Year base price:</b>	2022		

**Comments:****Current PQS status:** *pre-qualified ; 17 May 2022***Valid until :** *May 2024***Note:** If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased