



E003: Refrigerators and freezers



Product name:	HBD265
IMD-PQS Code:	E003-131
Status:	Prequalified
Date of acceptance:	23 Mar 2023
Appliance type:	E003
Product description:	Combined icelined refrigerator/waterpacks freezer
Manufacturer:	Qingdao Haier Biomedical Co., Ltd
Manufacturers Reference:	HBD265
Country of Manufacture:	China
Address:	No.280 Fengyuan Road High-tech Zone Qingdao Shandong China
Telephone:	+86-15066227057
Email:	tianhui.yu@haierbiomedical.com
Website address:	http://www.haiermedical.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Climate zone	Hot	Minimum rated ambient temperature (°C)	0.00
Hold over time (hours:mn)	11:25	External dimensions length (mm)	940.00
External dimensions width (mm)	717.00	External dimensions height (mm)	1647.00
Performance at" (°C)	43.00	Fuel and refrigeration cycle	Electric - compression
Energy source	Electric Mains 220-240V 50/60Hz	Energy Consumption During Freezing (kWh/24h)	1.4000
Energy Consumption Continuous Power (kWh/24h)	1.4200		

PRODUCT SPECIFICATIONS - ADDITIONAL REFRIGERATOR

Refrigerator vaccine storage capacity(L)	211.000	Refrigerant type	R600a
Solar radiation ref period (kWh/m2/day)	1.00		

PRODUCT SPECIFICATIONS - ADDITIONAL FREEZER

Freezer's gross volume (Litres)	265.00
---------------------------------	--------

Comments	Optional: Remote temperature monitoring device 1. Display: Solar LED Temperature Display 2. Low power consumption
List of accessories	1. Instructions 2. Key 3. Handle component 4. Basket 5. Alcohol thermometer 6. 30 days temperature logger
List of spare parts, Reference & Price	"1 No. Dry filter Part-0074180001 2 No. overload protector 0274300008B 1 No. Switch 0075050003 1 No. Thermo sensor-0274000357A 1 No. Thermostat with sensor 0274000352A 1 No. Compressor-0060705994A 1 No. Fan 0074000172 2 No. Indicator 0075050018 2 No. Displayer 0274000357 1 No. Sticker-0270501015B"

PRICE AND SHIPPING

Shipping volume (m3)	1.4	Shipping weight (kg)	160.00
Price 1	USD 1313	Year of base price	2023

QUALITY STANDARD

Quality Standard	Other	Quality standard (Other)	ISO 9001:2008
Specification Reference	Applicable PQS specification: Applicable PQS VP(s):		
VERIFICATION			
Verification Laboratory	CHEARI	Address	#29 Xia Jie Street, Xuan Wu district Beijing 100053 R. P. de Chine China
Phone	+86 1063162443	Email	gongcx@cheari.com
Verification Report Reference	WB-22-0203		
PRODUCT SITES			
Organization name	Qingdao Haier Biomedical Co., Ltd		
Address	No.280 Fengyuan Road High-tech Zone Qingdao Shandong China		
Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.