



E003: Refrigerators and freezers



Product name:	GHR 150 AC
IMD-PQS Code:	E003-139e
Status:	Prequalified
Date of acceptance:	25 Jan 2024
Appliance type:	E003
Product description:	Icelined refrigerator
Manufacturer:	Godrej & Boyce Manufacturing Company Limited
Manufacturers Reference:	GHR 150 AC
Country of Manufacture:	India
Address:	Plant 11, Pirojshanagar, Vikhroli Mumbai Maharashtra India
Telephone:	+91 22 67966346
Email:	medicalref@godrej.com
Website address:	http://www.godrejappliances/medical
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Climate zone	Hot	Minimum rated ambient temperature (°C)	10.00
Freeze protection	Grade A	Hold over time (hours:mn)	40.22
External dimensions length (mm)	842.00	External dimensions width (mm)	845.00
External dimensions height (mm)	1010.00	Appliance tested at (°C)	43.00
Performance at" (°C)	43.00	Fuel and refrigeration cycle	Electric - compression
Energy source	Electric Mains 220-240V 50/60Hz or Electric Mains 115/60Hz optional	Energy Consumption Cool Down (kWh/24h)	1.6800
Energy Consumption Continuous Power (kWh/24h)	1.7700		

PRODUCT SPECIFICATIONS - ADDITIONAL REFRIGERATOR

Refrigerator vaccine storage capacity(L)	164.500	Refrigerator's gross volume (Litres)	210.000
Refrigerant type	#R290		

Comments	APPROVED FOR EMS LEVEL 3
List of accessories	Baskets, lock and key, Basket Covers, Castors, - EMS 3 - 9994x that complies with " WHO/PQS/E006/EM01-VP.1 " and " WHO/PQS /E006/DL01-VP.2
List of spare parts, Reference & Price	1) Compressor 2) Electronic Controller 3) Condenser 4) Fan Motor 5) Capacitor

PRICE AND SHIPPING

Shipping volume (m3)	1.02	Shipping weight (kg)	133.20
Incoterms	EXW	Price 1	USD 1,150.00 (1-99 units)
Price 2	USD 1,127.00 (100 -199 units)	Price 3	USD 1,104.00 (200+)
Year of base price	2024		

QUALITY STANDARD

Quality Standard	Other	Quality standard (Other)	ISO 9001:2008
-------------------------	-------	---------------------------------	---------------

Specification Reference Applicable PQS specification:
E003/RF03.5 Applicable PQS
VP(s): E003/RF04-VP.4

VERIFICATION

Verification Laboratory	Intertek India	Address	Intertek India Private Limited E-20, Block B1, Mohan Cooperative Industrial Estate Mathura Road New Delhi - 110044 IndiaIndia
Telephone	+91-9820213510	Email	manas.das@intertek.com
Verification Report Reference	CE-JOB-NDA-23-000594-003		

PRODUCT SITES

Organization name	Godrej & Boyce Manufacturing Company Limited
Address	Plant 11, Pirojshanagar, Vikhroli Mumbai Maharashtra India

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.