



E004: Cold boxes and vaccine carriers



Product name:	BK-CB-LR-FF-17L
IMD-PQS Code:	E004-086
Status:	Prequalified
Date of acceptance:	12 Jan 2026
Appliance type:	Cold Box
Product description:	Cold Box
Manufacturer:	Blowkings
Manufacturers Reference:	BK-CB-LR-FF-17L
Country of Manufacture:	India
Address:	53-C Mittal Court Nariman Point Mumbai Maharashtra India
Telephone:	+91-(0)22-22840120
Email:	munjal@blowkings.co.in
Website address:	http://www.blowkings.co.in/
Valid until:	31 May 2027

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine storage capacity (Litre)	17.33	Gross storage volume (Litre)	17.3300
External dimensions length (mm)	858.00	External dimensions height (mm)	500.00
External dimensions width (mm)	600.00	Internal dimensions (length, mm)	624.00
Internal dimensions (height, mm)	270.00	Internal dimensions (width, mm)	372.00
Cold life at +43°C *(hrs:mn)	101 hrs 30 min	Warm life at - 20°C *(hrs:mn)	53 Hour 56 Min
Cool life at +43°C *(hrs:mn)	NA	Weight empty(Kg)	25.0000
Weight fully loaded (Kg)	49.88	Minimum Rated Ambient Temp (°C)	15.00
Coolant packs	0.6 L Cool Packs Type 4	Number coolant packs	20.00
Type coolant packs	Water packs	Model coolant pack	BK 6 PQS E005/012
Coolant packs supplied	Yes		

PRODUCT SPECIFICATIONS - MATERIALS & CONSTRUCTION

Insulation materials	CFC FREE POLYURETHANE	Insulation thickness (mm)	100 MM
Container type	Cold box	External materials	PE
Internal lining materials	PE	Lid type fixing	Fixed hinges

Comments	Freeze-preventive cold box LR
----------	-------------------------------

PRICE AND SHIPPING

Shipping volume (m3)	0.28	Shipping weight (kg)	24.50
Shipping weight per carton(kg)	26.5	Pieces per package	1 Nos
Minimum order	1.00	Price 1	260
Price 2	>10 Pcs 240 US \$	Year of base price	2025
Incoterms	EXW		

QUALITY STANDARD

Quality Standard	ISO 9001: 2000;ISO 13485	Specification Reference	WHO_PQS_E004/CB05.4
------------------	--------------------------	-------------------------	---------------------

VERIFICATION

Verification Laboratory	0013X00003cPkp6QAC	Verification Report Reference	PQS Test Report No. CE-JOB-NDA-24-000406-003
Current PQS status:	Prequalified	Valid until:	31 May 2027

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.