



E006: Temperature monitoring devices



Product name:	Blulog
IMD-PQS Code:	E006-048
Status:	Prequalified
Date of acceptance:	04 Sep 2018
Appliance type:	E006
Product description:	Remote Temperature Monitoring Device
Manufacturer:	Blulog Sp. z o.o
Manufacturers Reference:	Blulog
Country of Manufacture:	Poland
Address:	Konarszewska 4 60-101 Poznan Poland
Telephone:	+48606641176
Email:	robert.neumann@blulog.eu
Website address:	http://blulog.eu/en/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Mode of operation	Electronic	Minimum logging intervals (mins)	1.00
Maximum logging intervals (mins)	120.00	Sensor type(s)	Digital Temperature sensor
Maximum sensor lead length (metres)	3.00	Maximum number of sensor channels	2.00
External dimensions height (mm)	5.00	External dimensions length (mm)	85.00
External dimensions width (mm)	55.00	Activation mechanism	Press buttons
Material	metal, plastic	Fixing mechanism	Adhesive
CE marking	Yes	Minimum working temperature (°C)	-40.00
Maximum working temperature (°C)	60.00		

MECHANICAL AND ELECTRONIC DEVICES ONLY

Accuracy (+/- °C)	0.50	Power source	Non-replaceable battery
Battery type	Lithium non-replaceable	Battery shelf life (year)	3 years
Activated battery life	3 years	Alarm type	Visual
Memory size (Mb)	0.0000080000	Memory capacity (readings)	5000.00
Programmability	Other	Programmability (Other)	temperature treshold, alert delay, start delay, measurement interval, alert sending
User interface	Web application, Adroid application, windows softw	Display medium	Lenovo Tablet
Recording medium	Electronic	Ingress protection rating IP	IP 67

PRICE AND SHIPPING

Shipping volume (m3)	0.005	Weight per Package (Kg)	0.0300
Units per package	5.00	Weight per Unit (Kg)	1.7000
Minimum order	5.00	Incoterms	EXW
Price 1	1-50 units: 210USD / Tablet + USB hub + holder	Price 2	50-500 units: 200USD / Tablet + USB hub + holder
Price 3	500 - 1000 units: 190USD / Tablet + USB hub + hold	Year of base price	2018

QUALITY STANDARD

Quality Standard	Other	Quality standard (Other)	ISO 9001:2008
Specification Reference	Applicable PQS specification: E006/TR03.2 Applicable PQS VP(s): E006/TR03-VP.2		

VERIFICATION

Verification Laboratory	CEMAFROID SAS	Address	5 Avenue des Pres CS 20029 F94 266 Fresnes Cedex FranceFrance
Telephone	+33 1 49 84 84 84	Email	gerald.cavalier@cemafroid.fr
Verification Report Reference	EN 12830		

PRODUCT SITES

Organization name	Blulog Sp. z o.o
Address	Konarzewska 4 60-101 Poznan Poland

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.