

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization

Product name: MEDECO® INJECT AD 0.5 ml

IMD-PQS Code:E008-004Status:PrequalifiedDate of acceptance:25 Sep 2005

Appliance type: E008

Product description: AD syringe 0.5 ml

Manufacturer: Abu Dhabi Medical Devices Co. L.L.C

Manufacturers Reference: MEDECO® INJECT AD 0.5 ml

Country of Manufacture: United Arab Emirates

Address: MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124

ABU DHABI United Arab Emirates

Telephone: +971-2-6199100
Email: info@admd.ae
Website address: http://www.admd.ae
Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL) 0.50 Graduations (ml) 0.50

MarkingsCE mark 1639RUP mechanismvalve in the syringe hub

Packaging & Sterility Yes Other needle size options Yes

Assurance

Units per package 1300.00 AD Number pieces 3 pieces

Auto-disabling at: Start of injection Needle Size 23Gx1" (0.60 x 25 mm)

AD Sub Category Simple AD AD Primary packaging Blister pack

Comments available in two carton sizes, 1,400pcs and 3,000 pcs Minimum order (units): 1400 or 3000

Shipping volume (m3): 1,400 pcs 0.055 m3, 3,000 0.112m3 Weight per package (kg): 1,400 pcs 5.7

kg, 3,000 11.1kg

PRICE AND SHIPPING

Pieces per package 1400 Price 1 100 000 units and more: 0.054

US\$ per piece

Year of base price 2016 Incoterms EXW

QUALITY STANDARD

Quality Standard ISO 13485;Other Quality standard (Other) ISO 9001; ISO 7886-3

Specification Reference Applicable PQS specification:

ISO 7886-3 Applicable PQS

VP(s): ISO 7886-3

PRODUCT SITES

Organization name Abu Dhabi Medical Devices Co. L.L.C

Address MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124

ABU DHABI

United Arab Emirates

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.