

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization



IMD-PQS Code:E008-040Status:PrequalifiedDate of acceptance:10 Aug 2009

Appliance type: E008

Product description: AD Syringe 0.5ml

Manufacturer:Revital Health Care (EPZ) LimitedManufacturers Reference:AD Syringe 0.5ml - Revital CADY

Country of Manufacture: Kenya

Address: P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239,

Takaungu 80100 Mombasa Kenya

Kenya

Telephone: +254722412900

Email: qa@rhcare-epz.com

Website address: www.revitalhcare.com

Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL) 0.50 Graduations (ml) 0.50

MarkingsCE 0123RUP mechanismPlunger Disables/breaks

Packaging & Sterility Yes Other needle size options Yes

Assurance

Units per package 100.00 AD Number pieces 2 pieces

Auto-disabling at:End of injectionNeedle Size23Gx1" (0.60 x 25 mm)AD Sub CategorySimple ADAD MechanismLocked / Broken Plunger

AD Primary packaging Blister pack

Comments Pieces per carton 4,000/3600 Shipper cartons of 8600 boxes of 100 pieces Note: this syringe is to

be used with caution by health care workers as per instructions of use. Shipping volume (m3):

0.134/0.112

PRICE AND SHIPPING

 Shipping volume (m3)
 0.134
 Weight per package (Kg)
 6.0000

 Pieces per package
 3600
 Minimum order (unit)
 860000.00

 Price 1
 Per unit: 0.06 €
 Year of base price
 2008

Incoterms FCA

QUALITY STANDARD

Quality Standard ISO 13485 **Specification Reference** Applicable PQS specification:

ISO 7886-3 Applicable PQS

VP(s): ISO 7886-3

PRODUCT SITES

Organization name Revital Health Care (EPZ) Limited

Address P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya

Kenya

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.