

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization



IMD-PQS Code:E008-041Status:PrequalifiedDate of acceptance:11 Dec 2009

Appliance type: E008

Product description: AD syringe 0.5ml

Manufacturer: Anhui Tiankang Medical Technology Co., Ltd

Manufacturers Reference: AD syringe 0.5ml

Country of Manufacture: China

Address: No. 228, Weiyi Road Economic Development Zone

Other needle size options

Tianchang Anhui China

Telephone: +86 550 7308592

Email: marshall988@163.net

Website address: http://www.tkmedical.com

Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL) 0.50 Graduations (ml) 0.50

MarkingsCE markRUP mechanismClip blocking plunger

Packaging & Sterility Yes

Assurance

Units per package 100.00 AD Number pieces 3 pieces

Auto-disabling at: Start of injection Needle Size 23Gx1" (0.60 x 25 mm)

AD Sub Category Simple AD AD Primary packaging Blister pack

Comments Cartons of 7680 boxes of 100 pieces Shipping volume (m3): Container 28 m3

PRICE AND SHIPPING

Shipping volume (m3)28.0Weight per package (Kg)3840.0000Pieces per package7680Minimum order (unit)768000.00

Price 1 Up to 768 000 units: 0.0405 € Price 2 Up to 1 590 000 units: 0.0378 €

Year of base price 2008 Incoterms FCA

QUALITY STANDARD

Quality Standard ISO 13485;Other Quality standard (Other) ISO 7886-3

Specification Reference Applicable PQS specification:

ISO 7886-3 Applicable PQS

VP(s): ISO 7886-3

PRODUCT SITES

Organization name Anhui Tiankang Medical Technology Co., Ltd

Address No. 228, Weiyi Road Economic Development Zone

Tianchang Anhui China

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.