

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization



Product name: MEDECO INJECT AD 0.1ml BCG

IMD-PQS Code:E008-070Status:PrequalifiedDate of acceptance:06 Jan 2017

Appliance type: E008

Product description: 0.1ml AD BCG syringe

Manufacturer: Abu Dhabi Medical Devices Co. L.L.C

Manufacturers Reference: MEDECO INJECT AD 0.1ml BCG

Country of Manufacture: United Arab Emirates

Address: MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124

ABU DHABI United Arab Emirates

Telephone: +971-2-6199100
Email: info@admd.ae
Website address: http://www.admd.ae
Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL) 0.10 Graduations (ml) 0.10

MarkingsCE markSyringe material(s)Polypropylene

RUP mechanism One way valve auto-disable Packaging & Sterility Yes

mechanism Assurance

Other needle size options Yes AD Number pieces 3 pieces

Auto-disabling at: Start of injection Needle Size 27Gx3/8" (0.40 x 10 mm)

AD Primary packaging Blister pack

Comments Available in two carton sizes 1400 Pcs or 3000 Pcs Shipping volume (m3): 0.053 M3 for 1400 Pcs

pack. 0.112 M3 for 3000 Pcs Units per package: 1400 or 3000 Weight per package (kg): 5.46 Kg for

1400 Pcs pack and 11.1 Kg for 3000 Pcs

PRICE AND SHIPPING

Pieces per package1400Minimum order (unit)1400.00Price 10.068 USDYear of base price2017

Incoterms EXW

QUALITY STANDARD

Quality Standard ISO 13485;Other Quality standard (Other) ISO 9001

Specification Reference Applicable PQS specification:

ISO 7886-3 Applicable PQS

VP(s): ISO 7886-3

PRODUCT SITES

Organization name Abu Dhabi Medical Devices Co. L.L.C

Address MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124

ABU DHABI

United Arab Emirates

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.