

## WHO Immunisation Devices - Performance, Quality & Safety

## **E008: Injection devices for immunization**

Hongda 0.05ml AD syringe Product name:

IMD-PQS Code: E008-074 Prequalified Status: Date of acceptance: 07 Aug 2019

Appliance type: E008

Product description: 0.05ml AD syringe

Manufacturer: Jiangxi Hongda Medical Equipment Group Ltd

Manufacturers Reference: Hongda 0.05ml AD syringe

Country of Manufacture:

Address: 39 South Shengi Road Jinxian county, Nanchang City

Jiangxi province China; People's Republic of

+86 791 85628828 Telephone: Email: hongda008@jxhd.cn Website address: http://www.jxhd.cn Valid until: 31 May 2026

## **Specifications**

**PRODUCT SPECIFICATIONS - MAIN** 

Vaccine capacity (mL) 0.05 Graduations (ml) 0.05

**Markings** CE mark Syringe material(s) Polypropylene

Packaging & Sterility Yes Other needle size options Yes

**Assurance** 

**AD Number pieces** 

End of injection 3 pieces Auto-disabling at: **Needle Size** 27Gx3/8" (0.40 x 10 mm) **AD Primary packaging** Blister pack

Comments Available in cartons of 1,500 & 2,000 Units per package: 1,500/2,000

PRICE AND SHIPPING

Shipping volume (m3) 19.0000 0.097954 Weight per package (Kg) Pieces per package 1500 Minimum order (unit) 1000000.00

Price 1 available on request Year of base price 2019

FOB Incoterms

**QUALITY STANDARD** 

Applicable PQS specification: **Quality Standard** ISO 13485 Specification Reference

7886-3 Applicable PQS VP(s):

**PRODUCT SITES** 

Organization name Jiangxi Hongda Medical Equipment Group Ltd

**Address** 39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of

China

**Current PQS status:** Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.