

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization

Product name: **KOJAK SELINGE 0.2 ml**

IMD-PQS Code: E008-092 Prequalified Status: Date of acceptance: 17 May 2022

Appliance type: E008

Product description: AD Syringe 0.2 ml

Manufacturer: Hindustan Syringes & Medical Devices Ltd.

Manufacturers Reference: KOJAK SELINGE 0.2 ml

Country of Manufacture:

Address: 174, 178/25 Ballabgarh

Faridabad Haryana India

Telephone: +91 98 102 90320

Email: regulatory@hmdhealthcare.com

Website address: www.hmdhealthcare.com

Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

RUP mechanism 0.20 Locked / Broken Plunger Vaccine capacity (mL)

Other needle size options Packaging & Sterility Yes Yes

Assurance 2000.00 Units per package **AD Number pieces** 3 pieces

Auto-disabling at: Start of injection **Needle Size** 23Gx1" (0.60 x 25 mm)

AD Primary packaging Blister & ribbon pack

Average deadspace 0.021mL Comments

PRICE AND SHIPPING

0.07 9.8100 Shipping volume (m3) Weight per package (Kg) Pieces per package 2000 Minimum order (unit) 20000.00 Price 1

USD 4.70/100 Pcs Year of base price 2025

Incoterms **FCA**

QUALITY STANDARD

Quality Standard ISO 13485;Other **Quality standard (Other)** ISO 9001

Specification Reference Applicable PQS specification:

> ISO 7886-3 Applicable PQS VP(s):

PRODUCT SITES

Organization name Hindustan Syringes & Medical Devices Ltd.

Address 174, 178/25 Ballabgarh

Faridabad Haryana India

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.