

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization

Product name: **KOJAK SELINGE 0.25 ml**

IMD-PQS Code: E008-095 Prequalified Status: 06 Feb 2022 Date of acceptance:

Appliance type: E008

Product description: 0.25 ml AD syringe

Manufacturer: Hindustan Syringes & Medical Devices Ltd.

Manufacturers Reference: KOJAK SELINGE 0.25 ml

Country of Manufacture:

Address: 174, 178/25 Ballabgarh

Faridabad Haryana India

Telephone: +91 98 102 90320

Email: regulatory@hmdhealthcare.com

Website address: www.hmdhealthcare.com

Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

0.25 0.25 Vaccine capacity (mL) Graduations (ml) **RUP** mechanism Packaging & Sterility Locked / Broken Plunger Yes

Assurance

Other needle size options Yes **AD Number pieces** 3 pieces

Auto-disabling at: Start of injection **Needle Size** 23Gx1" (0.60 x 25 mm)

AD Primary packaging Blister & ribbon pack

Average deadspace 0.021mL Comments

PRICE AND SHIPPING

0.07 Weight per package (Kg) 9.8100 Shipping volume (m3) Pieces per package 2000 Minimum order (unit) 20000.00 Price 1 USD 4.70/100 Pcs Year of base price 2025

Incoterms **FCA**

QUALITY STANDARD

Quality Standard ISO 13485;Other **Quality standard (Other)** ISO 9001

Specification Reference Applicable PQS specification:

ISO 7886-3 Applicable PQS

VP(s):

PRODUCT SITES

Organization name Hindustan Syringes & Medical Devices Ltd.

Address 174, 178/25 Ballabgarh

Faridabad Haryana India

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.