

WHO Immunisation Devices - Performance, Quality & Safety

E008: Injection devices for immunization

Product name: Revital Cady 0.5ml 3piece Autodisable Syringe

IMD-PQS Code: E008-097 Prequalified Status: 16 Jun 2023 Date of acceptance:

Appliance type: E008

Product description: 0.5mL AD syringe Manufacturer: Revital Health Care (EPZ) Limited

Manufacturers Reference: Revital Cady 0.5ml 3piece Autodisable Syringe

Country of Manufacture:

Address: P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239,

Takaungu Mombasa Mombasa Kenya

Telephone: +254722412900 Email: qa@rhcare-epz.com Website address: www.revitalhcare.com

Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

No image

available

Graduations (ml) Vaccine capacity (mL)

Markings CE mark **RUP** mechanism Plunger disables / breaks

Packaging & Sterility Yes Other needle size options Yes

Assurance

AD Number pieces

Auto-disabling at: Start of injection 3 pieces

Needle Size 23Gx1" (0.60 x 25 mm) **AD Mechanism** Locked / Broken Plunger

AD Primary packaging Blister pack

Comments Units per package: 4000/3600

PRICE AND SHIPPING

Shipping volume (m3) 0.1031 Weight per package (Kg) 11.6000 Pieces per package 4000 Minimum order (unit) 640000.00 Price 1 POA Year of base price 2022 Incoterms FCA

QUALITY STANDARD

ISO 13485 **Quality Standard**

PRODUCT SITES

Organization name Revital Health Care (EPZ) Limited

Address P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu

Mombasa Mombasa Kenya

Current PQS status: Valid until: Prequalified 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.