

# WHO Immunisation Devices - Performance, Quality & Safety



## **E008: Injection devices for immunization**

Product name: REVITAL CADY 0.3ML 2PC AD SYRINGE

IMD-PQS Code:E008-100Status:PrequalifiedDate of acceptance:25 Oct 2023Product description:0.3mL AD syringe

Manufacturer: Revital Health Care (EPZ) Limited

Manufacturers Reference: REVITAL CADY 0.3ML 2PC AD SYRINGE

Country of Manufacture: Kenya

Address: P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239,

Takaungu Mombasa Mombasa Kenya

Telephone: +254722412900

Email: qa@rhcare-epz.com

Website address: www.revitalhcare.com

Valid until: 31 May 2026

### **Specifications**

#### **PRODUCT SPECIFICATIONS - MAIN**

Graduations (ml)0.30MarkingsCE markSyringe material(s)PolypropyleneOther needle size optionsYes

Auto-disabling at: End of injection Needle Size 23Gx1" (0.60 x 25 mm)

AD Mechanism Locked / Broken Plunger AD Primary packaging Blister pack

PRICE AND SHIPPING

Shipping volume (m3)0.112Weight per package (Kg)11.5700Pieces per package3600Minimum order (unit)648000.00Price 1on requestYear of base price2023

Incoterms FCA

#### QUALITY STANDARD

Quality Standard ISO 13485

Current PQS status:PrequalifiedValid until:31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.