



E008: Injection devices for immunization



IMD-PQS Code:	E008-112
Status:	Prequalified
Date of acceptance:	07 May 2025
Product description:	AD syringe 0.5ml
Manufacturer:	PT Oneject Indonesia
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor Jawa Barat Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Syringe material(s)	Polypropylene	RUP mechanism	Locked / Broken plunger
Other needle size options	Yes	Units per package	100.00
AD Number pieces	3	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.0048	Weight per Unit (Kg)	0.6420
Minimum order (unit)	1000.00	Price 1	Upon request to manufacturer
Year of base price	2025	Incoterms	FOB

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	MDSAP
Specification Reference	1. EN ISO 13485 : 2016 (MD 760987) 2. US 21 CFR part 820 3. GMP-MD Indonesia by Ministry of Health		

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.