



E008: Injection devices for immunization



Product name:	ONEJECT II 0.05ml AD syringe
IMD-PQS Code:	E008-129
Status:	Prequalified
Date of acceptance:	17 Aug 2026
Appliance type:	E008
Product description:	AD syringe 0.05ml
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	ONEJECT II 0.05ml AD syringe
Country of Manufacture:	Indonesia
Address:	JL. Anggrek VII No.41, Kawasan Industri Terpadu Indonesia China (KITIC), Kab. Bekasi Jawa Barat Indonesia
Telephone:	+62 21 879 17 422
Email:	pqs.registration@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2027

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE Mark	Syringe material(s)	Polypropylene
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD location	Upon completion of injection
AD Mechanism	Locked / Broken Plunger	AD Primary packaging	Blister pack

Comments	Syringe is compatible with other needle sizes beyond that listed.
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.051	Weight per package (Kg)	8.2300
Pieces per package	100	Minimum order (unit)	16000.00
Price 1	Upon request to manufacturer	Year of base price	2013
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3; Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Jl. Anggrek VII No.41, Kawasan Industri Terpadu Indonesia China (KITIC), Kab. Bekasi Jawa Barat Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2027</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.