

WHO Immunisation Devices - Performance, Quality & Safety

E010: Waste management equipment



Product name: POLYSAFE® 5L
IMD-PQS Code: E010-001
Status: Prequalified
Date of acceptance: 07 Oct 2008
Appliance type: E010

 Product description:
 Safety box 5L

 Manufacturer:
 POLYNOR AS

 Manufacturers Reference:
 POLYSAFE® 5L

Country of Manufacture: Norway

Address: Oslo Science Park Gaustadalléen 21 N- 0349 Oslo Norway

Norway

Telephone: +47 932 59 486

Email: maria@polynor.no

Website address: http://www.polynor.no

Valid until: 31 May 2026

Specifications

Volume (Litre) 5.00 **Temperature Zones** Temperate Capacity (# units) 155.00 Diameter syringe hole (mm) 38.00 Material recycled carton board Colour White Walls thickness (mm) 1.30 Weight fully assembled 300.00 (grams) **Ext.Dimension After** 274.00 **Ext.Dimension After** 155.00

Assembly Height(mm)

Assembly Length(mm)

Ext.Dimension After 121.00 Ext.Dimension Before 5.00

Assembly Width(mm)

Assembly Height(mm)

Ext.Dimension Before 590.00 Ext.Dimension Before 280.00 Assembly Length(mm) Assembly Width(mm)

PRICE AND SHIPPING

Shipping volume (m3)0.02Weight per Unit (Kg)8.0000Pieces per package25 unitsMinimum order25.00Price 1On requestYear of base price2008

Incoterms FCA

QUALITY STANDARD

Quality Standard Other Quality standard (Other) ISO 9001

Specification Reference Applicable PQS specification: E010/SB01.1 Applicable PQS

VP(s): E010/SB0-VP.1

VERIFICATION

Verification LaboratoryForceTechnologyAddressFORCE Technology Brøndby

Park Allé 345 2605 Brøndby

DenmarkDenmark

Telephone +45 43250591 Email njs@force.dk

Verification Report Reference 115-26166.e rev.1

PRODUCT SITES

Organization name POLYNOR AS

Address Oslo Science Park Gaustadalléen 21 N- 0349 Oslo Norway

Norway

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.