



E013: Injection devices for therapeutic purposes



| | |
|---------------------------------|---|
| Product name: | Haious syringe 10ml |
| IMD-PQS Code: | E013-036 |
| Status: | Prequalified |
| Date of acceptance: | 30 Apr 2009 |
| Appliance type: | E013 |
| Product description: | RUP syringe 10ml |
| Manufacturer: | Guangdong Haiou Medical Apparatus Co., Ltd |
| Manufacturers Reference: | Haious syringe 10ml |
| Country of Manufacture: | China |
| Address: | Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China |
| Telephone: | +86-663-2758666 |
| Email: | market01@haiou.net.cn |
| Website address: | http://www.haiou.net.cn |
| Valid until: | 31 May 2026 |

Specifications

PRODUCT SPECIFICATIONS - MAIN

| | | | |
|----------------------------------|--|--|----------------------|
| Intended purpose | Medication variable dose administration; Medication variable dose reconstitution | Restrictive use for | None |
| Graduations (ml) | 1.00 | Needle Size | Other |
| Needle Size (Other) | 21G x 40mm | Needle Fixation | Detached |
| Nominal capacity (ml) | 10.00 | Packaging & Sterility Assurance | Yes |
| Other needle size options | Yes | Plunger aspiration | Yes |
| Number of pieces | 3.00 | RUP mechanism | Plunger break |
| RUP Type | Plunger lock | Location RUP activation | End of injection |
| RUP Needle Fixation | Not Fixed | RUP Needle Sizes | 21G x 1 1/2" (38 mm) |
| RUP Nominal Capacity (ml) | 10.00 | RUP plunger Aspiration | Multiple |
| RUP Primary Packaging | Blister pack | RUP Sub-category | Simple RUP |
| CE marking | Yes | | |

PRICE AND SHIPPING

| | | | |
|-----------------------------|------------|-----------------------------|-----------|
| Shipping volume (m3) | 0.11 | Shipping weight (kg) | 15.50 |
| Pieces per package | 1200 | Minimum order | 100000.00 |
| Price 1 | On request | Year of base price | 2008 |
| Incoterms | FOB | | |

QUALITY STANDARD

| | | | |
|-------------------------|-----------|--------------------------------|---|
| Quality Standard | ISO 13485 | Specification Reference | Applicable PQS specification: ISO 7886-4 Applicable PQS VP(s): ISO 7886-4 |
|-------------------------|-----------|--------------------------------|---|

PRODUCT SITES

| | |
|--------------------------|---|
| Organization name | Guangdong Haiou Medical Apparatus Co., Ltd |
| Address | Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China |

| | | | |
|----------------------------|---------------------|---------------------|--------------------|
| Current PQS status: | <i>Prequalified</i> | Valid until: | <i>31 May 2026</i> |
|----------------------------|---------------------|---------------------|--------------------|

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.