

WHO Immunisation Devices - Performance, Quality & Safety

E013: Injection devices for therapeutic purposes

RUP syringe 3ml



IMD-PQS Code: E013-069 Status: Prequalified Date of acceptance: 12 Jan 2011

E013 Appliance type:

Product description:

Manufacturer: Mitra Rajawali Banjaran, PT Manufacturers Reference: SKIFA RUP syringe 3ml

Country of Manufacture: Indonesia

Address: JL. Raya Banjaran KM16 Bandung, Jawa Barat Indonesia

Indonesia

Telephone: +62 22 594 0151

Email: nurasfiyahqo@gmail.com Website address: http://www.mrb.co.id

Valid until: 31 May 2025

Specifications

PRODUCT SPECIFICATIONS - MAIN

Intended purpose Medication variable dose Restrictive use for None

administration; Medication variable dose reconstitution

Graduations (ml) **Needle Size** Other 0.10 **Needle Size (Other)** 23G x 32mm **Needle Fixation** Detached

Nominal capacity (ml) 3.00 **Packaging & Sterility** Yes

Assurance

Other needle size options Yes Plunger aspiration Yes

Number of pieces 3.00 **RUP** mechanism Plunger locked at end of

injection

RUP Type Plunger lock **Location RUP activation** End of injection **RUP Needle Fixation** Not Fixed **RUP Needle Sizes** 23G x 1 1/4" (30 mm)

RUP Nominal Capacity (ml) 3.00 **RUP plunger Aspiration** Multiple **RUP Primary Packaging** Blister pack **RUP Sub-category** Simple RUP

CE marking No

PRICE AND SHIPPING

0.17 15.00 Shipping volume (m3) Shipping weight (kg) Pieces per package 2000 20000.00 Minimum order Price 1 upon request to manufacturer Year of base price 2011

Incoterms **EXW**

QUALITY STANDARD

Quality Standard Quality standard (Other) ISO 9001 ISO 13485;Other

Specification Reference Applicable PQS specification:

ISO 7886-4 Applicable PQS

VP(s): ISO 7886-4

PRODUCT SITES

Organization name Mitra Rajawali Banjaran, PT

Address JL. Raya Banjaran KM16 Bandung, Jawa Barat Indonesia

Indonesia

Current PQS status: Prequalified Valid until: 31 May 2025

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.