

WHO Immunisation Devices - Performance, Quality & Safety

E013: Injection devices for therapeutic purposes

Product name: RUP syringe 2mL

IMD-PQS Code:E013-105Status:PrequalifiedDate of acceptance:07 Jan 2015

Appliance type: E013

Product description: RUP syringe 2ml

Manufacturer: Sanavita Pharmaceuticals GMBH

Manufacturers Reference: RUP syringe 2mL
Address: Spaldingstr. 110B

Hamburg Hamburg Germany

Telephone: +49 40 730 8853-0

Email: guenter.kunde@sanavita.net

Website address: www.sanavita.net
Valid until: 31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Intended purpose Medication variable dose Restrictive use for None administration; Medication

variable dose reconstitution

Graduations (ml)0.20Needle SizeOtherNeedle Size (Other)21Gx1 ½" (38mm)Needle FixationFixed

Syringe material(s)
Polypropylene
Nominal capacity (ml)
Packaging & Sterility
Yes
Other needle size options
Yes

Packaging & Sterility Ye Assurance

Plunger aspiration Yes Number of pieces 3.00

RUP mechanismPlunger lockRUP TypePlunger lockLocation RUP activationEnd of injectionRUP Needle FixationFixedRUP Needle Sizes21G x 1 1/2" (38 mm)RUP Nominal Capacity (ml)2.00

RUP Needle Sizes 21G x 1 1/2" (38 mm) RUP Nominal Capacity (ml) 2.00

RUP plunger Aspiration Multiple RUP Primary Packaging Blister pack

RUP Sub-category Simple RUP CE marking Yes

PRICE AND SHIPPING

Shipping volume (m3)0.127Shipping weight (kg)13.00Pieces per package3000Minimum order3000.00Price 1Upon request to manufacturerYear of base price2016

Incoterms FCA

QUALITY STANDARD

Quality Standard ISO 13485 **Specification Reference** Applicable PQS specification:

ISO 7886-4 Applicable PQS

VP(s): ISO 7886-4

PRODUCT SITES

Organization name Wuxi Yushou Medical Appliances Co., Ltd

Address No.115 Nongxinhe Road Xishan District

Wuxi city Jiangsu China

Current PQS status: Prequalified Valid until: 31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.