



E008

Single-use Injection Devices

The products listed in this section are auto-disable (AD) syringes used to administer vaccines. WHO strongly recommends the systematic and exclusive use of auto-disable (AD) syringes only (and not disposable and reusable syringes) for either the administration or reconstitution of vaccines.

The correct use of injection devices for immunization is described in the WHO's [Immunization in practice: A practical guide for health staff](#).

Currently, auto-disable syringes are the only equipment that WHO strongly recommends for the administration of injectable vaccines. Performance standards for AD syringes are defined in [ISO 7886-3](#).

Details regarding the products included in each sub-category of E008, as well as guidelines, performance specifications and verification protocols for this category, can be accessed through the drop-down sections below.

E008.1: Auto-disable syringes

WHO-prequalified syringes must comply with [ISO 7886-3 – Sterile hypodermic syringes for single use – Part 3: Auto-disable syringes for fixed-dose immunization](#).

The auto-disable syringe is the only type of single-use equipment that may be used for the administration of vaccines in all settings; it carries the lowest risk of person-to-person transmission of blood-borne pathogens because it cannot be reused. Refer to the [Safety of injections : WHO-UNICEF-UNFPA joint statement on the use of auto-disable syringes in immunization services, 2nd rev-](#).

Compliance with the IMD-PQS Product Specification and Verification Protocol is required (refer to "Category Documentation" below).

E008.2: Standard disposal syringes

WHO strongly recommends the systematic and exclusive use of auto-disable (AD) syringes only (and not disposable and reusable syringes) for either the administration or reconstitution of vaccines in the context of immunization. Refer to the [Joint Policy](#)

[Statement: Promoting the exclusive use of injection safety devices for all immunization activities.](#)

WARNING: There is clear evidence that the reuse of disposable syringes is widespread globally. WHO warns governments and donor agencies that the reuse of standard disposable syringes and needles places the general public at high risk of disease and death. Refer to the [WHO guideline on the use of safety-engineered syringes for intramuscular, intradermal and subcutaneous injections in health care settings.](#)

E008.3: Needle-free syringe injectors

Needle-free jet injectors deliver a sterile, single dose of liquid medication by pressurizing the dose in a chamber from which it is ejected through a small orifice on an auto-disabling syringe with sufficient force to penetrate human tissues. It is intended for clinical use by medical personnel on humans. It may be used to deliver intra-dermal (ID) and/or subcutaneous (SC) and/or intra-muscular (IM) injections.

Single-use and auto-disabling refer to the needle-free syringe/cartridge, not the injector itself.

Compliance with the IMD-PQS Product Specification [PQS performance specification E008/JI01: Single-use auto-disable needle-free syringe injectors](#) and Verification Protocol [PQS type-examination protocol E008/JI01-VP.1: Single-use auto-disable needle-free syringe injectors](#) is required.

E008.4: Disposal of injection equipment

Additional equipment is necessary to ensure the safety of the contaminated waste stream.

Compliance with the IMD-PQS Product Specification [E010/SB01.1: Safety boxes for the collection of used syringes, needles and sharps](#) and Verification Protocol [PQS independent type-testing protocol E010/SB01-VP.1: Safety box for the disposal of used sharps](#) is required.

CATEGORY DOCUMENTATION, GUIDANCE FOR MANUFACTURERS/SUPPLIERS & SUPPORTING INFORMATION FOR USERS

Performance specifications

[PQS performance specification E008/JI01: Single-use auto-disable needle-free syringe injectors](#)

[ISO 7886-3 – Sterile hypodermic syringes for single use – Part 3: Auto-disable syringes for fixed-dose immunization.](#)

[E010/SB01.1: Safety boxes for the collection of used syringes, needles and sharps.](#)

Verification protocols

[PQS type-examination protocol E008/JI01-VP.1: Single-use auto-disable needle-free syringe injectors](#)

[PQS independent type-testing protocol E010/SB01-VP.1: Safety box for the disposal of used sharps](#)

Guidance for manufacturers or suppliers

[Generic guide for the field evaluation of new technologies for WHO PQS prequalification
GENERIC/GUIDE 1.1](#)

[IMD-PQS Guidelines for Prequalification Holders](#)

Supporting information for users

[Guiding principles to ensure injection device security](#)

[Immunization in practice: a practical guide for health staff – 2015 update](#)

[Safety of injections: WHO-UNICEF-UNFPA joint statement on the use of auto-disable syringes in immunization services](#)

[Joint Policy Statement: Promoting the exclusive use of injection safety devices for all immunization activities](#)



E008: Injection devices for immunization



Product name:	MEDECO® INJECT AD 0.5 ml
IMD-PQS Code:	E008-004
Status:	Prequalified
Date of acceptance:	25 Sep 2005
Appliance type:	E008
Product description:	AD syringe 0.5 ml
Manufacturer:	Abu Dhabi Medical Devices Co. L.L.C
Manufacturers Reference:	MEDECO® INJECT AD 0.5 ml
Country of Manufacture:	United Arab Emirates
Address:	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates
Telephone:	+971-2-6199100
Email:	info@admd.ae
Website address:	http://www.admd.ae
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark 1639	RUP mechanism	valve in the syringe hub
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	1300.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	available in two carton sizes, 1,400pcs and 3,000 pcs Minimum order (units): 1400 or 3000 Shipping volume (m3): 1,400 pcs 0.055 m3, 3,000 0.112m3 Weight per package (kg): 1,400 pcs 5.7 kg, 3,000 11.1kg
-----------------	--

PRICE AND SHIPPING

Pieces per package	1400	Price 1	100 000 units and more: 0.054 US\$ per piece
Year of base price	2016	Incoterms	EXW

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001; ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Abu Dhabi Medical Devices Co. L.L.C
Address	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Kojak Seling 0.5ml
IMD-PQS Code:	E008-007
Status:	Prequalified
Date of acceptance:	22 Dec 2004
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	Kojak Seling 0.5ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
RUP mechanism	Locked / broken plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	100.00
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Sub Category	Simple AD
AD Primary packaging	Blister & ribbon pack		
Comments	Option: Two types of packs with 100 units or 200 units		

PRICE AND SHIPPING

Shipping volume (m3)	0.09	Weight per package (Kg)	13.7000
Pieces per package	2000	Minimum order (unit)	20000.00
Price 1	USD 3.5/100 Pcs	Year of base price	2025
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Hindustan Syringes & Medical Devices Ltd.
Address	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	K1 OneJect AD 0.5ml
IMD-PQS Code:	E008-019
Status:	Prequalified
Date of acceptance:	10 Dec 2005
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	K1 OneJect AD 0.5ml
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE 0044	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	1600.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	24Gx1" (0.55 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.06	Weight per package (Kg)	7.7000
Pieces per package	1600	Minimum order (unit)	16000.00
Price 1	Upon request to manufacturer	Year of base price	2005
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Yushou® AD syringe 0.5ml
IMD-PQS Code:	E008-020
Status:	Prequalified
Date of acceptance:	22 Sep 2005
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	Wuxi Yushou Medical Appliances Co., Ltd
Manufacturers Reference:	Yushou® AD syringe 0.5ml
Country of Manufacture:	China
Address:	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China
Telephone:	+86 (510)3777555
Email:	mike@chinasyringe.com
Website address:	http://www.chinasyringe.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Clip locking plunger at start
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	200.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.088	Weight per package (Kg)	11.0000
Pieces per package	3000	Minimum order (unit)	200000.00
Price 1	Price available on request	Year of base price	2005
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Wuxi Yushou Medical Appliances Co., Ltd
Address	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	VanishPoint II® Retractable AD syringe 0.5ml
IMD-PQS Code:	E008-025
Status:	Prequalified
Date of acceptance:	15 Jun 2005
Appliance type:	E008
Product description:	Retractable AD syringe 0.5ml
Manufacturer:	Retractable Technologies, Inc.
Manufacturers Reference:	VanishPoint II® Retractable AD syringe 0.5ml
Country of Manufacture:	China
Address:	511 Lobo Lane, Little Elm, TX 75068 United States of America
Telephone:	+1 972 294 1010
Email:	kduesman@retractable.com
Website address:	www.retractable.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
RUP mechanism	Retraction of needle into barrel by spring	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	100.00
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Sub Category	Retractable
AD Primary packaging	Blister pack		

Comments	The retraction of the needle is a reuse prevention feature (RUP) and a sharp injury Protection (SIP)
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.0432	Weight per package (Kg)	4.5000
Pieces per package	800	Minimum order (unit)	800.00
Price 1	On request	Year of base price	2005
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Retractable Technologies, Inc.
Address	511 Lobo Lane, Little Elm, TX 75068 United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



No image
available

E008: Injection devices for immunization

IMD-PQS Code:	E008-026
Status:	Suspended
Date of acceptance:	14 Jul 2006
Appliance type:	E008

Specifications

Current PQS status:	<i>Suspended</i>	Valid until:
---------------------	------------------	--------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	K1 Oneject 0.05 ml
IMD-PQS Code:	E008-030
Status:	Prequalified
Date of acceptance:	21 May 2007
Appliance type:	E008
Product description:	AD syringe 0.05 ml
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	K1 Oneject 0.05 ml
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE 0044	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	26Gx3/8" (0.45 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.06	Weight per package (Kg)	6.3000
Pieces per package	1600	Minimum order (unit)	16000.00
Price 1	On request	Year of base price	2007
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	YUSHOU® AD syringe 0.1ml
IMD-PQS Code:	E008-032
Status:	Prequalified
Date of acceptance:	01 May 2009
Appliance type:	E008
Product description:	AD syringe 0.1ml
Manufacturer:	Wuxi Yushou Medical Appliances Co., Ltd
Manufacturers Reference:	YUSHOU® AD syringe 0.1ml
Country of Manufacture:	China
Address:	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China
Telephone:	+86 (510)3777555
Email:	mike@chinasyringe.com
Website address:	http://www.chinasyringe.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	RUP mechanism	locking clip, locked trapped piston
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	200.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.088	Weight per package (Kg)	11.0000
Pieces per package	3000	Minimum order (unit)	200000.00
Price 1	On request	Year of base price	2008
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Wuxi Yushou Medical Appliances Co., Ltd
Address	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Yushou® AD syringe 0.05ml
IMD-PQS Code:	E008-033
Status:	Prequalified
Date of acceptance:	01 May 2009
Appliance type:	E008
Product description:	AD syringe 0.05ml
Manufacturer:	Wuxi Yushou Medical Appliances Co., Ltd
Manufacturers Reference:	Yushou® AD syringe 0.05ml
Country of Manufacture:	China
Address:	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China
Telephone:	+86 (510)3777555
Email:	mike@chinasyringe.com
Website address:	http://www.chinasyringe.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	RUP mechanism	locking clip, locked trapped piston
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	200.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.088	Weight per package (Kg)	11.0000
Pieces per package	3000	Minimum order (unit)	200000.00
Price 1	On request	Year of base price	2008
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Wuxi Yushou Medical Appliances Co., Ltd
Address	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.5 ml SKIFA
IMD-PQS Code:	E008-034
Status:	Prequalified
Date of acceptance:	14 Jan 2009
Appliance type:	E008
Product description:	AD syringe 0.5 ml
Manufacturer:	Mitra Rajawali Banjaran, PT
Manufacturers Reference:	AD syringe 0.5 ml SKIFA
Country of Manufacture:	Indonesia
Address:	JL. Raya Banjaran KM16 Bandung, Jawa Barat Indonesia Indonesia
Telephone:	+62 22 594 0151
Email:	nurasfiyahqo@gmail.com
Website address:	http://www.mrb.co.id
Valid until:	31 May 2025

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	locked middle injection -broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	Middle of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Plastic ribbon

PRICE AND SHIPPING

Shipping volume (m3)	0.071	Weight per package (Kg)	9.6400
Pieces per package	2000	Minimum order (unit)	10000.00
Price 1	On request	Year of base price	2008
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES


Organization name	Mitra Rajawali Banjaran, PT
Address	JL. Raya Banjaran KM16 Bandung, Jawa Barat Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2025</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Auto-disable syringe 0.5ml BD SoloShot™ Mini
IMD-PQS Code:	E008-035
Status:	Prequalified
Date of acceptance:	04 Aug 2009
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	Becton Dickinson
Manufacturers Reference:	Auto-disable syringe 0.5ml BD SoloShot™ Mini
Country of Manufacture:	Spain
Address:	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) Spain
Telephone:	+34 974470900
Email:	kathy.cooman@bd.com
Website address:	http://www.bd.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	locking clip; locked trapped piston
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	2 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	Product is available in 100 pc, 200 pc and 240 pcs per shelf pack, which results respectively in 3300 pcs and 4320 pcs per carton Minimum order (units): 3300 / 3600 / 4320 Weight per package (kg): 10.45 / 10.20 kg
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.102	Pieces per package	3300
Price 1	On request	Year of base price	2009
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001, ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES


Organization name	Becton Dickinson
Address	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.1ml BD SoloShot™ Mini
IMD-PQS Code:	E008-036
Status:	Prequalified
Date of acceptance:	04 Aug 2009
Appliance type:	E008
Product description:	AD syringe 0.1ml
Manufacturer:	Becton Dickinson
Manufacturers Reference:	AD syringe 0.1ml BD SoloShot™ Mini
Country of Manufacture:	Spain
Address:	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) Spain
Telephone:	+34 974470900
Email:	kathy.cooman@bd.com
Website address:	http://www.bd.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	RUP mechanism	locking clip; locked trapped piston
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	2 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	Product is available in 100pc per shelf pack, which results respectively in 3300 pcs per carton. Minimum order (units): 3300 / 3600 Weight per package (kg): 10.45 kg / 10.2 kg
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.102	Pieces per package	3300
Price 1	On request	Year of base price	2009
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001, ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES


Organization name	Becton Dickinson
Address	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.05ml BD SoloShot™ Mini
IMD-PQS Code:	E008-037
Status:	Prequalified
Date of acceptance:	04 Aug 2009
Appliance type:	E008
Product description:	AD syringe 0.05ml
Manufacturer:	Becton Dickinson
Manufacturers Reference:	AD syringe 0.05ml BD SoloShot™ Mini
Country of Manufacture:	Spain
Address:	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) Spain
Telephone:	+34 974470900
Email:	kathy.cooman@bd.com
Website address:	http://www.bd.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	RUP mechanism	locking clip; locked trapped piston
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	2 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	Product is available in 100 pcper shelf pack, which results in 3300 pcs per carton. Weight per package (kg): 10.45 kg / 10.2 kg
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.102	Pieces per package	3300
Minimum order (unit)	3300.00	Price 1	On request
Year of base price	2009	Incoterms	FCA

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001, ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Becton Dickinson
Address	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	KOJAK SELINGE 0.5ml
IMD-PQS Code:	E008-039
Status:	Prequalified
Date of acceptance:	15 Apr 2009
Appliance type:	E008
Product description:	AD syringe 0.5 ml
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	KOJAK SELINGE 0.5ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2025

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
RUP mechanism	locked / broken plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	100.00
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Sub Category	Simple AD
AD Primary packaging	Blister & ribbon pack		
Comments	Two types of packs with 100 units or 200 units		

PRICE AND SHIPPING

Shipping volume (m3)	0.09	Weight per package (Kg)	13.7000
Pieces per package	2000	Minimum order (unit)	20000.00
Price 1	USD 3.5/100 Pcs	Year of base price	2025
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Hindustan Syringes & Medical Devices Ltd.
Address	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India India

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2025</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD Syringe 0.5ml - Revital CADY
IMD-PQS Code:	E008-040
Status:	Prequalified
Date of acceptance:	10 Aug 2009
Appliance type:	E008
Product description:	AD Syringe 0.5ml
Manufacturer:	Revital Health Care (EPZ) Limited
Manufacturers Reference:	AD Syringe 0.5ml - Revital CADY
Country of Manufacture:	Kenya
Address:	P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya
Telephone:	+254722412900
Email:	qa@rhcare-epz.com
Website address:	www.revitalhcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE 0123	RUP mechanism	Plunger Disables/breaks
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	2 pieces
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Mechanism	Locked / Broken Plunger
AD Primary packaging	Blister pack		

Comments	Pieces per carton 4,000/3600 Shipper cartons of 8600 boxes of 100 pieces Note: this syringe is to be used with caution by health care workers as per instructions of use. Shipping volume (m3): 0.134/0.112
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.134	Weight per package (Kg)	6.0000
Pieces per package	3600	Minimum order (unit)	860000.00
Price 1	Per unit: 0.06 €	Year of base price	2008
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Revital Health Care (EPZ) Limited
Address	P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.5ml
IMD-PQS Code:	E008-041
Status:	Prequalified
Date of acceptance:	11 Dec 2009
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	Anhui Tiankang Medical Technology Co., Ltd
Manufacturers Reference:	AD syringe 0.5ml
Country of Manufacture:	China
Address:	No. 228, Weiyi Road Economic Development Zone 239300 Tianchang China
Telephone:	+86 550 7308592
Email:	marshall988@163.net
Website address:	http://www.tkmedical.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Clip blocking plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	Cartons of 7680 boxes of 100 pieces Shipping volume (m3): Container 28 m3
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	28.0	Weight per package (Kg)	3840.0000
Pieces per package	7680	Minimum order (unit)	768000.00
Price 1	Up to 768 000 units: 0.0405 €	Price 2	Up to 1 590 000 units: 0.0378 €
Year of base price	2008	Incoterms	FCA

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Anhui Tiankang Medical Technology Co., Ltd
Address	No. 228, Weiyi Road Economic Development Zone 239300 Tianchang China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.5ml
IMD-PQS Code:	E008-043
Status:	Prequalified
Date of acceptance:	09 Feb 2010
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	Jiangxi Sanxin Medtec Co.,Ltd
Manufacturers Reference:	AD syringe 0.5ml
Country of Manufacture:	China
Address:	999 Fushan Road, Xiaolan Economic Development Zone, Nanchang, Jiangxi, P.C. 330200 China; People's Republic of China
Telephone:	+86 791 85988111
Email:	fqb@sanxin-med.com
Website address:	www.sanxin-med.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Plunger lock
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.1	Weight per package (Kg)	15.0000
Pieces per package	3000	Minimum order (unit)	2000.00
Price 1	On request	Year of base price	2009
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Jiangxi Sanxin Medtec Co.,Ltd
Address	999 Fushan Road, Xiaolan Economic Development Zone, Nanchang, Jiangxi, P.C. 330200 China; People's Republic of China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Kojak AD syringe 0.1ml
IMD-PQS Code:	E008-045
Status:	Prequalified
Date of acceptance:	05 Oct 2011
Appliance type:	E008
Product description:	AD syringe 0.1ml
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	Kojak AD syringe 0.1ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
RUP mechanism	Locked / broken plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	100.00
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	26Gx3/8" (0.45 x 10 mm)	AD Sub Category	Simple AD
AD Primary packaging	Blister pack		

Comments	Two types of packs with 100 units or 200 units
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.09	Weight per package (Kg)	12.2000
Pieces per package	3200	Minimum order (unit)	32000.00
Price 1	USD 4/100 Pcs	Year of base price	2025
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Hindustan Syringes & Medical Devices Ltd.
Address	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India India

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Apple K1 AD syringe
IMD-PQS Code:	E008-048
Status:	Prequalified
Date of acceptance:	11 Feb 2012
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	AMSON VACCINES & PHARMA (PVT) LTD.
Manufacturers Reference:	Apple K1 AD syringe
Country of Manufacture:	Pakistan
Address:	115, Industrial Triangle, Kahuta road, Islamabad Pakistan Pakistan
Telephone:	+92 51 4492022
Email:	naveed@amson.org.pk
Website address:	www.amson.org.pk
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	24Gx1" (0.55 x 25 mm)	AD Sub Category	Simple AD
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.07	Weight per package (Kg)	8.6200
Pieces per package	1200	Minimum order (unit)	100000.00
Price 1	On request to manufacturer	Year of base price	2012
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	Other	Quality standard (Other)	ISO 9001; ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	AMSON VACCINES & PHARMA (PVT) LTD.
Address	115, Industrial Triangle, Kahuta road, Islamabad Pakistan Pakistan

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Yi Xin 0.1ml
IMD-PQS Code:	E008-049
Status:	Prequalified
Date of acceptance:	13 Jan 2013
Appliance type:	E008
Product description:	AD syringe 0.1ml
Manufacturer:	Jiangxi Sanxin Medtec Co.,Ltd
Manufacturers Reference:	Yi Xin 0.1ml
Country of Manufacture:	China
Address:	999 Fushan Road, Xiaolan Economic Development Zone, Nanchang, Jiangxi, P.C. 330200 China; People's Republic of China
Telephone:	+86 791 85988111
Email:	fqb@sanxin-med.com
Website address:	www.sanxin-med.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	RUP mechanism	Plunger lock
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	26Gx3/8" (0.45 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.104	Weight per package (Kg)	11.0000
Pieces per package	3000	Minimum order (unit)	100000.00
Year of base price	2013	Incoterms	FOB

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Jiangxi Sanxin Medtec Co.,Ltd
Address	999 Fushan Road, Xiaolan Economic Development Zone, Nanchang, Jiangxi, P.C. 330200 China; People's Republic of China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Pharmajet Stratis®
IMD-PQS Code:	E008-050
Status:	Prequalified
Date of acceptance:	31 Jan 2013
Appliance type:	E008
Product description:	single-use auto-disable needle-free jet injector
Manufacturer:	PharmaJet, Inc.
Manufacturers Reference:	Pharmajet Stratis®
Country of Manufacture:	United States of America
Address:	400 Corporate Circle, Suite N Golden, Colorado 80401 USA United States of America
Telephone:	+1 303 526-4278
Email:	brent.carpenter@pharmajet.com
Website address:	http://www.pharmajet.com/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	FDA 510(k); CE 556625	RUP mechanism	Plunger break
Packaging & Sterility Assurance	Yes	Other needle size options	No
AD Number pieces	2 pieces	Auto-disabling at:	Start of injection
AD Sub Category	Jet injector	AD Primary packaging	Blister pack

Comments	Single-use auto-disable needle-free jet injector Note: Jet Injectors (single-use auto-disable needle-free syringe injectors) as medical devices (Class II) are regulated by National Regulatory Authorities. A Jet injector prequalified by the WHO PQS programme to deliver medications and vaccines is to be used according to label to deliver ONLY those medications and vaccines that have been approved by the relevant Authorities. Reset station Injector: Reference manufacturer C501-20 Needle free syringe: Reference manufacturer C501-32 & C501-34 Filling adapter: Reference manufacturer C501-43 & C501-46 Injector: Reference manufacturer C501-10 Minimum order (units): 1 set Units per package: Syringes, 250/box; adaptors, 75/box
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.037	Weight per package (Kg)	3.2300
Pieces per package	1	Price 1	Upon request to mfr
Year of base price	2012	Incoterms	EXW

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: E008/JI01.1 Applicable PQS VP(s): E008/JI01-VP.1
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	PharmaJet, Inc.
Address	400 Corporate Circle, Suite N Golden, Colorado 80401 USA United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	JMI AD syringe 0.5ml
IMD-PQS Code:	E008-051
Status:	Prequalified
Date of acceptance:	29 May 2013
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	JMI Syringes & Medical Devices Ltd.
Manufacturers Reference:	JMI AD syringe 0.5ml
Country of Manufacture:	Bangladesh
Address:	Unique Heights 117 Kazi Nazrul Islam Avenue Dhaka-1217 Bangladesh. Bangladesh
Telephone:	+88 01939909332
Email:	hasan.pmd@jmigroup-bd.com
Website address:	http://www.jmisyringe.com/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Plunger lock
Packaging & Sterility Assurance	Yes	Other needle size options	No
Units per package	200.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.133	Weight per package (Kg)	15.6000
Pieces per package	2400	Minimum order (unit)	192000.00
Price 1	Upon request to manufacturer	Year of base price	2013
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	JMI Syringes & Medical Devices Ltd.
Address	Noapara, Chauddagam, Cumilla, Bangladesh 3550, Cumilla, Bangladesh Bangladesh

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	OnejectII 0.05ml AD syringe
IMD-PQS Code:	E008-052
Status:	Prequalified
Date of acceptance:	06 May 2013
Appliance type:	E008
Product description:	AD syringe 0.05ml
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	OnejectII 0.05ml AD syringe
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.051	Weight per package (Kg)	8.2300
Pieces per package	1600	Minimum order (unit)	16000.00
Price 1	Upon request to manufacturer	Year of base price	2013
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	OnejectII 0.5 ml AD syringe
IMD-PQS Code:	E008-053
Status:	Prequalified
Date of acceptance:	06 May 2013
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	OnejectII 0.5 ml AD syringe
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	24Gx1" (0.55 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.052	Weight per package (Kg)	7.6600
Pieces per package	1600	Minimum order (unit)	16000.00
Price 1	Upon request to manufacturer	Year of base price	2013
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Haiou Retractable auto disable syringe 0.5ml
IMD-PQS Code:	E008-056
Status:	Prequalified
Date of acceptance:	29 Jul 2013
Appliance type:	E008
Product description:	Retractable AD syringe 0.5ml
Manufacturer:	Guangdong Haiou Medical Apparatus Co., Ltd
Manufacturers Reference:	Haiou Retractable auto disable syringe 0.5ml
Country of Manufacture:	China
Address:	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China
Telephone:	+86-663-2758666
Email:	market01@haiou.net.cn
Website address:	http://www.haiou.net.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Plunger lock / needle retraction
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Retractable	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.087	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	100000.00
Price 1	Upon request to manufacturer	Year of base price	2013
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Guangdong Haiou Medical Apparatus Co., Ltd
Address	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.05ml
IMD-PQS Code:	E008-057
Status:	Prequalified
Date of acceptance:	29 Jul 2013
Appliance type:	E008
Product description:	AD syringe 0.05ml
Manufacturer:	Anhui Tiankang Medical Technology Co., Ltd
Manufacturers Reference:	AD syringe 0.05ml
Country of Manufacture:	China
Address:	No. 228, Weiye Road Economic Development Zone 239300 Tianchang China
Telephone:	+86 550 7308592
Email:	marshall988@163.net
Website address:	http://www.tkmedical.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	RUP mechanism	Plunger lock by stainless jaw at start of injection
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.069	Weight per package (Kg)	10.0000
Pieces per package	2000	Minimum order (unit)	200000.00
Price 1	Upon request	Year of base price	2013
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Anhui Tiankang Medical Technology Co., Ltd
Address	No. 228, Weiye Road Economic Development Zone 239300 Tianchang China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Auto-disable syringe 0.25ml
IMD-PQS Code:	E008-058
Status:	Prequalified
Date of acceptance:	29 Jul 2013
Appliance type:	E008
Product description:	AD syringe 0.25ml
Manufacturer:	Anhui Tiankang Medical Technology Co., Ltd
Manufacturers Reference:	Auto-disable syringe 0.25ml
Country of Manufacture:	China
Address:	No. 228, Weiye Road Economic Development Zone 239300 Tianchang China
Telephone:	+86 550 7308592
Email:	marshall988@163.net
Website address:	http://www.tkmedical.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.25	Graduations (ml)	0.25
Markings	CE mark	RUP mechanism	Plunger lock by stainless steel jaw at start of injection
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.069	Weight per package (Kg)	10.0000
Pieces per package	2000	Minimum order (unit)	200000.00
Price 1	Upon request	Year of base price	2013
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Anhui Tiankang Medical Technology Co., Ltd
Address	No. 228, Weiye Road Economic Development Zone 239300 Tianchang China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Destroject 0.5ml AD (1000) syringe
IMD-PQS Code:	E008-060
Status:	Prequalified
Date of acceptance:	04 Jan 2014
Appliance type:	E008
Product description:	AD syringe 0.5ml
Manufacturer:	AMSON VACCINES & PHARMA (PVT) LTD.
Manufacturers Reference:	Destroject 0.5ml AD (1000) syringe
Country of Manufacture:	Pakistan
Address:	115, Industrial Triangle, Kahuta road, Islamabad Pakistan Pakistan
Telephone:	+92 51 4492022
Email:	naveed@amson.org.pk
Website address:	www.amson.org.pk
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Plunger lock by cremalliere at start
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	2000.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.097	Weight per package (Kg)	12.5000
Pieces per package	100	Minimum order (unit)	32000.00
Price 1	Upon request	Year of base price	2014
Incoterms	EXW		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	AMSON VACCINES & PHARMA (PVT) LTD.
Address	115, Industrial Triangle, Kahuta road, Islamabad Pakistan Pakistan

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.5mL
IMD-PQS Code:	E008-062
Status:	Prequalified
Date of acceptance:	07 Jan 2015
Appliance type:	E008
Product description:	AD syringe 0.5mL
Manufacturer:	Sanavita Pharmaceuticals GMBH
Manufacturers Reference:	AD syringe 0.5mL
Country of Manufacture:	China
Address:	Spaldingstr. 110B Hamburg Hamburg Germany
Telephone:	+49 40 730 8853-0
Email:	guenter.kunde@sanavita.net
Website address:	www.sanavita.net
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	metal clip	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	3000.00
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Sub Category	Simple AD
AD Mechanism	Metal clip	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.087	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	3000.00
Price 1	Upon request to manufacturer	Year of base price	2015
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Sanavita Pharmaceuticals GMBH
Address	Spaldingstr. 110B Hamburg Hamburg Germany

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.05mL
IMD-PQS Code:	E008-063
Status:	Prequalified
Date of acceptance:	07 Jan 2015
Appliance type:	E008
Product description:	AD syringe 0.05mL
Manufacturer:	Sanavita Pharmaceuticals GMBH
Manufacturers Reference:	AD syringe 0.05mL
Country of Manufacture:	China
Address:	Spaldingstr. 110B Hamburg Hamburg Germany
Telephone:	+49 40 730 8853-0
Email:	guenter.kunde@sanavita.net
Website address:	www.sanavita.net
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	metal clip	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	3000.00
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	27Gx3/8" (0.40 x 10 mm)	AD Sub Category	Simple AD
AD Mechanism	Metal clip	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.087	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	3000.00
Price 1	Upon request to manufacturer	Year of base price	2015
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Sanavita Pharmaceuticals GMBH
Address	Spaldingstr. 110B Hamburg Hamburg Germany

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	AD syringe 0.1mL
IMD-PQS Code:	E008-064
Status:	Prequalified
Date of acceptance:	07 Jan 2015
Appliance type:	E008
Product description:	AD syringe 0.1mL
Manufacturer:	Sanavita Pharmaceuticals GMBH
Manufacturers Reference:	AD syringe 0.1mL
Country of Manufacture:	China
Address:	Spaldingstr. 110B Hamburg Hamburg Germany
Telephone:	+49 40 730 8853-0
Email:	guenter.kunde@sanavita.net
Website address:	www.sanavita.net
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	metal clip	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	3000.00
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	27Gx3/8" (0.40 x 10 mm)	AD Sub Category	Simple AD
AD Mechanism	Metal clip	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.087	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	3000.00
Price 1	Upon request to manufacturer	Year of base price	2015
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Sanavita Pharmaceuticals GMBH
Address	Spaldingstr. 110B Hamburg Hamburg Germany

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Oneject II AD Syringe 0.1 mL
IMD-PQS Code:	E008-065
Status:	Prequalified
Date of acceptance:	09 Dec 2015
Appliance type:	E008
Product description:	AD Syringe 0.1 mL
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	Oneject II AD Syringe 0.1 mL
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	1600.00	Auto-disabling at:	End of injection
Needle Size	27Gx3/8" (0.40 x 10 mm)	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.09	Weight per package (Kg)	10.6000
Pieces per package	1600	Minimum order (unit)	1600.00
Price 1	Upon request to manufacturer	Year of base price	2015
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Haiou Retractable auto disable syringe 0.05ml
IMD-PQS Code:	E008-068
Status:	Prequalified
Date of acceptance:	06 Jan 2017
Appliance type:	E008
Product description:	Retractable 0.05ml Auto Disable Syringe
Manufacturer:	Guangdong Haiou Medical Apparatus Co., Ltd
Manufacturers Reference:	Haiou Retractable auto disable syringe 0.05ml
Country of Manufacture:	China
Address:	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China
Telephone:	+86-663-2758666
Email:	market01@haiou.net.cn
Website address:	http://www.haiou.net.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	RUP mechanism	Manual retractable
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	27Gx3/8" (0.40 x 10 mm)	AD Sub Category	Retractable
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.094	Weight per package (Kg)	15.0000
Pieces per package	3600	Minimum order (unit)	100000.00
Price 1	On request	Year of base price	2017
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Guangdong Haiou Medical Apparatus Co., Ltd
Address	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



No image available

Product name:	MEDECO INJECT AD 0.05ml BCG
IMD-PQS Code:	E008-069
Status:	Prequalified
Date of acceptance:	06 Jan 2017
Appliance type:	E008
Product description:	0.05mlAD BCG syringe
Manufacturer:	Abu Dhabi Medical Devices Co. L.L.C
Manufacturers Reference:	MEDECO INJECT AD 0.05ml BCG
Country of Manufacture:	United Arab Emirates
Address:	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates
Telephone:	+971-2-6199100
Email:	info@admd.ae
Website address:	http://www.admd.ae
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	One way valve auto disable mechanism	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Primary packaging	Blister pack		
Comments	Available in 1400 Pcs or 3000 Pcs per carton Minimum order (units): 1400 or 3000 Shipping volume (m3): 0.053 M3 for 1400 Pcs pack. 0.112 M3 for 3000 Pcs Weight per package (kg): 5.46 Kg for 1400 Pcs pack and 11.1 Kg for 3000 Pcs		

PRICE AND SHIPPING

Pieces per package	1400	Price 1	0.068 USD
Year of base price	2017	Incoterms	EXW

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Abu Dhabi Medical Devices Co. L.L.C		
Address	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates		
Current PQS status:	Prequalified	Valid until:	31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	MEDECO INJECT AD 0.1ml BCG
IMD-PQS Code:	E008-070
Status:	Prequalified
Date of acceptance:	06 Jan 2017
Appliance type:	E008
Product description:	0.1ml AD BCG syringe
Manufacturer:	Abu Dhabi Medical Devices Co. L.L.C
Manufacturers Reference:	MEDECO INJECT AD 0.1ml BCG
Country of Manufacture:	United Arab Emirates
Address:	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates
Telephone:	+971-2-6199100
Email:	info@admd.ae
Website address:	http://www.admd.ae
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	One way valve auto-disable mechanism	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Primary packaging	Blister pack		

Comments	Available in two carton sizes 1400 Pcs or 3000 Pcs Shipping volume (m3): 0.053 M3 for 1400 Pcs pack. 0.112 M3 for 3000 Pcs Units per package: 1400 or 3000 Weight per package (kg): 5.46 Kg for 1400 Pcs pack and 11.1 Kg for 3000 Pcs
-----------------	--

PRICE AND SHIPPING

Pieces per package	1400	Minimum order (unit)	1400.00
Price 1	0.068 USD	Year of base price	2017
Incoterms	EXW		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Abu Dhabi Medical Devices Co. L.L.C
Address	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	PharmaJet® Tropis®
IMD-PQS Code:	E008-071
Status:	Prequalified
Date of acceptance:	06 Apr 2018
Appliance type:	E008
Product description:	Single Use auto-disable needle-free jet injector
Manufacturer:	PharmaJet, Inc.
Manufacturers Reference:	PharmaJet® Tropis®
Country of Manufacture:	United States of America
Address:	400 Corporate Circle, Suite N Golden, Colorado 80401 USA United States of America
Telephone:	+1 303 526-4278
Email:	brent.carpenter@pharmajet.com
Website address:	http://www.pharmajet.com/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Markings	-----
Packaging & Sterility Assurance	Yes	Auto-disabling at:	End of injection
AD Sub Category	Jet injector	AD Primary packaging	Blister pack
Comments	Single-use auto-disable needle-free jet injector Note: Jet Injectors (single-use auto-disable needle-free syringe injectors) as medical devices (Class II) are regulated by National Regulatory Authorities. A Jet injector prequalified by the WHO PQS programme to deliver medications and vaccines is to be used according to label to deliver ONLY those medications and vaccines that have been approved by the relevant Authorities.		

PRICE AND SHIPPING

Price 1	Contact manufacturer	Year of base price	2018
----------------	----------------------	---------------------------	------

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: E008/JI01.1 Applicable PQS VP(s): E008/JI01-VP.1
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	PharmaJet, Inc.		
Address	400 Corporate Circle, Suite N Golden, Colorado 80401 USA United States of America		
<hr/>			
Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Hongda 0.05ml AD syringe
IMD-PQS Code:	E008-074
Status:	Prequalified
Date of acceptance:	07 Aug 2019
Appliance type:	E008
Product description:	0.05ml AD syringe
Manufacturer:	Jiangxi Hongda Medical Equipment Group Ltd
Manufacturers Reference:	Hongda 0.05ml AD syringe
Country of Manufacture:	China
Address:	39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of China
Telephone:	+86 791 85628828
Email:	hongda008@jxhd.cn
Website address:	http://www.jxhd.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	Syringe material(s)	Polypropylene
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	27Gx3/8" (0.40 x 10 mm)	AD Primary packaging	Blister pack

Comments	Available in cartons of 1,500 & 2,000 Units per package: 1,500/2,000
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.097954	Weight per package (Kg)	19.0000
Pieces per package	1500	Minimum order (unit)	1000000.00
Price 1	available on request	Year of base price	2019
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: 7886-3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Jiangxi Hongda Medical Equipment Group Ltd
Address	39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Apple K-1 3ring AD syringe
IMD-PQS Code:	E008-075
Status:	Prequalified
Date of acceptance:	16 Jul 2019
Appliance type:	E008
Product description:	0.5ml Auto Disable syringe
Manufacturer:	AMSON VACCINES & PHARMA (PVT) LTD.
Manufacturers Reference:	Apple K-1 3ring AD syringe
Country of Manufacture:	Pakistan
Address:	115, Industrial Triangle, Kahuta road, Islamabad Pakistan Pakistan
Telephone:	+92 51 4492022
Email:	naveed@amson.org.pk
Website address:	www.amson.org.pk
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	Locked/broken Plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.07	Weight per package (Kg)	8.6200
Pieces per package	1200	Minimum order (unit)	1000000.00
Price 1	Price available on request	Year of base price	2019
Incoterms	EXW		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: 7883-3 Applicable PQS VP(s):		

PRODUCT SITES

Organization name	AMSON VACCINES & PHARMA (PVT) LTD.
Address	115, Industrial Triangle, Kahuta road, Islamabad Pakistan Pakistan

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Haiou 0.5ml Auto Disable Syringe
IMD-PQS Code:	E008-076
Status:	Prequalified
Date of acceptance:	11 Apr 2019
Appliance type:	E008
Product description:	0.5ml Auto Disable Syringe
Manufacturer:	Guangdong Haiou Medical Apparatus Co., Ltd
Manufacturers Reference:	Haiou 0.5ml Auto Disable Syringe
Country of Manufacture:	China
Address:	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China China
Telephone:	+86-663-2758666
Email:	market01@haiou.net.cn
Website address:	http://www.haiou.net.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	Plunger break	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

Comments	Title: 0.5ml Auto Disbale Syringe Product standard: ISO 7886-3
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.086	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	100000.00
Price 1	on request	Price 2	on request
Year of base price	2019	Incoterms	FOB

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	Guangdong Haiou Medical Apparatus Co., Ltd
Address	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Haiou 0.05ml Auto Disable Syringe
IMD-PQS Code:	E008-078
Status:	Prequalified
Date of acceptance:	11 Oct 2020
Appliance type:	E008
Product description:	0.05ml AD syringe
Manufacturer:	Guangdong Haiou Medical Apparatus Co., Ltd
Manufacturers Reference:	Haiou 0.05ml Auto Disable Syringe
Country of Manufacture:	China
Address:	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China
Telephone:	+86-663-2758666
Email:	market01@haiou.net.cn
Website address:	http://www.haiou.net.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.05	Graduations (ml)	0.05
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	Broken plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	27Gx3/8" (0.40 x 10 mm)
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.086	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	100000.00
Price 1	Price available on request	Year of base price	2020
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Guangdong Haiou Medical Apparatus Co., Ltd
Address	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



No image available

Product name:	0.3ml AD syringe
IMD-PQS Code:	E008-080
Status:	Prequalified
Date of acceptance:	31 Dec 2020
Appliance type:	E008
Product description:	0.3ml AD syringe
Manufacturer:	Wuxi Yushou Medical Appliances Co., Ltd
Manufacturers Reference:	0.3ml AD syringe
Country of Manufacture:	China
Address:	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China
Telephone:	+86 (510)3777555
Email:	mike@chinasyringe.com
Website address:	http://www.chinasyringe.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	CE mark
RUP mechanism	AD	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	3000.00
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack
Comments	Average deadspace 0.0056mL		

PRICE AND SHIPPING

Weight per package (Kg)	11.0000	Pieces per package	3000
Minimum order (unit)	200000.00	Price 1	On request to manufacturer
Year of base price	2020	Incoterms	FOB

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: 7883-3 Applicable PQS VP(s): 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Wuxi Yushou Medical Appliances Co., Ltd
Address	No.115 Nongxinhe Road Xishan District Wuxi city Jiangsu Province China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	MEDECO INJECT AD 0.3ml
IMD-PQS Code:	E008-081
Status:	Prequalified
Date of acceptance:	13 Jan 2021
Appliance type:	E008
Product description:	0.3ml AD syringe
Manufacturer:	Abu Dhabi Medical Devices Co. L.L.C
Manufacturers Reference:	MEDECO INJECT AD 0.3ml
Country of Manufacture:	United Arab Emirates
Address:	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates
Telephone:	+971-2-6199100
Email:	info@admd.ae
Website address:	http://www.admd.ae
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	CE mark
Syringe material(s)	Polypropylene	RUP mechanism	Valve in the syringe hub
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	3000.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

Comments	Graduation 0.3ml. Average deadspace 0.0231mL
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.1	Weight per package (Kg)	11.5000
Pieces per package	3000	Minimum order (unit)	100000.00
Price 1	0.054 USD	Year of base price	2021
Incoterms	EXW		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: 7883-3 Applicable PQS VP(s): 78863-3
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	Abu Dhabi Medical Devices Co. L.L.C
Address	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	0.3mL AD Soloshot Mini
IMD-PQS Code:	E008-082
Status:	Prequalified
Date of acceptance:	29 Jan 2021
Appliance type:	E008
Product description:	0.3mL AD
Manufacturer:	Becton Dickinson
Manufacturers Reference:	0.3mL AD Soloshot Mini
Country of Manufacture:	Spain
Address:	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) Spain
Telephone:	+34 974470900
Email:	kathy.cooman@bd.com
Website address:	http://www.bd.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	CE mark
RUP mechanism	locking clip; locked trapped piston	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	2 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

Comments	Average deadspace 0.0069mL
-----------------	----------------------------

PRICE AND SHIPPING

Shipping volume (m3)	0.1032	Weight per package (Kg)	10.2890
Pieces per package	200	Minimum order (unit)	3600.00
Price 1	price on request	Year of base price	2021

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s):		

PRODUCT SITES

Organization name	Becton Dickinson
Address	Crtra. Mequinenza (N-211) S/N, 22520, Fraga (Huesca) United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	TKMD 0.3ml AD syringe with fixed needle
IMD-PQS Code:	E008-083
Status:	Prequalified
Date of acceptance:	04 Aug 2021
Appliance type:	E008
Product description:	0.3ml AD syringe
Manufacturer:	Anhui Tiankang Medical Technology Co., Ltd
Manufacturers Reference:	TKMD 0.3ml AD syringe with fixed needle
Country of Manufacture:	China
Address:	No. 228, Weiyi Road Economic Development Zone 239300 Tianchang China
Telephone:	+86 550 7308592
Email:	marshall988@163.net
Website address:	http://www.tkmedical.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	CE mark
Syringe material(s)	Polypropylene	RUP mechanism	Plunger lock by stainless steel jaw at
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack
Comments	Graduation 0.3ml. Average deadspace 0.005mL		

PRICE AND SHIPPING

Shipping volume (m3)	0.068656	Weight per package (Kg)	10.0000
Pieces per package	2000	Minimum order (unit)	200000.00
Price 1	price on request	Year of base price	2021
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: 7886-3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Anhui Tiankang Medical Technology Co., Ltd
Address	No. 228, Weiyi Road Economic Development Zone 239300 Tianchang China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Kojak Seling 0.3ml
IMD-PQS Code:	E008-084
Status:	Prequalified
Date of acceptance:	30 Apr 2021
Appliance type:	E008
Product description:	0.3ml AD syringe
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	Kojak Seling 0.3ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	RUP mechanism	Locked/broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister & ribbon pack
Comments	Average deadspace 0.025mL		

PRICE AND SHIPPING

Shipping volume (m3)	0.07	Weight per package (Kg)	9.8100
Pieces per package	2000	Minimum order (unit)	20000.00
Price 1	USD 4.50/100 Pcs	Year of base price	2025
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: 7883-3 Applicable PQS VP(s):		

PRODUCT SITES

Organization name	Hindustan Syringes & Medical Devices Ltd.
Address	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Haiou AD Syringe 0.3ml
IMD-PQS Code:	E008-085
Status:	Prequalified
Date of acceptance:	06 Nov 2021
Appliance type:	E008
Product description:	AD Syringe 0.3ml
Manufacturer:	Guangdong Haiou Medical Apparatus Co., Ltd
Manufacturers Reference:	Haiou AD Syringe 0.3ml
Country of Manufacture:	China
Address:	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China
Telephone:	+86-663-2758666
Email:	market01@haiou.net.cn
Website address:	http://www.haiou.net.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	CE mark
Syringe material(s)	Polypropylene	RUP mechanism	Broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.086	Weight per package (Kg)	13.0000
Pieces per package	3000	Minimum order (unit)	100000.00
Price 1	price available on request	Year of base price	2021
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Guangdong Haiou Medical Apparatus Co., Ltd
Address	Nanyuan Industrial Zone, North Liusha, Puning City, Guangdong Province, China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Oneject 0.3ml AD
IMD-PQS Code:	E008-086
Status:	Prequalified
Date of acceptance:	31 Aug 2021
Appliance type:	E008
Product description:	Oneject 0.3ml AD
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	Oneject 0.3ml AD
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	CE mark
Syringe material(s)	Polypropylene	RUP mechanism	Locked / broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	24Gx1" (0.55 x 25 mm)	AD Sub Category	Simple AD
AD Primary packaging	Blister pack		

Comments	Graduation 0.3mL. Average deadspace 0.021mL
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.057	Weight per package (Kg)	7.8000
Pieces per package	1600	Minimum order (unit)	16000.00
Price 1	Upon request to manufacturer	Year of base price	2021
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886/3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	SOL-GUARD Auto-Disable Syringe 0.3ml
IMD-PQS Code:	E008-088
Status:	Prequalified
Date of acceptance:	17 Feb 2022
Appliance type:	E008
Product description:	Auto-Disable Syringe 0.3ml
Manufacturer:	Sol-Millennium Medical Inc.
Manufacturers Reference:	SOL-GUARD Auto-Disable Syringe 0.3ml
Country of Manufacture:	China
Address:	311 South Wacker Drive, Suite 4100 Chicago 60606 Illinois United States of America
Telephone:	+41 76 675 65 86
Email:	jwang@sol-m.com
Website address:	www.sol-m.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.30	Markings	Device license (Canada)
Syringe material(s)	Polypropylene	RUP mechanism	AD
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack

Comments	Graduation 0.3mL. Average deadspace 0.003mL
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.088	Weight per package (Kg)	9.6000
Pieces per package	2400	Minimum order (unit)	100800.00
Price 1	price on request	Year of base price	2022
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s): ISO 7886-3		

PRODUCT SITES

Organization name	Sol-Millennium Medical Inc.
Address	311 South Wacker Drive, Suite 4100 Chicago 60606 Illinois United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Sol-Guard 0.5mL Auto Disable syringe
IMD-PQS Code:	E008-090
Status:	Prequalified
Date of acceptance:	20 Apr 2022
Appliance type:	E008
Product description:	0.5mL Auto Disable syringe
Manufacturer:	Sol-Millennium Medical Inc.
Manufacturers Reference:	Sol-Guard 0.5mL Auto Disable syringe
Country of Manufacture:	China
Address:	311 South Wacker Drive, Suite 4100 Chicago 60606 Illinois United States of America
Telephone:	+41 76 675 65 86
Email:	jwang@sol-m.com
Website address:	www.sol-m.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	Device license (Canada)	Syringe material(s)	Polypropylene
RUP mechanism	locking clip, locked trapped plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.088	Weight per package (Kg)	9.6000
Pieces per package	2400	Minimum order (unit)	100800.00
Price 1	price available on request	Year of base price	2022
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: 7886-3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Sol-Millennium Medical Inc.
Address	311 South Wacker Drive, Suite 4100 Chicago 60606 Illinois United States of America

Current PQS status:	<i>Prequalified</i>	Valid until:	31 May 2026
----------------------------	---------------------	---------------------	-------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	0.1 mL AD syringe
IMD-PQS Code:	E008-091
Status:	Prequalified
Date of acceptance:	17 May 2022
Appliance type:	E008
Product description:	0.1 mL AD syringe
Manufacturer:	Zhejiang Lingyang Medical Apparatus Co., Ltd
Manufacturers Reference:	0.1 mL AD syringe
Country of Manufacture:	China
Address:	Bai Shuiyang 317031 Linhai Peoples's Republic of China China
Telephone:	+86-576 85800258
Email:	lingyangmedical@163.com
Website address:	www.ly-medical.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	----	Syringe material(s)	Polypropylene
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	4000.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	26Gx3/8" (0.45 x 10 mm)
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.138	Weight per package (Kg)	16.6000
Pieces per package	4000	Minimum order (unit)	4000.00
Price 1	Price available on request	Year of base price	2022
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	Zhejiang Lingyang Medical Apparatus Co., Ltd
Address	Bai Shuiyang 317031 Linhai Peoples's Republic of China China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	KOJAK SELINGE 0.2 ml
IMD-PQS Code:	E008-092
Status:	Prequalified
Date of acceptance:	17 May 2022
Appliance type:	E008
Product description:	AD Syringe 0.2 ml
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	KOJAK SELINGE 0.2 ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.20	RUP mechanism	Locked / Broken Plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	2000.00	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister & ribbon pack		

Comments	Average deadspace 0.021mL
-----------------	---------------------------

PRICE AND SHIPPING

Shipping volume (m3)	0.07	Weight per package (Kg)	9.8100
Pieces per package	2000	Minimum order (unit)	20000.00
Price 1	USD 4.70/100 Pcs	Year of base price	2025
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s):		

PRODUCT SITES

Organization name	Hindustan Syringes & Medical Devices Ltd.
Address	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India

Current PQS status:	Prequalified	Valid until:	31 May 2026
----------------------------	--------------	---------------------	-------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	KOJAK SELINGE 0.25 ml
IMD-PQS Code:	E008-095
Status:	Prequalified
Date of acceptance:	06 Feb 2022
Appliance type:	E008
Product description:	0.25 ml AD syringe
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	KOJAK SELINGE 0.25 ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.25	Graduations (ml)	0.25
RUP mechanism	Locked / Broken Plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3 pieces
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister & ribbon pack		

Comments	Average deadspace 0.021mL
-----------------	---------------------------

PRICE AND SHIPPING

Shipping volume (m3)	0.07	Weight per package (Kg)	9.8100
Pieces per package	2000	Minimum order (unit)	20000.00
Price 1	USD 4.70/100 Pcs	Year of base price	2025
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001
Specification Reference	Applicable PQS specification: ISO 7886-3 Applicable PQS VP(s):		

PRODUCT SITES

Organization name	Hindustan Syringes & Medical Devices Ltd.
Address	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India

Current PQS status:	Prequalified	Valid until:	31 May 2026
----------------------------	--------------	---------------------	-------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



No image available

Product name:	Revital Cady 0.5ml 3piece Autodisable Syringe
IMD-PQS Code:	E008-097
Status:	Prequalified
Date of acceptance:	16 Jun 2023
Appliance type:	E008
Product description:	0.5mL AD syringe
Manufacturer:	Revital Health Care (EPZ) Limited
Manufacturers Reference:	Revital Cady 0.5ml 3piece Autodisable Syringe
Country of Manufacture:	Kenya
Address:	P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya
Telephone:	+254722412900
Email:	qa@rhcare-epz.com
Website address:	www.revitalhcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Plunger disables / breaks
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Mechanism	Locked / Broken Plunger
AD Primary packaging	Blister pack		

Comments Units per package: 4000/3600

PRICE AND SHIPPING

Shipping volume (m3)	0.1031	Weight per package (Kg)	11.6000
Pieces per package	4000	Minimum order (unit)	640000.00
Price 1	POA	Year of base price	2022
Incoterms	FCA		

QUALITY STANDARD

Quality Standard ISO 13485

PRODUCT SITES

Organization name	Revital Health Care (EPZ) Limited
Address	P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya

Current PQS status: *Prequalified* **Valid until:** *31 May 2026*

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Iscon 0.5mL AD syringe
IMD-PQS Code:	E008-098
Status:	Prequalified
Date of acceptance:	07 May 2023
Appliance type:	E008
Product description:	Auto Disable Syringe for fixed-dose immunization
Manufacturer:	ISCON SURGICALS LTD.
Manufacturers Reference:	Iscon 0.5mL AD syringe
Country of Manufacture:	India
Address:	B-70 Marudhar Ind. Area, MIA 2nd Phase, Basni, Jodhpur, Rajasthan 342005, India
Telephone:	+91 291 2740558
Email:	nitin@pricon.co.in
Website address:	https://www.isconsurgicals.net/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	2000.00	AD Number pieces	3 pieces
Auto-disabling at:	Middle of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Shipping volume (m3)	0.08	Weight per package (Kg)	9.8000
Pieces per package	2000	Minimum order (unit)	2000.00
Price 1	upto 100,000 / 4.75\$ per 100	Price 2	upto 500,000 / 4.50\$ per 100
Price 3	above 500,00 / 4.25\$ per 100	Year of base price	2023
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

PRODUCT SITES

Organization name	ISCON SURGICALS LTD.
Address	B-70 Marudhar Ind. Area, MIA 2nd Phase, Basni, Jodhpur, Rajasthan 342005, India

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



No image
available

Product name:	Oneject 0.5 AD
IMD-PQS Code:	E008-099
Status:	Prequalified
Date of acceptance:	25 Sep 2023
Appliance type:	E008
Product description:	0.5mL AD syringe
Manufacturer:	PT Oneject Indonesia
Manufacturers Reference:	Oneject 0.5 AD
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	Locked / broken Plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	200.00
AD Number pieces	3 pieces	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.007	Weight per package (Kg)	0.8000
Pieces per package	200	Minimum order (unit)	1000.00
Price 1	Upon request to manufacturer	Year of base price	2023
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

PRODUCT SITES


Organization name	PT Oneject Indonesia
Address	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	REVITAL CADY 0.3ML 2PC AD SYRINGE
IMD-PQS Code:	E008-100
Status:	Prequalified
Date of acceptance:	25 Oct 2023
Product description:	0.3mL AD syringe
Manufacturer:	Revital Health Care (EPZ) Limited
Manufacturers Reference:	REVITAL CADY 0.3ML 2PC AD SYRINGE
Country of Manufacture:	Kenya
Address:	P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya
Telephone:	+254722412900
Email:	qa@rhcare-epz.com
Website address:	www.revitalhcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.30	Markings	CE mark
Syringe material(s)	Polypropylene	Other needle size options	Yes
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Mechanism	Locked / Broken Plunger	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.112	Weight per package (Kg)	11.5700
Pieces per package	3600	Minimum order (unit)	648000.00
Price 1	on request	Year of base price	2023
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	KOJAK SELINGE 0.5ml AD Syringe with S.I.P. feature
IMD-PQS Code:	E008-101
Status:	Prequalified
Date of acceptance:	15 Nov 2023
Product description:	0.5ml AD Syringe with SIP feature
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	KOJAK SELINGE 0.5ml AD Syringe with S.I.P. feature
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.50	Syringe material(s)	Polypropylene
Other needle size options	Yes	Needle Size	23Gx1" (0.60 x 25 mm)
AD location	Start of injection	AD Mechanism	Locked / Broken Plunger
AD Primary packaging	Blister pack		

PRICE AND SHIPPING

Pieces per package	2000	Price 1	USD 7.50/100 Pcs
Year of base price	2025	Incoterms	FCA

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3; ISO 23908
Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	KOJAK SELINGE 1ml AD Syringe with S.I.P. feature
IMD-PQS Code:	E008-102
Status:	Prequalified
Date of acceptance:	25 May 2024
Product description:	1ml AD Syringe with SIP feature
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	KOJAK SELINGE 1ml AD Syringe with S.I.P. feature
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Syringe material(s)	Polypropylene	Other needle size options	Yes
Needle Size	23Gx1" (0.60 x 25 mm)	AD location	Start of injection
AD Mechanism	Locked / Broken Plunger	AD Primary packaging	Blister & ribbon pack
Comments	Kojak Seling 1ml AD Syringe with S.I.P. feature having dual safety feature. One for the patient safety by preventing the Syringe from re-use with the AD mechanism and second for the healthcare personnel from accidental needle stick injury by using Safety feature (SIP) in needle cover. (Complies with ISO 7886-3 & ISO 23908) Graduations: 1ml		

PRICE AND SHIPPING

Shipping volume (m3)	0.09	Weight per package (Kg)	16.4600
Pieces per package	2000	Price 1	USD 7.50/100 Pcs
Year of base price	2025	Incoterms	FCA

QUALITY STANDARD

Quality Standard	ISO 9001: 2000;ISO 13485
-------------------------	--------------------------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Iscon 0.1ml AD syringe
IMD-PQS Code:	E008-103
Status:	Prequalified
Date of acceptance:	21 Oct 2024
Product description:	0.1ml AD syringe
Manufacturer:	ISCON SURGICALS LTD.
Manufacturers Reference:	Iscon 0.1ml AD syringe
Country of Manufacture:	India
Address:	B-70 Marudhar Ind. Area, MIA 2nd Phase, Basni, Jodhpur, Rajasthan 342005, India
Telephone:	+91 291 2740558
Email:	nitin@pricon.co.in
Website address:	https://www.isconsurgicals.net/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.10	Markings	CE mark
Syringe material(s)	Polypropylene	Other needle size options	Yes
Needle Size	26Gx3/8" (0.45 x 10 mm)	AD location	Middle of injection
AD Mechanism	Locked / Broken Plunger	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.109	Weight per package (Kg)	12.3500
Pieces per package	3200	Minimum order (unit)	32000.00
Price 1	3.9\$ upto 32,000pcs	Price 2	3.8\$ 32,000 to 160,000pcs
Price 3	3.7\$ 160,000 and above	Year of base price	2024
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Jiangxi 0.5ml AD syringe
IMD-PQS Code:	E008-104
Status:	Prequalified
Date of acceptance:	21 Oct 2024
Product description:	0.5ml AD syringe
Manufacturer:	Jiangxi Sanxin Medtec Co.,Ltd
Manufacturers Reference:	Jiangxi 0.5ml AD syringe
Country of Manufacture:	China
Address:	999 Fushan Road, Xiaolan Economic Development Zone, Nanchang, Jiangxi, P.C. 330200 China; People's Republic of China
Telephone:	+86 791 85988111
Email:	fgb@sanxin-med.com
Website address:	www.sanxin-med.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.50	Markings	CE mark
Syringe material(s)	Polypropylene	Other needle size options	Yes
Needle Size	23Gx1" (0.60 x 25 mm)	AD location	Start of injection
AD Mechanism	Clip locking plunger	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.1	Weight per package (Kg)	15.0000
Pieces per package	3000	Minimum order (unit)	870000.00
Price 1	0.03-0.04	Year of base price	2024
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	TKMD 0.5ml AD syringe
IMD-PQS Code:	E008-105
Status:	Prequalified
Date of acceptance:	12 Nov 2024
Product description:	TKMD 0.5ml AD syringe
Manufacturer:	TKMD RWANDA Ltd.
Manufacturers Reference:	TKMD 0.5ml AD syringe
Country of Manufacture:	Rwanda
Telephone:	+250798685288
Email:	development@tkmedical.com
Website address:	https://www.tkmedical.com/
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.50	Markings	CE mark
Syringe material(s)	Polypropylene	Other needle size options	Yes
Needle Size	23Gx1" (0.60 x 25 mm)	AD location	Start of injection
AD Mechanism	Clip locking plunger	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.055	Weight per package (Kg)	8.8000
Pieces per package	1800	Minimum order (unit)	500000.00
Year of base price	2024	Incoterms	FCA

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



IMD-PQS Code:	E008-106
Status:	Prequalified
Date of acceptance:	19 Nov 2024
Product description:	AD Syringe 0.25ml
Manufacturer:	Abu Dhabi Medical Devices Co. L.L.C
Address:	MUSSAFAH INDUSTRIAL AREA, M-43, PLOT 124, 30485, ABU DHABI United Arab Emirates
Telephone:	+971-2-6199100
Email:	info@admd.ae
Website address:	http://www.admd.ae
Valid until:	31 May 2026

Specifications

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Shandong 0.3mL AD syringe
IMD-PQS Code:	E008-107
Status:	Prequalified
Date of acceptance:	25 Nov 2024
Product description:	0.3mL AD syringe
Manufacturer:	Shangdong Ande Healthcare Apparatus Co., Ltd
Manufacturers Reference:	Shandong 0.3mL AD syringe
Country of Manufacture:	China
Telephone:	+86-533-3917816
Email:	trade@andemed.com
Website address:	www.andehealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Markings	CE	Syringe material(s)	Polypropylene
Other needle size options	Yes	AD Number pieces	3
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Mechanism	Metal clip	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.087	Weight per package (Kg)	11.0000
Pieces per package	2000	Minimum order (unit)	2000.00
Year of base price	2024	Incoterms	FOB

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Shandong 0.5mL AD syringe
IMD-PQS Code:	E008-108
Status:	Prequalified
Date of acceptance:	25 Nov 2024
Product description:	0.5mL AD syringe
Manufacturer:	Shangdong Ande Healthcare Apparatus Co., Ltd
Manufacturers Reference:	Shandong 0.5mL AD syringe
Country of Manufacture:	China
Telephone:	+86-533-3917816
Email:	trade@andemed.com
Website address:	www.andehealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.50	Markings	CE Mark
Syringe material(s)	Polypropylene	Other needle size options	Yes
Needle Size	21Gx1" (0.80 x 25 mm)	AD location	Start of injection
AD Mechanism	Locking clip	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.0926	Weight per package (Kg)	11.0000
Pieces per package	2000	Minimum order (unit)	2000.00
Price 1	On request	Year of base price	2024
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Safety Syringe 0.05 ml Auto Disable
IMD-PQS Code:	E008-109
Status:	Prequalified
Date of acceptance:	10 Dec 2024
Manufacturer:	International Company for Medical Necessities
Manufacturers Reference:	Safety Syringe 0.05 ml Auto Disable
Country of Manufacture:	Egypt, Arab Rep.
Address:	Industrial Zone Block No. 19&67 Abu Tig Assuit Egypt Egypt
Telephone:	+2 (088) 49 50200 - 202 - 250
Email:	mak@medical-ico.com
Website address:	www.medical-ico.com & aak@medical-ico.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Graduations (ml)	0.05	Markings	CE mark
Syringe material(s)	Polypropylene	Other needle size options	Yes
Needle Size	26Gx3/8" (0.45 x 10 mm)	AD location	Start of injection
AD Mechanism	Locked / Broken Plunger	AD Primary packaging	Blister pack

Comments	Number of components: 3
-----------------	-------------------------

PRICE AND SHIPPING

Shipping volume (m3)	0.08	Weight per Unit (Kg)	13.2000
Pieces per package	2000	Minimum order (unit)	500000.00
Price 1	100,000 units .039 \$	Price 2	1,000,000 units 0.038 \$
Price 3	10,000,000 units 0.037 \$	Year of base price	2022

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Revital Cady 0.5ml 3 piece Autodisable Syringe
IMD-PQS Code:	E008-110
Status:	Prequalified
Date of acceptance:	11 Nov 2024
Manufacturer:	Revital Health Care (EPZ) Limited
Manufacturers Reference:	Revital Cady 0.5ml 3 piece Autodisable Syringe
Country of Manufacture:	Kenya
Address:	P.O. BOX 80713-80100 MOMBASA LR No. 5025/1239, Takaungu 80100 Mombasa Kenya
Telephone:	+254722412900
Email:	qa@rhcare-epz.com
Website address:	www.revitalhcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
Other needle size options	Yes	Auto-disabling at:	End of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD location	Upon completion of injection
AD Mechanism	Locked / Broken Plunger	AD Primary packaging	Blister pack
Comments	Number of components: 3		

PRICE AND SHIPPING

Shipping volume (m3)	0.1031	Weight per package (Kg)	15.9200
Pieces per package	4000	Minimum order (unit)	640000.00
Price 1	POA	Year of base price	2024
Incoterms	FCA		

QUALITY STANDARD

Quality Standard	ISO 13485		
Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



IMD-PQS Code:	E008-112
Status:	Prequalified
Date of acceptance:	07 May 2025
Product description:	AD syringe 0.5ml
Manufacturer:	PT Oneject Indonesia
Country of Manufacture:	Indonesia
Address:	Sentul Industrial Estate Jl.Olympic Raya Kav B9 Bogor 16810, Indonesia Indonesia
Telephone:	+62 21 879 17 422
Email:	niken.permatasari@oneject.co.id
Website address:	www.oneject.co.id
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Syringe material(s)	Polypropylene	RUP mechanism	Locked / Broken plunger
Other needle size options	Yes	Units per package	100.00
AD Number pieces	3	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack

PRICE AND SHIPPING

Shipping volume (m3)	0.0048	Weight per Unit (Kg)	0.6420
Minimum order (unit)	1000.00	Price 1	Upon request to manufacturer
Year of base price	2025	Incoterms	FOB

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	MDSAP
Specification Reference	1. EN ISO 13485 : 2016 (MD 760987) 2. US 21 CFR part 820 3. GMP-MD Indonesia by Ministry of Health		

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	KOJAK SELINGE 0.5ml
IMD-PQS Code:	E008-114
Status:	Prequalified
Date of acceptance:	07 May 2025
Product description:	AD Syringe 0.5ml
Manufacturer:	Hindustan Syringes & Medical Devices Ltd.
Manufacturers Reference:	KOJAK SELINGE 0.5ml
Country of Manufacture:	India
Address:	174, Sector - 25, Ballabgarh - 121 004 (Haryana) India
Telephone:	+91 129-4289000
Email:	hmduk@hmdhealthcare.com
Website address:	www.hmdhealthcare.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE	Syringe material(s)	Polypropylene
RUP mechanism	Locked / Broken plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	2000.00
AD Number pieces	3	Auto-disabling at:	Start of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Sub Category	Simple AD
AD Primary packaging	Blister pack		

Comments	1. Average Dead Space= 0.03 ml. 2. Other Needle Sizes available on request.
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.09	Weight per package (Kg)	13.1000
Minimum order (unit)	20000.00	Price 1	Over 1 million - USD 4.50/100 Pcs
Price 2	Over 10 millions - USD 4.00/100 Pcs	Price 3	Over 50 millions - USD 3.75/100 Pcs
Year of base price	2025	Incoterms	FCA

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 9001:2015; MDSAP
Current PQS status:	Prequalified	Valid until:	31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Hongda 0.1ml AD syringe
IMD-PQS Code:	E008-122
Status:	Prequalified
Date of acceptance:	10 Dec 2018
Appliance type:	E008
Product description:	0.1ml AD syringe
Manufacturer:	Jiangxi Hongda Medical Equipment Group Ltd
Manufacturers Reference:	Hongda 0.1ml AD syringe
Country of Manufacture:	China
Address:	39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of China
Telephone:	+86 791 85628828
Email:	hongda008@jxhd.cn
Website address:	http://www.jxhd.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.10	Graduations (ml)	0.10
Markings	CE mark	RUP mechanism	AD
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	25Gx3/4" (0.50 x 19 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	Pieces per carton: 1500 or other quantity. Other quantities per carton available.
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.0979	Weight per package (Kg)	19.0000
Pieces per package	1500	Minimum order (unit)	100000.00
Price 1	price available on request	Year of base price	2017
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485;Other	Quality standard (Other)	ISO 7886-3
Specification Reference	Applicable PQS specification: Applicable PQS VP(s):		

PRODUCT SITES

Organization name	Jiangxi Hongda Medical Equipment Group Ltd		
Address	39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of China		
Current PQS status:	Prequalified	Valid until:	31 May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	Hongda 0.5ml AD syringe
IMD-PQS Code:	E008-123
Status:	Prequalified
Date of acceptance:	10 Dec 2018
Appliance type:	E008
Product description:	0.5ml AD syringe
Manufacturer:	Jiangxi Hongda Medical Equipment Group Ltd
Manufacturers Reference:	Hongda 0.5ml AD syringe
Country of Manufacture:	China
Address:	39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of China
Telephone:	+86 791 85628828
Email:	hongda008@jxhd.cn
Website address:	http://www.jxhd.cn
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	RUP mechanism	Locked/broken plunger
Packaging & Sterility Assurance	Yes	Other needle size options	Yes
Units per package	100.00	AD Number pieces	3 pieces
Auto-disabling at:	End of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Sub Category	Simple AD	AD Primary packaging	Blister pack

Comments	Pieces per carton: Pieces per carton: 1500 or other quantity available.
-----------------	---

PRICE AND SHIPPING

Shipping volume (m3)	0.0979	Weight per package (Kg)	19.0000
Pieces per package	1500	Minimum order (unit)	100000.00
Price 1	price available on request	Year of base price	2017
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: 7889-3 Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	---

PRODUCT SITES

Organization name	Jiangxi Hongda Medical Equipment Group Ltd
Address	39 South Shengi Road Jinxian county, Nanchang City Jiangxi province China; People's Republic of China

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization



Product name:	ICO 0.5ml Auto Disable syringe
IMD-PQS Code:	E008-124
Status:	Prequalified
Date of acceptance:	30 Apr 2019
Appliance type:	E008
Product description:	0.5ml Auto Disable syringe
Manufacturer:	International Company for Medical Necessities
Manufacturers Reference:	ICO 0.5ml Auto Disable syringe
Country of Manufacture:	Egypt, Arab Rep.
Address:	Industrial Zone Block No. 19&67 Abu Tig Assuit Egypt
Telephone:	+2 (088) 49 50200 - 202 - 250
Email:	mak@medical-ico.com
Website address:	www.medical-ico.com & aak@medical-ico.com
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	CE mark	Syringe material(s)	Polypropylene
RUP mechanism	Locking plunger	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	Units per package	2000.00
AD Number pieces	3 pieces	Auto-disabling at:	End of injection
Needle Size	23Gx1" (0.60 x 25 mm)	AD Primary packaging	Blister pack
Comments	Shipping volume (m3): 0.710 x 0.385 x 0.308		

PRICE AND SHIPPING

Weight per package (Kg)	13.2000	Pieces per package	2000
Minimum order (unit)	500000.00	Price 1	100,000 US\$0.039
Price 2	1,000,000 US\$0.038	Price 3	>1,000,000 US\$0.0378
Year of base price	2019	Incoterms	EXW

QUALITY STANDARD

Quality Standard	ISO 13485	Specification Reference	Applicable PQS specification: Applicable PQS VP(s):
-------------------------	-----------	--------------------------------	--

PRODUCT SITES

Organization name	International Company for Medical Necessities
Address	Industrial Zone Block No. 19&67 Abu Tig Assuit Egypt

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.



E008: Injection devices for immunization

Product name:	0.5ML AD SYRINGE FOR FIXED DOSE IMMUNIZATION
IMD-PQS Code:	E008-125
Status:	Prequalified
Date of acceptance:	20 Aug 2025
Product description:	0.5ML AD SYRINGE FOR FIXED DOSE IMMUNIZATION
Manufacturer:	IM Medicare
Manufacturers Reference:	0.5ML AD SYRINGE FOR FIXED DOSE IMMUNIZATION
Country of Manufacture:	Morocco
Address:	PARC INDUSTRIEL SIDI BOU OTHMANE, LOT 21 COMMUNE SIDI BOU OTHMANE, BENGUERIR Marrakech-Safi Morocco
Telephone:	+212 631 700 900
Email:	contact@immedicare.ma
Website address:	www.immedicare.ma
Valid until:	31 May 2026

Specifications

PRODUCT SPECIFICATIONS - MAIN

Vaccine capacity (mL)	0.50	Graduations (ml)	0.50
Markings	Under review	Syringe material(s)	Other
Syringe Material(s) (Other)	POLYPROPYLENE, THERMOPLASTIC ELASTOMER	Packaging & Sterility Assurance	Yes
Other needle size options	Yes	AD Number pieces	3
Auto-disabling at:	Start of injection	Needle Size	23Gx1" (0.60 x 25 mm)
AD Primary packaging	Blister pack		

Comments	MEDICATION FIXED DOSE IMMUNIZATION AD Syringe with needle is used for immunization with a fixed dose. It is used for administering medications /Vaccines. It consists of a transparent barrel to ensure controlled administration of medicines with graduation, a plunger with seal/gasket and a fixed hypodermic needle (capped for user protection).
-----------------	--

PRICE AND SHIPPING

Shipping volume (m3)	0.057	Weight per Unit (Kg)	6.7000
Pieces per package	1800	Minimum order (unit)	180000.00
Price 1	On request	Year of base price	2025
Incoterms	FOB		

QUALITY STANDARD

Quality Standard	ISO 13485
-------------------------	-----------

Current PQS status:	<i>Prequalified</i>	Valid until:	<i>31 May 2026</i>
----------------------------	---------------------	---------------------	--------------------

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.