

WHO Prequalification of Diagnostics Programme PUBLIC REPORT

**Product: COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0
(TaqMan 48)
Number: PQDx 0126-046-00**

The COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48) with product codes 05212294190 (COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0) and 03587797190 (COBAS AmpliPrep/COBAS TaqMan Wash Reagent) manufactured by Roche Molecular Systems, Inc., CE-marked regulatory version, was accepted for the WHO list of prequalified diagnostics and listed on 01 May 2012. This version of the product is intended to be used in conjunction with the following instruments/software: 03279332001 (COBAS TaqMan 48), 03051315001 (COBAS AmpliPrep), 07347308001 (Amplilink Software version 3.4) or 05807875001 (Amplilink Software version 3.3). Optional: 05527503001 (cobas p 630 instrument) and 28127387001 (Docking station).

Summary of prequalification status for COBAS AmpliPrep / COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48)

	Date	Outcome
Status on PQ list	01-May-2012	listed
Dossier assessment	06-Mar-2012	MR
Inspection status	30-Mar-2017	MR
Laboratory evaluation	FT	MR

MR: Meets Requirements

FT: Fast-tracked

Report amendments and/or product changes

This public report has since been amended. Amendments may have arisen because of changes to the prequalified product for which WHO has been notified and has undertaken a review. Amendments to the report are summarized in the following table, and details of each amendment are provided below.

Version	Summary of amendment	Date of report amendment
2.0	1. Inclusion of a specimen collection device, the Cobas Plasma Separation Card (PSC, with product code 07963084190), and subsequent addition of dried plasma spot as an additional specimen type that can be used with the test. 2. The removal of a fulfilled prequalification commitment was also reflected in this amendment.	5-Sep-2018
3.0	Change of the notified body from LRQA to BSI	6-Nov-2019

Intended use:

According to the claim of intended use from Roche Molecular Systems, Inc., “*COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0 (v2.0) is an in vitro nucleic acid amplification test for the quantitation of Human Immunodeficiency Virus Type 1 (HIV-1) RNA in human EDTA plasma or from a cobas Plasma Separation Card (PSC) dried plasma spot using the COBAS AmpliPrep Instrument for automated specimen processing and the COBAS TaqMan Analyzer or COBAS TaqMan 48 Analyzer for automated amplification and detection. The test can quantitate HIV-1 RNA over the range of 20 - 10,000,000 copies/mL (33 to 1.67×10^7 International Units [IU]/mL in EDTA plasma and 738 – 10,000,000 copies/mL (1230 to 1.67×10^7 International Units [IU]/mL) in a PSC dried plasma spot). One copy of HIV-1 RNA is equivalent to 1.67 IU based on the WHO 1st International Standard for HIV-1 RNA for Nucleic Acid-Based Techniques (NAT) (NIBSC 97/656).*

This test is intended for use in conjunction with clinical presentation and other laboratory markers of disease progress for the clinical management of HIV-1 group M and HIV-1 group O infected patients. The test can be used to assess patient prognosis by measuring the baseline HIV-1 RNA level or to monitor the effects of antiretroviral therapy by measuring changes in HIV-1 RNA levels during the course of antiretroviral treatment.

The COBAS AmpliPrep/COBAS TaqMan HIV-1 Test v2.0 is not intended for use as a screening test for the presence of HIV-1 in blood or blood products or as a diagnostic test to confirm the presence of HIV-1 infection”.

Assay description:

According to the claim of assay description from Roche Molecular Systems, Inc., “*COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, v2.0 is a nucleic acid amplification test for the quantitation of Human Immunodeficiency Virus Type 1 (HIV-1) RNA in human EDTA plasma or from a PSC dried plasma spot. The COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, v2.0 is based on three major processes:*

- (1) Specimen preparation to isolate HIV-1 RNA;*
- (2) Reverse transcription of the target RNA to generate complementary DNA (cDNA),*

(3) Simultaneous PCR amplification of target cDNA and detection of cleaved dual-labeled oligonucleotide detection probe specific to the target.

The COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, v2.0 permits automated specimen preparation followed by automated reverse transcription, PCR amplification and detection of HIV-1 target RNA and HIV-1 Quantitation Standard (QS) Armored RNA. The Master Mix reagent contains primers and probes specific for both HIV-1 RNA and HIV-1 QS RNA. The Master Mix has been developed to ensure equivalent quantitation of group M subtypes of HIV-1 and of HIV-1 group O. The detection of amplified DNA is performed using target-specific and QS-specific dual-labeled oligonucleotide probes that permit independent identification of HIV-1 amplicon and HIV-1 QS amplicon.

The quantitation of HIV-1 viral RNA is performed using the HIV-1 QS. It compensates for effects of inhibition and controls the preparation and amplification processes, allowing a more accurate quantitation of HIV-1 RNA in each specimen. The HIV-1 QS is a non-infectious Armored RNA construct that contains HIV sequences with identical primer binding sites as the HIV-1 target RNA and a unique probe binding region that allows HIV-1 QS amplicon to be distinguished from HIV-1 target amplicon.

The HIV-1 QS is added to each specimen at a known copy number and is carried through the subsequent steps of specimen preparation, reverse transcription, simultaneous PCR amplification and detection of cleaved dual-labeled oligonucleotide detection probes. The COBA TaqMan Analyzer or COBAS TaqMan 48 Analyzer calculates the HIV-1 RNA concentration in the test specimens by comparing the HIV-1 signal to the HIV-1 QS signal for each specimen and control”.

In order to perform the assay, the following components are required:

MATERIALS PROVIDED		
Reagents:		Product code
	COBAS AmpliPrep/COBAS TaqMan HIV-1 Test version 2.0	05212294190
	COBAS AmpliPrep/COBAS TaqMan Wash Reagent	03587797190
	cobas Specimen Pre-Extraction Reagent (SPEX)	06989861190
MATERIALS REQUIRED BUT NOT PROVIDED		
Instrumentation:		Product code
	COBAS TaqMan Analyzer or COBAS TaqMan 48 Analyzer	03279332001
	COBAS Ampliprep Instrument	03051315001
Software:		Series
	AMPLILINK Software, Version 3.3	05807875001
	AMPLILINK Software, Version 3.4	07347308001

	Control unit for the AMPLILINK software with printer	07140827001
Optional:		Product code
	cobas p 630 instrument	05527503001
	Docking station	28127387001
Disposables:		Product code
	Sample processing units (SPUs)	03755525001
	Sample input tubes (S-tubes) with barcode clips	03137040001
	Racks of K-tips	03287343001
	K-tube Box of 12 x 96	03137082001
OTHER MATERIALS REQUIRED BUT NOT PROVIDED		
For EDTA plasma specimen application		
	Sample Rack (SK 24 rack)	028122172001
	Reagent Rack	028122199001
	SPU rack	05471664001
	K-tube capper, motorized	03516539001
	K-tube capper	03339874001
	K-carrier	03341488001
	K-carrier Transporter	03517519001
	K-carrier rack	03286436001
	Pipettors with aerosol barrier or positive displacement RNase-free tips (capacity 1000 µL)	N/A
	Disposable gloves, powderless	N/A
	Vortex mixer	N/A
For PSC dried plasma spot specimen application		
	Cobas Plasma Separation Card	07963084190
	Sterile or disposable forceps or tweezers	N/A
	140 µL capillary with compatible dispenser	N/A
	Single use lancing device	N/A
	Sample bag and silica gel desiccant sachets	N/A
	Transport bag	N/A
	Pipette	N/A
	Eppendorf Thermomixer with Thermoblock for 24 cryo tubes	N/A

Storage:

Item	Product code	Temperature
COBAS AmpliPrep/COBAS TaqMan HIV-1 Test version 2.0	05212294190	2 to 8°C
COBAS AmpliPrep/COBAS TaqMan Wash Reagent	03587797190	2 to 30°C
cobas Specimen Pre-Extraction Reagent (SPEX)	06989861190	2 to 8°C

cobas Plasma Separation Card storage and handling requirements are specified in the PSC method sheet ms_07963084190

Shelf-life:

Item	Product code	Shelf-life
COBAS AmpliPrep/COBAS TaqMan HIV-1 Test version 2.0	05212294190	18 months
COBAS AmpliPrep/COBAS TaqMan Wash Reagent	03587797190	24 months
cobas Specimen Pre-Extraction Reagent (SPEX)	06989861190	22 months
cobas Plasma Separation Card	07963084190	12 months

Prioritization for prequalification

Roche Molecular Systems, Inc. submitted an application for prequalification of COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48). Based on the established eligibility criteria, COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48) was given priority for prequalification assessment.

Product dossier assessment

Roche Molecular Systems, Inc. submitted a product dossier for COBAS AmpliPrep/ COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48) as per the *'Instructions for compilation of a product dossier'* (PQDx_018 v1). The information submitted in the product dossier was reviewed by WHO staff and external experts (assessors) appointed by WHO in accordance with the internal report on the screening and assessment of a product dossier (PQDx_009 v2). Based on the product dossier screening and assessment findings, a recommendation was made to accept the product dossier for COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48) for prequalification.

Commitment for prequalification

Roche Molecular Systems, Inc commits to submit a revision to the CD-ROM test definition file label for COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, v2.0, COBAS TaqMan 48 Analyzer by 30 June 2020.

Manufacturing site inspection

An inspection was performed at the site of manufacture (1080 US Hwy 202 South, Branchburg, NJ 08876-3771, USA) of the COBAS AmpliPrep/-COBAS TaqMan HIV-1 Test v2.0 on 8,9 December 2016 as described in 'Information for manufacturers on WHO

prequalification inspection procedures for the sites of manufacture of diagnostics (PQDx_014)'.

The inspection found that the manufacturer had a well-established quality management system and manufacturing practices in place that would ensure the manufacture of product of consistent quality.

The manufacturer's response to the inspection finding was accepted, and the inspection thus finalized, on 30 March 2017.

Product performance evaluation

Given the regulatory version of the product submitted for prequalification and the quality of the data submitted as part of the product dossier to support the claims for its intended use, the COBAS AmpliPrep/COBAS TaqMan HIV-1 Test, version 2.0 (TaqMan 48) assay was found eligible to undergo the WHO fast-track¹ procedure. Subsequently, the product was not required to undergo a laboratory evaluation for its use with human plasma.


¹ Product performance evaluation went through WHO's Fast Track procedure at the date of prioritization for assessment. Fast Track procedure was phased out end of 2013.

Labelling


- 1. Labels**
- 2. Instructions for use**

1. Labels






1.1 Plasma separation card




1.2 COBAS AmpliPrep/COBAS TaqMan HIV-1 Test Kit Carton



Cont.	
HIV-1 v2.0 CS1	1 x 48 Tests
HIV-1 v2.0 CS2	1 x 48 Tests
HIV-1 v2.0 CS3	1 x 48 Tests
Puze	1 x 3.8 mL
EB	1 x 7.0 mL
HIV-1 v2.0 CS4	1 x 48 Tests
HIV-1 OS	1 x 3.8 mL
HIV-1 MMX	1 x 2.5 mL
CAP/CTM Mn ²⁺	1 x 19.8 mL
⊕ HIV-1 H(+)/C, v2.0	4 x 1.0 mL
⊕ HIV-1 L(-)/C, v2.0	4 x 1.0 mL
⊕ CTM (-) C	4 x 1.0 mL
HIV-1 H(+)/C, v2.0 Clip	1 x 4 Clips
HIV-1 L(-)/C, v2.0 Clip	1 x 4 Clips
HIV-1 (-) C Clip	1 x 4 Clips

Xi	 HIV-1 v2.0 CS1 Iritant/Iritant/Iritant/Lokalirritant/Lirritant/Ústředný/Ústřední látka 02% (w/v, Gew.-%, m/m, p/p, v/v, V/V) Isopropanol/ Isopropanol/ Isopropanol
F	 HIV-1 v2.0 CS1 Highly flammable/Lichtstanzendlich/Facilmente inflamabile/Facilmente inflamabile/ Facilmente inflamabile/Facilmente inflamabile/ Mogyt brennlíky/ Mogyt brennlíky/ Substancia letavolná/ Mirovatiné horľavé 02% (w/v, Gew.-%, m/m, p/p, v/v, V/V) Isopropanol/ Isopropanol/ Isopropanol
Xn	 HIV-1 v2.0 CS2 Harmful/Gesundheitsschädlich/ Nociv/ Nociv/ Sanjehetshakidakh/ Hilsschädlich/Škodlivý s účinkem/ Zdravie škodlivé 02.5% (w/v, Gew.-%, m/m, p/p, v/v, V/V) Quantité dangereuse/Quantidade perigosa/ Toxicitate de gravitate/ Toxicitate di gravitate/ Toxicante de gravitate/ Gauriditocypant/ Toxycypant gauridyp/ Gauriditocypant
N	 HIV-1 v2.0 CS2 Harmful to the Environment/Únevredlivý/ Nocivo a Terrestrial/ Nocivo per Terrestrial/Prejudicial para el medioambiente/Prejudicial para o meio ambiente/ Škodlivý for milieu/ Škodlivý for miljøet/ Substancia nociva dla środowiska/ Škodlivý pro životní prostředí <14% (w/v, Gew.-%, m/m, p/p, v/v, V/V) Polibocanol/ polycanol/ polibocanol/ Polibocanol/ polybocanol/ polybocanol/ Polibocanol
Xn	 HIV-1 v2.0 CS3 Harmful/Gesundheitsschädlich/ Nociv/ Nociv/ Sanjehetshakidakh/ Hilsschädlich/Škodlivý s účinkem/ Zdravie škodlivé 5.72% (w/v, Gew.-%, m/m, p/p, v/v, V/V) Protetanol/Protetanol/Protetanol/Protetanol/Protetanol/Protetanol/Protetanol

05328292001-03



(01) 2 0875197 00301 1

HI2CAP

ROCHE, COBAS, AMPLIPREP and TAQMAN are trademarks of Roche. ©2010 Roche Diagnostics




Distributed by

Roche Molecular Systems, Inc., Branchburg, NJ 08876 USA
A Member of the Roche Group/ Ein Mitglied der Roche-Gruppe/ Un membre du Groupe Roche/
Un membro del gruppo Roche/ Un miembro del Grupo Roche/ Membro do Grupo Roche/
Medlem af Roche-gruppen/ Medlem i Roche-koncernen / Członak Grupy Roche/ Člen skupiny Roche Group

Roche Diagnostics (Schweiz) AG Industristrasse 7 6040 Rotkreuz, Switzerland Roche Diagnostics GmbH Sandhofer Straße 116 66305 Mannheim, Germany Roche Diagnostics, S.L. Avenida Generalitat, 171-173 E-08174 Sant Cugat del Valles Barcelona, Spain	Roche Diagnostics 201, boulevard Armand-Frappier H7V 4A2 Laval, Quebec, Canada (For Technical Assistance call: appel: kt: 1-877-273-3433) Roche Diagnostics 2, Avenue du Vercois 38240 Meylan, France	Distributore in Italia: Roche Diagnostics S.p.A. Viale G. B. Stucchi 110 20052 Monza, Milano, Italy Distribuidor em Portugal: Roche Sistemas de Diagnósticos Lda. Estrada Nacional, 240-1 2720-413 Amadora, Portugal	Roche Diagnostica Brasil Ltda. Av. Engenheiro Billings, 1720 Jaguari, Building 10 05321-010 São Paulo, SP Brazil
--	--	---	---

Made in USA

NOTICE

The purchase of this product allows the purchaser to use it for amplification and detection of nucleic acid sequences by polymerase chain reaction (PCR) and related processes for human *in vitro* diagnostics. No general patent or other license of any kind other than this specific right of use from purchase is granted hereby.

HINWEIS

Mit dem Kauf dieses Produkts erwirbt der Käufer das Recht, das Produkt für die Amplifikation und den Nachweis von Nukleinsäuresequenzen mittels Polymerase-Kettenreaktion (PCR) und für verwandte Verfahren in der *in-vitro*-Humandiagnostik zu benutzen. Durch diesen Kauf wird weder ein allgemeines Patent noch eine Lizenz irgendwelcher Art verliehen, sondern nur dieses spezifische Nutzungsrecht.

AVERTISSEMENT

L'achat de ce produit permet à l'acheteur de l'utiliser pour procéder à une amplification et une détection des séquences d'acides nucléiques par réaction de polymérisation en chaîne (PCR) et autres méthodes qui y sont associées lors de diagnostics *in vitro* chez l'homme. Aucun brevet général ou licence autre que le droit d'utilisation spécifique délivré lors de l'achat ne sera autorisé.

AVVISO

L'acquisto di questo prodotto ne consente l'uso ai fini dell'amplificazione e della rilevazione delle sequenze di acido nucleico tramite reazione polimerasica a catena (PCR) e processi associati per la diagnostica umana *in vitro*. L'acquisto non comporta la concessione di alcun brevetto generale né di altra licenza di qualsiasi tipo eccetto lo specifico diritto d'uso qui indicato.

AVISO

La adquisición de este producto permite que el comprador lo utilice para la amplificación y detección de secuencias de ácido nucleico mediante la reacción en cadena de la polimerasa (PCR) y otros procedimientos relacionados, en el diagnóstico *in vitro* del ser humano. No se adjudica patente general o licencia alguna de cualquier tipo que no sea el derecho específico de uso de esta adquisición concedido en este documento.

AVISO

A aquisição deste produto permite ao comprador utilizá-lo para a amplificação e detecção de sequências de ácidos nucleicos por reação de polimerização em cadeia (PCR) e procedimentos relacionados para diagnósticos *in vitro* no ser humano. Não se concede por este meio nenhuma patente geral nem outra licença de qualquer tipo, salvo este específico direito de utilização decorrente da compra.

BEMÆRKNING

Købet af dette produkt giver køberen tilladelse til at anvende det til forstærkning og påvisning af nukleinsyresekvenser ved hjælp af polymerase kædereaktion (PKR) og relaterede metoder til menneskelig *in vitro* diagnostik. Der gives ved dette køb intet generelt patent eller anden licens af nogen som helst slags udover den specifikke brugsret.

MEDELÄNDE


Inköp av denna produkt medför rätt för köparen att använda produkten för förstärkning och detektering av nukleinsyresekvenser via polymeraskedjereaktion (PCR) och relaterade processer för *in vitro* diagnos hos människor. Inget allmänt patent eller någon annan sorts licens än denna specifika användningsrätt på grund av inköp, medges härmed.

UWAGA


Zakup niniejszego produktu umożliwia nabywcy używanie go do amplifikacji i detekcji sekwencji kwasu nukleinowego za pomocą łańcuchowej reakcji polimerazy (PCR) i odpowiednich procesów w celach diagnostyki *in vitro* próbek pochodzących od ludzi. Wraz z nabyciem niniejszego produktu nabywca nie otrzymuje żadnego patentu ogólnego lub innej licencji poza prawem do określonego, szczególnego zastosowania produktu.

OZNÁMENIE

Nákup tohto produktu oprávňuje kupujúceho použiť tento produkt na amplifikáciu a detekciu sekvencií nukleovej kyseliny prostredníctvom reakcie reťazca polymerázy (PCR) a súvisiacich procesov určených na ľudskú *in-vitro* diagnostiku. Okrem práv na používanie vyplývajúcich z nákupu sa neudel'uje žiadny patent a ani iná licencia.



1.3 Insert License Novartis HIV/HCV

Front	Back
<p>NOTICE Certain components of this product are covered by one or more U.S. Patents and their foreign equivalents in certain countries issued to Novartis Vaccines and Diagnostics, Inc. and licensed to Roche Molecular Systems, Inc. and F. Hoffmann-La Roche Ltd.</p> <p>HINWEIS Bestimmte Bestandteile dieses Produkts unterliegen einem oder mehreren U.S.-Patenten und ihren Entsprechungen außerhalb der USA, die Novartis Vaccines and Diagnostics, Inc. erteilt und an Roche Molecular Systems, Inc. und F. Hoffmann-La Roche Ltd. lizenziert wurden.</p> <p>AVERTISSEMENT Certains composants de ce produit sont protégés par un ou plusieurs brevets américains et leurs équivalents étrangers accordés à Novartis Vaccines and Diagnostics, Inc. et exploités sous licence par Roche Molecular Systems, Inc. et F. Hoffmann-La Roche Ltd.</p> <p>AVISO Alcuni componenti di questo prodotto sono protetti da uno o più brevetti statunitensi e dai loro equivalenti esteri, assegnati alla Novartis Vaccines and Diagnostics, Inc. e concessi su licenza alla Roche Molecular Systems, Inc. ed alla F. Hoffmann-La Roche Ltd.</p> <p>AVISO Certos componentes de este producto están cubiertos por una o más patentes de EE.UU. y sus equivalentes de otros países, concedidas a Novartis Vaccines and Diagnostics, Inc. y con licencia otorgada a Roche Molecular Systems, Inc. y a F. Hoffmann-La Roche Ltd.</p> <p>AVISO Certos componentes deste produto estão protegidos por uma ou mais patentes dos Estados Unidos e suas equivalentes internacionais, concedidas à Novartis Vaccines and Diagnostics, Inc. e com licenças concedidas a Roche Molecular Systems, Inc. e F. Hoffmann-La Roche Ltd.</p>	<p>BEMÆRKNING Nogle af dette produkts komponenter er dækket af et eller flere amerikanske patenter og deres udenlandske ækvivalenter, som er givet til Novartis Vaccines and Diagnostics, Inc. og givet i licens til Roche Molecular Systems, Inc. og F. Hoffmann-La Roche Ltd.</p> <p>MEDELÄNDE Vissa komponenter i denna produkt omfattas av ett eller flera patent i USA samt motsvarande patent i utlandet som utbejdats till Novartis Vaccines and Diagnostics, Inc. och licensierats till Roche Molecular Systems, Inc. och F. Hoffmann-La Roche Ltd.</p> <p>UWAGA Powne składniki tego produktu są chronione przez co najmniej jeden patent zarejestrowany w USA i ich parzystywe odpowiedniki zawierające licencję firmy Roche Molecular Systems, Inc. oraz firmy F. Hoffmann-La Roche Ltd. i przekazano firmie Novartis Vaccines and Diagnostics, Inc.</p> <p>OZNÁMENIE Na niektoré súčasť tohto produktu sa vzťahujú jeden alebo viaceré patenty USA a ich zahraničné ekvivalenty vystavené pre spoločnosť Novartis Vaccines and Diagnostics, Inc. a licencie poskytnuté spoločnosť Roche Molecular Systems, Inc. a F. Hoffmann-La Roche Ltd.</p> <p> Roche Molecular Systems, Inc. Branchburg, NJ 08876 USA</p> <p>051633027001-02</p>

1.4 Export Only Label

Actual Size: 5" x 3"
All prints black.
Die lines do not print.



1.5 Insert Product Info Card

COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0


KIT	LOT	IVD	HI2CAP
			LOT
	LOT		
HIV-1 v2.0 CS1		HIV-1 L(+)C, v2.0	
HIV-1 v2.0 CS2		HIV-1 H(+)C, v2.0	
HIV-1 v2.0 CS3		HIV-1 (-) C Clip	
HIV-1 v2.0 CS4		HIV-1 L(+)C, v2.0 Clip	
CTM (-) C		HIV-1 H(+)C, v2.0 Clip	

 COBAS® TaqMan® Analyzer:
 HI2PSC48 TDF v2.0 (AMPLILINK Software v3.3, v3.4; cp/mL)
 REF 08411778190


COBAS® TaqMan® Analyzer:
 HI2PSC96 TDF v2.0 (AMPLILINK Software v3.3, v3.4; cp/mL)
 REF 08410453190

COBAS® TaqMan® 48 Analyzer:
 HI2CAP48 TDF v2.1 (AMPLILINK Software v3.3, v3.4; cp/mL)
 REF 06584756190

COBAS® TaqMan® Analyzer:
 HI2CAP96 TDF v2.1 (AMPLILINK Software v3.3, v3.4; cp/mL)
 REF 08585655190




 website: <http://e-labdoc.roche.com>
 Method Sheet Catalog No.: 05212294190
 Doc Rev. 15.0

Please contact your local Roche affiliate if you require a printed copy free of charge or need technical support to access the package insert. Bei Ihrer zuständigen Roche-Vertretung erhalten Sie einen kostenfreien Ausdruck oder technische Unterstützung für den Zugriff auf die Packungsbeilage./ Veuillez contacter votre société Roche locale pour obtenir un exemplaire papier gratuit ou une assistance technique pour accéder à la notice./ Contattare il rappresentante Roche locale per ottenere gratuitamente una copia stampata o richiedere istruzioni per reperire il foglio illustrativo./ Póngase en contacto con su filial local de Roche si necesita una copia impresa gratuita o ayuda del servicio técnico para acceder al boletín técnico./ Se desejar uma cópia impressa gratuita ou necessitar de assistência técnica para aceder ao folheto informativo, entre em contacto com o representante local da Roche./ Kontakt den lokale Roche-representant, hvis du ønsker en gratis skriftlig kopi eller har brug for teknisk support for at få adgang til indlægsedlen./ Kontakta din Roche-representant om du vill ha en pappersversion kostnadsfritt eller om du behöver teknisk support för att komma åt bipacksedeln./ Aby otrzymać darmową drukowaną wersję lub jeśli zaistniały problemy techniczne z dostępem do ulotki dołączonej do opakowania, należy skontaktować się z miejscowym oddziałem firmy Roche./ Ak potrebujete bezplatnú vytlačení kópiu alebo ak potrebujete technickú pomoc ohľadne prístupu k príbalovým letákom, obráťte sa na miestne zastúpenie spoločnosti Roche.




 Roche Molecular Systems, Inc.
 1080 US Highway 202 South
 Branchburg, NJ 08876 USA

05328284001-10



1.6 HIV-1 v2.0 CS1 (HIV-1 v2.0, Magnetic Glass Particles Reagent Cassette)

<p>HIV-1 v2.0 CS1</p> <p>IVD</p> <p>48 Tests</p>  <p>Roche Molecular Systems, Inc. Branchburg, NJ 08876 USA</p>	<p>cobas[®] HIV-1 v2.0 CS1 COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0</p> <p>05212294190</p> <p>Cont.</p> <p>B: MGP 7.0 mL A: _____ C: _____</p>   <p>H225, H319, H336</p> <p>05340004001-03</p>
---	---

1.7 HIV-1 CS2 (HIV-1 v2.0, Lysis Reagent)

<p>HIV-1 v2.0 CS2</p> <p>IVD</p> <p>48 Tests</p>  <p>Roche Molecular Systems, Inc. Branchburg, NJ 08876 USA</p>	<p>cobas[®] HIV-1 v2.0 CS2 COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0</p> <p>05212294190</p> <p>Cont.</p> <p>LYS 78 mL</p>   <p>H302, H332, H315, H318, H412</p> <p>05339995001-04</p>
---	---


1.8 HIV-1 v2.0 CS3 (HIV-1 v2.0, Multi-Reagent Cassette)

<p>HIV-1 v2.0 CS3</p> <p>IVD</p> <p>48 Tests</p>  <p>Roche Molecular Systems, Inc. Branchburg, NJ 08876 USA</p>	<p>cobas[®]</p> <p>HIV-1 v2.0 CS3</p> <p>COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0</p> <p>05212294190</p> <p>Cont.</p> <p>B: Pase 3.8 mL</p> <p>A: _____</p> <p>C: EB 7.0 mL</p>  <p>H315, H318, H334</p> <p>05340012001-03</p>
---	--

1.9 HIV-1 v2.0 CS4 (HIV-1 v2.0, Test-Specific Reagent Cassette)

Actual Size: 4.095" x 2.125"

c(o)bas logo prints Black & PMS 340
 All else prints black.
 Die lines do not print.

<p>HIV-1 v2.0 CS4</p> <p>IVD</p> <p>48 Tests</p>  <p>Roche Molecular Systems, Inc. Branchburg, NJ 08876 USA</p>	<p>cobas[®]</p> <p>HIV-1 v2.0 CS4</p> <p>COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0</p> <p>05212294190</p> <p>Cont.</p> <p>B: HIV-1 QS 3.6 mL</p> <p>A: CAP/CTM Mn²⁺ 19.8 mL</p> <p>C: HIV-1 MMX 2.5 mL</p> <p>05340047001-02</p>
---	--

1.10 Quantitation Standard Vial Label

05160642001
HIV – 1 QS, v1.5
3.6 mL 2–8°C
Roche Molecular Systems, Inc.
Branchburg, NJ USA L1203–02

LOT: ???????
EXP: ???????

1.11 Master Mix Vial Label

05219680001
HIV – 1 MMX, v1.5
2.5 mL 2–8°C
Roche Molecular Systems, Inc.
Branchburg, NJ USA L1204–02

LOT: ???????
EXP: ???????

1.12 Manganese Vial


05060451001
CAP/CTM Mn²⁺
19.8 mL 2-8°C
Roche Molecular Systems, Inc.
Branchburg, NJ USA L1111-03
LOT: ???????
EXP: ???????

1.13 HIV-1 H (+) C v2.0 (HIV-1 v2.0, High Positive Control)


HIV-1 H(+), v2.0 05332591001
1.0 mL
IVD
Biohazard symbol
2-8°C
Roche Molecular Systems, Inc.
Branchburg, NJ USA
LOT /

7/16" x 13/16" unvarnished area

1.14 HIV-1 L (+) C v2.0 (HIV-1 v2.0, Low Positive Control)



1.15 CTM (-) C [COBAS TaqMan Negative Control (Human Plasma)]




1.16 High Positive Control Barcode Clip

4 Pcs. 05212294190


HIV-1 H(+)**C**, v2.0 Clip

HIV-1 v2.0 – COBAS® AmpliPrep/COBAS® TaqMan® Tests



KIT LOT

CTL LOT



Roche Roche Molecular Systems, Inc., Branchburg, NJ USA


L1267-02

1.17 Low Positive Control Barcode Clip

4 Pcs. 05212294190


HIV-1 L(+)**C**, v2.0 Clip

HIV-1 v2.0 – COBAS® AmpliPrep/COBAS® TaqMan® Tests



KIT LOT

CTL LOT



Roche Roche Molecular Systems, Inc., Branchburg, NJ USA


L1266-02

1.18 Negative Control Barcode Clip

4 Pcs. 05212294190

HIV-1 (-) C Clip



HIV-1 v2.0 – COBAS® AmpliPrep/COBAS® TaqMan® Tests



2°C 30°C


KIT LOT

CTL LOT

Roche Molecular Systems, Inc., Branchburg, NJ USA
L1268-02

1.19 CD-ROM Test Definition File 48 Tests



CD (base) prints white.

Type prints PMS 340.

PMS 340 overprints white.

Type knocks out to white.

Type overprints black.

This area does not print.

1.20 COBAS AmpliPrep/COBAS TaqMan Wash Reagent Labels

COBAS® AmpliPrep/COBAS® TaqMan® Wash Reagent

PG WR

5.1 L (5100 mL)

COBAS® AmpliPrep/COBAS® TaqMan® Tests

For USA* **Cont.**

Sodium citrate dihydrate
< 0.196 N-Methylisothiazolone HCl

* Composition information required by US only.

REF P/N: 03587797 190

IVD

2°C / 30°C

EC REP Roche Diagnostics GmbH
D-68298 Mannheim

CE

LOT **CM**

Unvarnished area = 1" x .5"

D Roche Diagnostics, 9115 Hague Road, Indianapolis, IN 46290-0467 USA
(For Technical Assistance call the Roche Response Center toll-free: 1-800-826-1247)

Roche Diagnostics (Schweiz) AG, Industriestrasse 7, 6848 Rotkreuz, Switzerland

Roche Diagnostics GmbH, Sandhofer Strasse 116, 66305 Mannheim, Germany

Roche Diagnostics, S.L. Avda. Generalitat, 171-179, E-08174 Sant Cugat del Valles, Barcelona, Spain

Roche Diagnostics, 201, Boulevard Armand-Frappier, H7V 4K2 Laval, Quebec, Canada
(For Technical Assistance call: Pour toute assistance technique, appelez le: 1-877-220-2963)

Roche Diagnostics, 2, Avenue du Vercoors, 98240 Meylan, France

Distributore in Italia: Roche Diagnostics S.p.A., Viale G. B. Stucchi 110, 20062 Monza, Milano, Italy

Distribuidor em Portugal: Roche Sistemas de Diagnósticos Lda, Estada Nacional, 249-1, 2720-419 Amadora, Portugal

Roche Diagnostica Brasil Ltda, Av. Engenharia Billings, 1729, Jaguaris Building 10, 08221-010 São Paulo, SP, Brazil

Roche Molecular Systems, Inc., Branchburg, NJ 08876 USA
A Member of the Roche Group/ Ein Mitglied der Roche-Gruppe/ Un membre du Groupe Roche/
Un membro del gruppo Roche/ Un miembro del Grupo Roche/ Membro do Grupo Roche/ Medlem af Roche-gruppen/
Medlem i Roche-lonnesmen/ Członek Grupy Roche/ Člen skupiny Roche Group

COBAS, AMPLIPREP and TAQMAN are trademarks of Roche. ©2011 Roche Molecular Systems, Inc. All rights reserved. Made in USA 03667367001-03

1.21 Low Positive Control bar code clippers

4 Pcs. **05212294190**

HIV-1 L(+)C, v2.0 Clip

HIV-1 v2.0 – COBAS® AmpliPrep/COBAS® TaqMan® Tests

2°C / 30°C

KIT **LOT** []

CTL **LOT** []

Roche **Roche Molecular Systems, Inc., Branchburg, NJ USA**


Print Form L1266-02

1.22 High Positive Control bar code clippers

4 Pcs. **05212294190**


HIV-1 H(+)**C**, v2.0 Clip


HIV-1 v2.0 – COBAS® AmpliPrep/COBAS® TaqMan® Tests



KIT LOT

CTL LOT





Roche Molecular Systems, Inc., Branchburg, NJ USA

Print Form


L1267-02

1.23 Negative Control bar code clippers

4 Pcs. **05212294190**


HIV-1 (-) **C** Clip


HIV-1 v2.0 – COBAS® AmpliPrep/COBAS® TaqMan® Tests



KIT LOT

CTL LOT





Roche Molecular Systems, Inc., Branchburg, NJ USA

Print Form

L1268-02

2. Instructions for use²

² English version of the IFU was the one that was assessed by WHO. It is the responsibility of the manufacturer to ensure correct translation into other languages

COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, version 2.0

cobas®

FOR *IN VITRO* DIAGNOSTIC USE.

COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0	HI2CAP	48 Tests	P/N: 05212294 190
COBAS® AmpliPrep/COBAS® TaqMan® Wash Reagent	PG WR	5.1 Liters	P/N: 03587797 190

INTENDED USE

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, version 2.0 (v2.0) is an *in vitro* nucleic acid amplification test for the quantitation of Human Immunodeficiency Virus Type 1 (HIV-1) RNA in human EDTA plasma or from a cobas® Plasma Separation Card (PSC) dried plasma spot using the COBAS® AmpliPrep Instrument for automated specimen processing and the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer for automated amplification and detection. The test can quantitate HIV-1 RNA over the range of 20 - 10,000,000 copies/mL (33 to 1.67×10^7 International Units [IU]/mL in EDTA plasma and 738 - 10,000,000 copies/mL (1230 to 1.67×10^7 International Units [IU]/mL) in a PSC dried plasma spot). One copy of HIV-1 RNA is equivalent to 1.67 IU based on the WHO 1st International Standard for HIV-1 RNA for Nucleic Acid-Based Techniques (NAT) (NIBSC 97/656)³⁶.

This test is intended for use in conjunction with clinical presentation and other laboratory markers of disease progress for the clinical management of HIV-1 group M and HIV-1 group O infected patients. The test can be used to assess patient prognosis by measuring the baseline HIV-1 RNA level or to monitor the effects of antiretroviral therapy by measuring changes in HIV-1 RNA levels during the course of antiretroviral treatment.

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 is not intended for use as a screening test for the presence of HIV-1 in blood or blood products or as a diagnostic test to confirm the presence of HIV-1 infection.

SUMMARY AND EXPLANATION OF THE TEST

Human Immunodeficiency Virus (HIV) is the etiologic agent of Acquired Immunodeficiency Syndrome (AIDS)¹⁻³. HIV infection can be transmitted by sexual contact, exposure to infected blood or blood products, or by an infected mother to the fetus⁴. Within three to six weeks of exposure to HIV, infected individuals generally develop a brief, acute syndrome characterized by flu-like symptoms and associated with high levels of viremia in the peripheral blood⁵⁻⁸. In most infected individuals this is followed by an HIV-specific immune response and a decline of plasma viremia, usually within four to six weeks of the onset of symptoms^{9,10}. After seroconversion, infected individuals typically enter a clinically stable, asymptomatic phase that can last for years¹¹⁻¹³. The asymptomatic period is characterized by persistent, low level plasma viremia¹⁴ and a gradual depletion of CD4⁺ T lymphocytes, leading to severe immunodeficiency, multiple opportunistic infections, malignancies and death¹⁵. Although virus levels in the peripheral blood are relatively low during the asymptomatic phase of the infection, virus replication and clearance appear to be dynamic processes in which high rates of virus production and infection of CD4⁺ cells are balanced by equally high rates of virus clearance, death of infected cells and replenishment of CD4⁺ cells, resulting in relatively stable levels of both plasma viremia and CD4⁺ cells¹⁶⁻¹⁸.

Quantitative measurements of HIV viremia in the peripheral blood have shown that higher virus levels may be correlated with increased risk of clinical progression of HIV disease, and that reductions in plasma virus levels may be associated with decreased risk of clinical progression¹⁹⁻²¹. Virus levels in the peripheral blood can be quantitated by measurement of the HIV p24 antigen in serum, by quantitative culture of HIV from

plasma, or by direct measurement of viral RNA in plasma using nucleic acid amplification or signal amplification technologies²²⁻²⁶.

The World Health Organization also now recommends the use of dried spot specimens to expand the reach of viral load testing in resource limited settings without ready access to phlebotomy services or robust EDTA plasma sample transportation capabilities²⁷. A dried plasma spot from the **PSC** which also stabilizes the HIV RNA in dried plasma can improve viral load testing coverage in these settings by enabling sample transportation over longer distances and harsher environmental conditions than EDTA plasma.

p24 antigen is the principal core protein of HIV and is found in serum either free or bound by anti-p24 antibody. Free p24 antigen can be measured with commercially available enzyme immunoassays (EIA), although the usefulness of p24 antigen as a marker of viral load is limited since the antigen is detectable in only 20% of asymptomatic patients and 40-50% of symptomatic patients. Procedures to dissociate antigen-antibody complexes improve the sensitivity of the p24 antigen tests, but the viral protein remains undetectable in most asymptomatic patients²².

Infectious HIV in plasma can be cultured by inoculation into activated peripheral blood mononuclear cells (PBMC) from normal donors. Quantitation is achieved by inoculating PBMC with serial dilutions of the plasma specimen. Quantitative culture has limited utility for monitoring virus levels in infected individuals since only a small fraction of virus particles is infectious *in vitro*. Infectious virus is often undetectable in asymptomatic individuals²².

HIV RNA in plasma can be quantitated by nucleic acid amplification technologies, such as the Polymerase Chain Reaction (PCR)²⁸⁻³⁰. The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 uses PCR technology to achieve maximum sensitivity and dynamic range for the quantitative detection of HIV-1 RNA in EDTA anti-coagulated plasma and in a **PSC** dried plasma spot.

PRINCIPLES OF THE PROCEDURE

The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 is a nucleic acid amplification test for the quantitation of Human Immunodeficiency Virus Type 1 (HIV-1) RNA in human EDTA plasma or from a **PSC** dried plasma spot. The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 is based on three major processes: (1) specimen preparation to isolate HIV-1 RNA; (2) reverse transcription of the target RNA to generate complementary DNA (cDNA), and (3) simultaneous PCR amplification of target cDNA and detection of cleaved dual-labeled oligonucleotide detection probe specific to the target.

The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 permits automated specimen preparation followed by automated reverse transcription, PCR amplification and detection of HIV-1 target RNA and HIV-1 Quantitation Standard (QS) Armored RNA. The Master Mix reagent contains primers and probes specific for both HIV-1 RNA and HIV-1 QS RNA. The Master Mix has been developed to ensure equivalent quantitation of group M subtypes of HIV-1 and of HIV-1 group O. The detection of amplified DNA is performed using target-specific and QS-specific dual-labeled oligonucleotide probes that permit independent identification of HIV-1 amplicon and HIV-1 QS amplicon.

The quantitation of HIV-1 viral RNA is performed using the HIV-1 QS. It compensates for effects of inhibition and controls the preparation and amplification processes, allowing a more accurate quantitation of HIV-1 RNA in each specimen. The HIV-1 QS is a non-infectious Armored RNA construct that contains HIV sequences with identical primer binding sites as the HIV-1 target RNA and a unique probe binding region that allows HIV-1 QS amplicon to be distinguished from HIV-1 target amplicon.

The HIV-1 QS is added to each specimen at a known copy number and is carried through the subsequent steps of specimen preparation, reverse transcription, simultaneous PCR amplification and detection of cleaved dual-labeled oligonucleotide detection probes. The COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer calculates the HIV-1 RNA concentration in the test specimens by comparing the HIV-1 signal to the HIV-1 QS signal for each specimen and control.

Target Selection

Selection of the target RNA sequence for HIV-1 depends on identification of regions within the HIV-1 genome that show maximum sequence conservation among the various HIV-1 group M subtypes and HIV-1 group O specimens. In order to address the high genetic variability of the virus, two regions of HIV genome

are simultaneously targeted for amplification and detection by the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0.

Two target-specific and one QS-specific dual-labeled oligonucleotide probes permit independent identification of the HIV-1 amplicon and of the HIV-1 QS amplicon. Accordingly, the appropriate selection of the primers and the dual-labeled oligonucleotide probes is critical to the ability of the test to amplify and detect the HIV-1 group M subtypes and HIV-1 group O. The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 uses reverse transcription and PCR amplification primers that define sequences within the highly conserved regions of the HIV-1 *gag* gene³³ and of the HIV-1 LTR region.

Specimen Preparation

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 utilizes automated specimen preparation on the COBAS® AmpliPrep Instrument by a generic silica-based capture technique. The procedure processes 850 µL of plasma or SPEX-based extract from the **PSC**. The HIV-1 virus particles are lysed by incubation at elevated temperature with a protease and chaotropic lysis/binding buffer that releases nucleic acids and protects the released HIV-1 RNA from RNases in plasma. Protease and a known number of HIV-1 QS Armored RNA molecules are introduced into each specimen along with the lysis reagent and magnetic glass particles. Subsequently, the mixture is incubated and the HIV-1 RNA and HIV-1 QS RNA are bound to the surface of the magnetic glass particles. Unbound substances, such as salts, proteins and other cellular impurities, are removed by washing the magnetic glass particles. After separating the magnetic glass particles and completing the washing steps, the adsorbed nucleic acids are eluted at elevated temperature with an aqueous solution. The processed specimen, containing the magnetic glass particles as well as released HIV-1 RNA and HIV-1 QS RNA, is added to the amplification mixture and transferred to the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer. The HIV-1 target RNA and the HIV-1 QS RNA are then reverse transcribed, amplified and simultaneously detected by cleavage of two target-specific and one QS-specific dual-labeled oligonucleotide probe.

Reverse Transcription and PCR Amplification

The reverse transcription and PCR amplification reaction is performed with the thermostable recombinant enzyme *Thermus specie* Z05 DNA Polymerase (Z05). In the presence of manganese (Mn^{2+}) and under the appropriate buffer conditions, Z05 has both reverse transcriptase and DNA polymerase activity^{31,32}. This allows both reverse transcription and PCR amplification to occur together with real-time detection of the amplicon.

Processed specimens are added to the amplification mixture in amplification tubes (K-tubes) in which both reverse transcription and PCR amplification occur. The reaction mixture is heated to allow the downstream primers to anneal specifically to the HIV-1 target RNA and to the HIV-1 QS RNA. In the presence of Mn^{2+} and excess deoxynucleotide triphosphates (dNTPs), including deoxyadenosine, deoxyguanosine, deoxycytidine, deoxyuridine and deoxythymidine triphosphates, Z05 polymerase extends the annealed primers forming DNA strands complementary to the RNA target.

Target Amplification

Processed specimens are added to the amplification mixture in amplification tubes (K-tubes) in which PCR amplification occurs. Following reverse transcription of the HIV-1 target RNA and the HIV-1 QS RNA, the Thermal Cycler in the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer heats the reaction mixture to denature the RNA:cDNA hybrids and to expose the specific primer target sequences. As the mixture cools, the primers anneal to the target DNA. Z05 in the presence of Mn^{2+} and excess deoxynucleotide triphosphates (dNTPs), including deoxyadenosine, deoxyguanosine, deoxycytidine, deoxyuridine and deoxythymidine triphosphates, extends the annealed primers along the target template to produce double-stranded DNA molecules termed amplicons. The COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer automatically repeats this process for a designated number of cycles, with each cycle intended to double the amount of amplicon DNA. The required number of cycles is preprogrammed into the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer. Amplification occurs only in the two regions of the HIV-1 genome between the primers; the entire HIV-1 genome is not amplified.

Selective Amplification

Selective amplification of target nucleic acid from the specimen is achieved in the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 by the use of AmpErase (uracil-N-glycosylase) enzyme and

deoxyuridine triphosphate (dUTP). The AmpErase enzyme recognizes and catalyzes the destruction of DNA strands containing deoxyuridine³⁴, but not DNA containing deoxythymidine.

Deoxyuridine is not present in naturally occurring DNA, but is always present in amplicon due to the use of deoxyuridine triphosphate as one of the dNTPs in the Master Mix reagent; therefore, only amplicon contains deoxyuridine. Deoxyuridine renders contaminating amplicon susceptible to destruction by the AmpErase enzyme prior to amplification of the target DNA. Also, any nonspecific product formed after initial activation of the Master Mix by manganese is destroyed by the AmpErase enzyme. The AmpErase enzyme, which is included in the Master Mix reagent, catalyzes the cleavage of deoxyuridine-containing DNA at the deoxyuridine residues by opening the deoxyribose chain at the C1-position. When heated in the first thermal cycling step, the amplicon DNA chain breaks at the position of the deoxyuridine, thereby rendering the DNA non-amplifiable. The AmpErase enzyme remains inactive for a prolonged period of time once exposed to temperatures above 55°C, i.e. throughout the thermal cycling steps, and therefore does not destroy target amplicon formed during amplification.

Detection of PCR Products in a COBAS[®] TaqMan[®] Test

The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 utilizes real-time^{35,36} PCR technology. The use of dual-labeled fluorescent probes allows for real-time detection of PCR product accumulation by monitoring of the emission intensity of fluorescent reporter dyes released during the amplification process. The probes consist of HIV-1 and HIV-1 QS-specific oligonucleotide probes with a reporter dye and a quencher dye. In the COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 the HIV-1 and HIV-1 QS probes are labeled with different fluorescent reporter dyes. When these probes are intact, the fluorescence of the reporter dye is suppressed by the proximity of the quencher dye due to Förster-type energy transfer effects. During PCR, the probe hybridizes to a target sequence and is cleaved by the 5' → 3' nuclease activity of the thermostable Z05 DNA polymerase. Once the reporter and quencher dyes are released and separated, quenching no longer occurs, and the fluorescent activity of the reporter dye is increased. The amplification of HIV-1 RNA and HIV-1 QS RNA are measured independently at different wavelengths. This process is repeated for a designated number of cycles, each cycle effectively increasing the emission intensity of the individual reporter dyes, permitting independent identification of HIV-1 RNA and HIV-1 QS RNA. The PCR cycle where a growth curve starts exponential growth is related to the amount of starting material at the beginning of the PCR.

Fundamentals of COBAS[®] TaqMan[®] Test Quantitation

The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 is inherently quantitative over a very wide dynamic range since the monitoring of amplicon is performed during the exponential phase of amplification. The higher the HIV-1 titer of a specimen, the earlier the fluorescence of the reporter dye of the HIV-1 probes rises above the baseline fluorescence level (see Figure 1). Since the amount of HIV-1 QS RNA is constant between all specimens, the fluorescence of the reporter dye of the HIV-1 QS probe should appear at a similar cycle for all specimens (see Figure 2). In specimens where the QS fluorescence is affected, the concentration is adjusted accordingly. The appearance of the specific fluorescent signals is reported as a critical threshold value (Ct). The Ct is defined as the fractional cycle number where reporter dye fluorescence exceeds a predetermined threshold (the Assigned Fluorescence Level), and starts the exponential growth phase of this signal (see Figure 3). A higher Ct value indicates a lower titer of initial HIV-1 target material. A 2-fold increase in titer correlates with a decrease of 1 Ct for target HIV-1 RNA, while a 10-fold increase in titer correlates with a decrease of 3.3 Ct.

Figure 1 shows the target growth curves for a dilution series spanning a 5-log₁₀ range. As the concentration of the virus increases, the growth curves shift to earlier cycles. Therefore, the leftmost growth curve corresponds to the highest viral titer level, whereas, the rightmost growth curve corresponds to the lowest viral titer level.

Figure 1
Target Growth Curves for a Dilution Series Spanning Over a 5- \log_{10} Range




Figure 2 shows the Quantitation Standard growth curves for specimens from a viral dilution series that spans a 5- \log_{10} range. The amount of Quantitation Standard added to each specimen is constant for each reaction. The Ct value of the Quantitation Standard is similar regardless of the viral titer.

Figure 2
Quantitation Standard Growth Curves for a Dilution Series of Virus Spanning a 5- \log_{10} Range





Figure 3 provides an example of how the fluorescence values at every cycle are normalized for each growth curve. The fractional cycle number (Ct) is calculated where the fluorescence signal crosses the Assigned Fluorescence Level.

Figure 3
Fluorescence Values at Every Cycle are Normalized for Each Growth Curve



HIV-1 RNA Quantitation

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 quantitates HIV-1 viral RNA by utilizing a second target sequence (HIV-1 Quantitation Standard) that is added to each test specimen at a known concentration. The HIV-1 QS is a non-infectious Armored RNA construct, containing fragments of HIV-1 sequences with primer binding regions identical to those of the HIV-1 *gag* target sequence. The HIV-1 QS contains HIV-1 primer binding regions and generates an amplification product of the same length and base composition as the HIV-1 *gag* target RNA. The detection probe binding region of the HIV-1 QS has been modified to differentiate HIV-1 QS amplicon from HIV-1 *gag* target amplicon.

During the annealing phase of the PCR in the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer, the specimens are illuminated and excited by filtered light and filtered emission fluorescence data are collected for each specimen. The readings from each specimen are then corrected for instrumental fluctuations. These fluorescence readings are sent by the instrument to the AMPLILINK software and stored in a database. Pre-Checks are used to determine if the HIV-1 RNA and HIV-1 QS RNA data represent sets that are valid, and flags are generated when the data lie outside the preset limits. After all Pre-Checks are completed and passed, the fluorescence readings are processed to generate Ct values for the HIV-1 RNA and the HIV-1 QS RNA. The lot-specific calibration constants provided with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 are used to calculate the titer value for the specimens and controls based upon the HIV-1 RNA and HIV-1 QS RNA Ct values. The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 is standardized against a World Health Organization International Standard for HIV-1 RNA. Titer results are reported in copies/mL (cp/mL) or International Units (IU/mL). The conversion factor between reported HIV-1 RNA cp/mL and HIV-1 IU/mL has been determined by Roche Molecular Systems, Inc. to be 0.6 cp/IU (1.67 IU/cp).

REAGENTS**COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0**
(P/N: 05212294 190)**H12CAP****48 Tests****HIV-1 v2.0 CS1****1 x 48 Tests**

(HIV-1 Magnetic Glass Particles Reagent Cassette)

Magnetic glass particles
93% Isopropanol**HIV-1 v2.0 CS2****1 x 48 Tests**

(HIV-1 Lysis Reagent Cassette)

Sodium citrate dihydrate
42.5% Guanidine thiocyanate
< 14% Polydocanol
0.9% Dithiothreitol**HIV-1 v2.0 CS3****1 x 48 Tests**

HIV-1 Multi-Reagent Cassette containing:

Pase

1 x 3.8 mL

(Proteinase Solution)

Tris buffer
< 0.05% EDTA
Calcium chloride
Calcium acetate
≤ 7.8% Proteinase
Glycerol**EB**

1 x 7.0 mL

(Elution Buffer)

Tris-base buffer
0.2% Methylparaben**HIV-1 v2.0 CS4****1 x 48 Tests**

HIV-1 Test-Specific Reagent Cassette containing:

HIV-1 QS

1 x 3.6 mL

(HIV-1 Quantitation Standard)

Tris-HCl buffer
EDTA
< 0.005% Poly rA RNA (synthetic)
< 0.001% Armored HIV-1 RNA construct containing HIV-1
primer binding sequences and a unique probe binding
region (non-infectious RNA in MS2 bacteriophage)
0.05% Sodium azide**HIV-1 MMX**

1 x 2.5 mL

(HIV-1 Master Mix)

Tricine buffer
Potassium acetate
Potassium hydroxide
20% Dimethylsulfoxide
Glycerol
< 0.04% dATP, dCTP, dGTP, dUTP, dTTP
< 0.003% Upstream and downstream primers to the *gag* and
the LTR region of HIV-1
< 0.003% Oligonucleotide aptamer

- < 0.003% Fluorescent-labeled oligonucleotide probes specific for HIV-1 and the HIV-1 Quantitation Standard
- < 0.05% Z05 DNA Polymerase (microbial)
- < 0.1% AmpErase (uracil-N-glycosylase) enzyme (microbial)
- 0.09% Sodium azide

CAP/CTM Mn²⁺

(CAP/CTM Manganese Solution)

- < 0.5% Manganese acetate
- Glacial acetic acid
- 0.09% Sodium azide

1 x 19.8 mL

HIV-1 H(+), v2.0

(HIV-1 High Positive Control, v2.0)

- < 0.001% Armored HIV-1 RNA construct containing HIV-1 sequences (non-infectious RNA in MS2 bacteriophage).

Negative Human Plasma, non-reactive by tests for antibody to HCV, antibody to HIV-1/2, HIV p24 antigen and HBsAg; HIV-1 RNA, HCV RNA and HBV DNA not detectable by PCR methods

- 0.1% ProClin[®] 300 preservative

4 x 1.0 mL

HIV-1 L(+), v2.0

(HIV-1 Low Positive Control, v2.0)

- < 0.001% Armored HIV-1 RNA construct containing HIV-1 sequences (non-infectious RNA in MS2 bacteriophage).

Negative Human Plasma, non-reactive by tests for antibody to HCV, antibody to HIV-1/2, HIV p24 antigen and HBsAg; HIV-1 RNA, HCV RNA and HBV DNA not detectable by PCR methods

- 0.1% ProClin[®] 300 preservative

4 x 1.0 mL

CTM (-) C

[(COBAS[®] TaqMan[™] Negative Control (Human Plasma))

Negative Human Plasma, non-reactive by tests for antibody to HCV, antibody to HIV-1/2, HIV p24 antigen and HBsAg; HIV-1 RNA, HCV RNA and HBV DNA not detectable by PCR methods

- 0.1% ProClin[®] 300 preservative

4 x 1.0 mL

HIV-1 H(+), v2.0 Clip

(HIV-1 High Positive Control, v2.0 Barcode Clip)

1 x 4 Clips

HIV-1 L(+), v2.0 Clip

(HIV-1 Low Positive Control, v2.0 Barcode Clip)

1 x 4 Clips

HIV-1 (-) C Clip

(HIV-1 Negative Control, v2.0 Barcode Clip)

1 x 4 Clips

COBAS[®] AmpliPrep/COBAS[®] TaqMan[™] Wash Reagent
(P/N: 03587797 190)

PG WR

1 x 5.1 L

PG WR

(COBAS[®] AmpliPrep/COBAS[®] TaqMan[™] Wash Reagent)

- Sodium citrate dihydrate
- < 0.1% N-Methylisothiazolone-HCl

SPEX

(cobas® Specimen Pre-Extraction Reagent)

42.5% guanidine thiocyanate

0.79% Sodium citrate

0.01% Citric acid

3.6% Polydocanol

1.8% Dithiothreitol

Note: This reagent is optional and should only to be used in conjunction with the PSC to generate dried plasma spot samples. See PSC Method Sheet ms_07963084190.

WARNINGS AND PRECAUTIONS

A. FOR IN VITRO DIAGNOSTIC USE.

B. This test is for use with human plasma collected in the anticoagulant EDTA or from a **PSC** dried plasma spot.

C. Do not pipette by mouth.

D. Do not eat, drink or smoke in laboratory work areas. Wear protective disposable gloves, laboratory coats and eye protection when handling specimens and kit reagents. Wash hands thoroughly after handling specimens and test reagents.

E. **Avoid microbial and ribonuclease contamination of reagents when removing aliquots from control vials.**

F. **The use of sterile disposable pipettes and RNase-free pipette tips is recommended.**

G. Do not pool controls from different lots or from different bottles of the same lot.

H. Do not mix reagent cassettes or controls from different kits.

I. Do not open COBAS® AmpliPrep cassettes and exchange, mix, remove or add bottles.

J. Dispose of unused reagents, waste and specimens in accordance with country, federal, state and local regulations.

K. Do not use a kit after its expiration date.

L. Safety Data Sheets (SDS) are available on request from your local Roche office.

M. Specimens and controls should be handled as if infectious using safe laboratory procedures such as those outlined in *Biosafety in Microbiological and Biomedical Laboratories*³⁸ and in the CLSI Document M29-A3³⁹. Thoroughly clean and disinfect all work surfaces with a freshly prepared solution of 0.5% sodium hypochlorite in deionized or distilled water.

Note: Commercial liquid household bleach typically contains sodium hypochlorite at a concentration of 5.25%. A 1:10 dilution of household bleach will produce a 0.5% sodium hypochlorite solution.

Note: If spillage of PSC dried plasma spot samples in cobas® Specimen Pre-Extraction Reagent (SPEX) (which contain guanidine thiocyanate) occurs, do not allow it to come in contact with sodium hypochlorite containing disinfectants such as bleach. This mixture can produce a highly toxic gas.

N. **CAUTION: CTM (-) C, HIV-1 L(+), v2.0 and HIV-1 H(+), v2.0** contain Human Plasma derived from human blood. The source material has been tested and found non-reactive for the presence of Hepatitis B Surface Antigen (HBsAg), antibodies to HIV-1/2 and HCV, and HIV p24 Antigen. Testing of Negative Human Plasma by PCR methods showed no detectable HIV-1 RNA, HCV RNA or HBV DNA. No known test methods can offer complete assurance that products derived from human blood will not transmit infectious agents. Therefore, all human sourced

material should be considered potentially infectious. **CTM (-) C, HIV-1 L(+)**C**, v2.0** and **HIV-1 H(+)**C**, v2.0** should be handled as if infectious using safe laboratory procedures such as those outlined in *Biosafety in Microbiological and Biomedical Laboratories*³⁸ and in the CLSI Document M29-A3³⁹. Thoroughly clean and disinfect all work surfaces with a freshly prepared solution of 0.5% sodium hypochlorite in deionized or distilled water.

- O. **HIV-1 QS, CAP/CTM Mn²⁺** and **HIV-1 MMX** contain sodium azide. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. While disposing of sodium azide-containing solutions down laboratory sinks, flush the drains with a large volume of water to prevent azide buildup.
- P. Wear eye protection, laboratory coats and disposable gloves when handling any reagent. Avoid contact of these materials with the skin, eyes or mucous membranes. If contact does occur, immediately wash with large amounts of water. Burns can occur if left untreated. If spills of these reagents occur, dilute with water before wiping dry.
- Q. Do not allow **HIV-1 v2.0 CS2, SPEX** (used in **PSC** dried plasma spot procedure) and liquid waste from the COBAS[®] AmpliPrep Instrument, which contain guanidine thiocyanate, to contact sodium hypochlorite (bleach) solution. These mixtures can produce a highly toxic gas.
- R. **SPEX** is light sensitive and shipped in light protective bottles.
- S. When disposing of used COBAS[®] AmpliPrep Sample Processing Units (SPUs), which contain guanidine thiocyanate, avoid any contact with sodium hypochlorite (bleach) solution. These mixtures can produce a highly toxic gas.
- T. Refer to **PSC** Method Sheet ms_07963084190 for additional warnings and precautions.

STORAGE AND HANDLING REQUIREMENTS

- A. **Do not freeze reagents or controls.**
- B. Before use, visually inspect each reagent cassette and vial to ensure that there are no signs of leakage. If there is any evidence of leakage, do not use that material for testing.
- C. Store **HIV-1 v2.0 CS1, HIV-1 v2.0 CS2, HIV-1 v2.0 CS3** and **HIV-1 v2.0 CS4** at 2-8°C. Unused, these reagents are stable until the expiration date indicated. Once used, these reagents are stable for 28 days at 2-8°C or until the expiration date, whichever comes first. **HIV-1 v2.0 CS1, HIV-1 v2.0 CS2, HIV-1 v2.0 CS3** and **HIV-1 v2.0 CS4** can be used for a maximum of 4 instrument cycles, up to a maximum of 64 hours cumulative on board the COBAS[®] AmpliPrep Instrument. Reagents must be stored at 2-8°C between instrument cycles.
- D. Store **HIV-1 H(+)**C**, v2.0, HIV-1 L(+)**C**, v2.0** and **CTM (-) C** at 2-8°C. The controls are stable until the expiration date indicated. Once opened, any unused portion must be discarded.
- E. Store Barcode clips [**HIV-1 H(+)**C**, v2.0 Clip, HIV-1 L(+)**C**, v2.0 Clip** and **HIV-1 (-) C Clip**] at 2-30°C.
- F. Store **PG WR** at 2-30°C. **PG WR** is stable until the expiration date indicated. Once opened, this reagent is stable for 28 days at 2-30°C or until the expiration date, whichever comes first.
- G. Store **SPEX** (used in **PSC** dried plasma spot procedure) at 2-8°C. **SPEX** is stable until the expiration date indicated. Once opened, this reagent is stable for 28 days at 2-30°C or until the expiration date, whichever comes first.
- H. **PSC** storage and handling requirements are specified in the **PSC** Method Sheet ms_07963084190.

MATERIALS PROVIDED

- A. **COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0**

HI2CAP

(P/N: 05212294 190)

HIV-1 v2.0 CS1

(HIV-1 Magnetic Glass Particles Reagent Cassette)

HIV-1 v2.0 CS2

(HIV-1 Lysis Reagent Cassette)

HIV-1 v2.0 CS3

(HIV-1 Multi-Reagent Cassette)

HIV-1 v2.0 CS4

(HIV-1 Test-Specific Reagent Cassette)

HIV-1 H(+)C, v2.0

(HIV-1 High Positive Control, v2.0)

HIV-1 L(+)C, v2.0

(HIV-1 Low Positive Control, v2.0)

CTM (-) C

[COBAS® TaqMan® Negative Control (Human Plasma)]

HIV-1 H(+)C, v2.0 Clip

(HIV-1 High Positive Control, v2.0 Barcode Clip)

HIV-1 L(+)C, v2.0 Clip

(HIV-1 Low Positive Control, v2.0 Barcode Clip)

HIV-1 (-) C Clip

(HIV-1 Negative Control Barcode Clip)

- B. **COBAS® AmpliPrep/COBAS® TaqMan® Wash Reagent**

PG WR

(P/N: 03587797 190)

- C. **COBAS® AmpliPrep/COBAS® TaqMan® Specimen Pre-Extraction Reagent**

SPEX

(P/N: 06989861190)

MATERIALS REQUIRED BUT NOT PROVIDED

Instrumentation and Software

- COBAS[®] AmpliPrep Instrument
- COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer
- Optional: **cobas p** 630 Instrument
- Optional: Docking Station
- AMPLILINK Software Version 3.3 or Version 3.4 Series
- Control Unit for the AMPLILINK Software, with printer
- Instrument and Software Manuals:
 - COBAS[®] AmpliPrep Instrument Manual for use with the AMPLILINK Software Version 3.3 and 3.4 Series
 - COBAS[®] TaqMan[®] Analyzer Instrument Manual for use with the AMPLILINK Software Version 3.3 and 3.4 Series
 - COBAS[®] TaqMan[®] 48 Analyzer Instrument Manual for use with the AMPLILINK Software Version 3.3 and 3.4 Series
 - AMPLILINK Software Version 3.3 Series Application Manual for use with COBAS[®] AmpliPrep Instrument, COBAS[®] TaqMan[®] Analyzer, COBAS[®] TaqMan[®] 48 Analyzer, COBAS[®] AMPLICOR[®] Analyzer and **cobas p** 630 Instrumentor
 - AMPLILINK Software Version 3.4 Series Application Manual
 - Optional: **cobas p** 630 Instrument Operator's Manual Software Version 2.2
- Test Definition File (TDF). See product information card, provided with the kit, for name and current version of the TDF.

Disposables

- Sample processing units: SPUs
- Sample input tubes (S-tubes) with barcode clips
- Racks of K-tips
- K-tube Box of 12 x 96

OTHER MATERIALS REQUIRED (FOR EDTA PLASMA SAMPLE APPLICATION ONLY) BUT NOT PROVIDED

- Sample Rack (SK 24 rack)
- Reagent Rack
- SPU rack
- K-tube capper, motorized
- K-tube capper
- K-carrier
- K-carrier Transporter
- K-carrier rack
- Pipettors with aerosol barrier or positive displacement RNase-free tips (capacity 1000 µL)*
- Disposable gloves, powderless
- Vortex mixer

* Pipettors should be accurate within 3% of stated volume. Aerosol barrier or positive displacement RNase-free tips must be used where specified to prevent specimen and amplicon cross-contamination.

OTHER MATERIALS AND CONSUMABLES REQUIRED (FOR PSC DRIED PLASMA SPOT SAMPLE APPLICATION ONLY) BUT NOT PROVIDED

- **cobas**[®] Plasma Separation Card*
- Sterile or disposable forceps or tweezers**
- 140 µL capillary (e.g., Vitrex plastic tube) with compatible dispenser (e.g., Vitrex pipette holder)*
- Single Use lancing device (e.g., Greiner Bio-one: MiniCollect[®] Safety Lancet penetration depth 2.00 mm)*
- Sample bag (plastic transparent resealable ziplock) and silica gel desiccant sachets (for a total of 4 grams. For **PSC** storage and delivery, see **PSC** Method Sheet ms_07963084190 for more information).
- Transport bag (e.g., Wicoseal 180 x 60 x 240 mm)
- Pipette (e.g., Multistep[®] pipette)
- Eppendorf Thermomixer[®] (e.g., model R 5355 or C or equivalent) with Thermoblock for 24 cryo tubes

* See **PSC** Method Sheet ms_07963084190 for more information about the **PSC** sample collection.

** To prevent cross-contamination, use only one pair of forceps for each patient! The usage of metal forceps that are autoclaved after single use is recommended.

EDTA PLASMA SPECIMEN COLLECTION, TRANSPORT AND STORAGE

Note: Handle all specimens and controls as if they are capable of transmitting infectious agents.


Note: This test has been validated for use with only human plasma collected in EDTA anticoagulant or from a PSC dried plasma spot. Testing of other specimen types may result in inaccurate results.

A. Specimen Collection

The COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 is for use with plasma specimens. Blood should be collected in sterile tubes using EDTA as the anticoagulant and mixed adequately according to the tube manufacturer's instructions.

Store whole blood at 2-25°C for no longer than 24 hours. Separate plasma from whole blood within 24 hours of collection by centrifugation at 800-1600 x g for 20 minutes at room temperature. Transfer plasma to a sterile polypropylene tube. Figure 4 shows specimen stability data from specimen stability studies performed with the COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test (P/N: 03543005 190).

Figure 4
HIV-1 Stability in Whole Blood (Collected in EDTA-Plasma Tubes)




B. Specimen Transport

Transportation of whole blood or plasma must comply with country, federal, state and local regulations for the transport of etiologic agents⁴⁰. Whole blood must be transported at 2-25°C and centrifuged within 24 hours of collection. Plasma may be transported at 2-8°C or frozen at -20°C to -80°C.

C. Specimen Storage


Plasma specimens may be stored at room temperature (25-30°C) for up to 1 day or at 2-8°C for up to 6 days. Plasma specimens were shown to be stable for six weeks if frozen at -20°C to -80°C. It is recommended that specimens be stored in 1100-1200 µL aliquots in sterile, 2.0 mL polypropylene screw-cap tubes (such as Sarstedt 72.694.006). Figure 5 shows the specimen stability data from specimen storage studies performed with the COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test (P/N: 03543005 190).

Figure 5
HIV-1 Stability in EDTA Plasma



Plasma specimens may be frozen and thawed up to five times without a significant loss of HIV-1 RNA. Figure 6 shows the data from a freeze-thaw study performed with the COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test (P/N: 03543005 190).

Figure 6
HIV-1 Results After Up to Five Freeze-Thaw (F-T) Cycles (EDTA-Plasma)



PSC DRIED PLASMA SPOT SAMPLE COLLECTION, TRANSPORT AND STORAGE

Note: Handle all specimens and controls as if they are capable of transmitting infectious agents.


Note: This test has been validated for use with only human plasma collected in EDTA anticoagulant or from a PSC dried plasma spot. Testing of other specimen types may result in inaccurate results.

A. Sample Collection

PSC dried plasma spot samples are collected by appropriate clinical procedures. The expiry date of the **PSC** must be checked before use. Proceed only if the **PSC** has not expired yet and if the bag in which the **PSC** is sealed is completely closed and intact. Label the **PSC** with patient's name, date of birth, date and time of sample collection. Apply 140 μ L of whole blood on each circle of the **PSC** membrane delineated by the spotting layer using an appropriate capillary and a dispenser. It is recommended to fill all three spots on the **PSC**, in order to allow retesting. Do not apply samples from more than one patient on the same **PSC**. Ensure that BOTH sides of the **PSC** spots (front: membrane with blood; back: spot with plasma) are saturated after 5 minutes. Check the back side through the transparent back layer. Do not allow the membranes to get in contact with gloves, tools or any potentially contaminated surfaces during this process.

Allow the **PSC** to dry at room temperature for at least 4 hours (to maximum overnight), protecting it from direct sunlight. Do not remove the spotting layer. This is done at the laboratory. After drying, store the **PSC** in an individual sample bag with 4 grams of desiccant and seal the bag (Figure 7). The collected sample bags must be packed in a transport bag together with their relative Patient Information Sheet. It is recommended to pack a maximum number of 25 **PSCs** per transport bag.

Figure 7
Overview of the Packaging Concept for PSC Transport



B. Sample Transport

If samples are to be shipped, they should be packaged and labeled in compliance with applicable country and/or international regulations covering the transport of samples and etiologic agents. Transport bags containing PSCs must be transported within 28 days at 18–45°C and up to 85% humidity.

C. Sample Storage

PSCs in individual sample bags with 4 grams of desiccant, within a transport bag may be stored after transportation at room temperature (18–30°C), at 2–8°C or at ≤–10°C for up to 56 days (with and without layer separation).

INSTRUCTIONS FOR USE

Note: For detailed operating instructions, a detailed description of the possible configurations, printing results and interpreting flags, comments and error messages, refer to: (1) the COBAS® AmpliPrep Instrument Manual for use with the AMPLILINK Software Version 3.3 and 3.4 Series; (2) the COBAS® TaqMan® Analyzer Instrument Manual for use with the AMPLILINK Software Version 3.3 and 3.4 Series; (3) the COBAS® TaqMan® 48 Analyzer Instrument Manual for use with the AMPLILINK Software Version 3.3 and 3.4 Series; (4) the AMPLILINK Software Version 3.3 Series Application Manual for use with COBAS® AmpliPrep Instrument,

COBAS® TaqMan® Analyzer, COBAS® Taqman® 48 Analyzer, COBAS® AMPLICOR® Analyzer and cobas p 630 Instrument or the AMPLILINK Software Version 3.4 Series Application Manual; (5) Optional: cobas p 630 Instrument Operator's Manual Software Version 2.2.

Information contained in Part A-C and F-J is general for both sample types (EDTA plasma and PSC dried plasma spot samples). Part D contains information about ordering and loading of specimens for EDTA plasma samples only. Part E contains information about ordering and loading of specimens for PSC dried plasma spot samples only.

Batch Size

Each kit contains reagents sufficient for 48 tests, which may be performed in batches of 12 to 24 tests. At least one of each control **[CTM (-) C, HIV-1 L(+)**C**, v2.0** and **HIV-1 H(+)**C**, v2.0]** must be included in each batch (see "Quality Control" section).

Workflow

The COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer run must be started within 120 minutes following completion of specimen and control preparation.

Note: Do not freeze or store processed specimens and controls at 2-8°C.

Specimen and Control Preparation

Note: If using frozen specimens, place the specimens at room temperature until completely thawed and vortex for 3-5 seconds before use. Controls should be removed from 2-8°C storage and equilibrated to ambient temperature before use.

COBAS® AmpliPrep Instrument Set-up

Part A. Maintenance and Priming

- A1. The COBAS® AmpliPrep Instrument is ready for operation in stand-by mode.
- A2. Turn the Control Unit for the AMPLILINK software **ON**. Prepare the Control Unit as follows:
 1. Log onto the Microsoft Windows Operating System.
 2. Double click the AMPLILINK software icon.
 3. Log onto AMPLILINK software by entering the assigned User ID and password.
- A3. Check the supply of **PG WR** using the **Status** Screen and replace if necessary.
- A4. Perform all Maintenance that is listed in the Due Tab. The COBAS® AmpliPrep Instrument will automatically prime the system.

Part B. Loading of Reagent Cassettes

Note: All reagent cassettes should be removed from 2-8°C storage, immediately loaded onto the COBAS® AmpliPrep Instrument and allowed to equilibrate to ambient temperature on the instrument for at least 30 minutes before the first specimen is to be processed. Do not let reagent cassettes come to ambient temperature outside the instrument as condensation may form on the barcode labels. Do not wipe off condensation if it appears on the barcode labels.

- B1. Place **HIV-1 v2.0 CS1** onto a reagent rack. Place **HIV-1 v2.0 CS2**, **HIV-1 v2.0 CS3** and **HIV-1 v2.0 CS4** onto a separate reagent rack.
- B2. Load the reagent rack containing **HIV-1 v2.0 CS1** onto rack position **A** of the COBAS® AmpliPrep Instrument.
- B3. Load the reagent rack containing **HIV-1 v2.0 CS2**, **HIV-1 v2.0 CS3** and **HIV-1 v2.0 CS4** onto rack position **B, C, D** or **E** of the COBAS® AmpliPrep Instrument. (See Table 1 for additional information).

Part C. Loading of Disposables

Note: Determine the number of COBAS® AmpliPrep reagent cassettes, Sample Processing Units (SPUs), Input Sample tubes (S-tubes), K-tips and K-tubes needed. One SPU, one Input S-tube, one K-tip and one K-tube are needed for each specimen or control.

Multiple workflows for use of the COBAS® AmpliPrep Instrument with the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer are possible. For reference, see Table 1 below. Depending on the workflow used, load the appropriate number of reagent cassette racks, sample racks with Input S-tubes, SPU racks, K-tip racks, K-tube racks and K-carriers on K-carrier racks onto the respective rack positions of the COBAS® AmpliPrep Instrument (see Table 1 for additional information).

- C1. Place the SPUs in the SPU rack(s) and load the rack(s) onto rack position **J, K or L** of the COBAS® AmpliPrep Instrument.
- C2. Depending on the workflow used, load full K-tube rack(s) onto rack position **M, N, O or P** of the COBAS® AmpliPrep Instrument.
- C3. Load full K-tip rack(s) onto rack position **M, N, O or P** of the COBAS® AmpliPrep Instrument.
- C4. For workflow 3 using the COBAS® TaqMan® 48 Analyzer, load K-carriers on K-carrier rack(s) onto rack position **M & N, or O & P** of the COBAS® AmpliPrep Instrument.

Table 1
Possible Workflows for Using COBAS® AmpliPrep Instrument with
COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer

Workflow		Transfer Mode to COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer	Racks, Carriers and Disposables	Position on COBAS® AmpliPrep Instrument
1	COBAS® AmpliPrep Instrument plus Docking Station plus COBAS® TaqMan® Analyzer	Automated transfer of K-carrier	K-tubes in full K-tube racks	M - P
			K-tips in full K-tip racks	M - P
			Input S-tubes containing specimens and controls on sample racks	F - H
			SPUs in SPU racks	J - L
			CS1 on Cassette rack	A
			CS2, CS3, CS4 on Cassette rack	B - E
2	COBAS® AmpliPrep Instrument plus COBAS® TaqMan® Analyzer	Manual transfer of K-tubes via sample rack(s) onto COBAS® TaqMan® Analyzer	K-tubes in full K-tube racks	M - P
			K-tips in full K-tip racks	M - P
			Input S-tubes containing specimens and controls on sample racks	F - H
			SPUs in SPU racks	J - L
			CS1 on Cassette rack	A
			CS2, CS3, CS4 on Cassette rack	B - E
			<u>After specimen processing is finished:</u> K-tubes on sample racks (ready for manual transfer)	Same as above (F - H)
3	COBAS® AmpliPrep Instrument plus COBAS® TaqMan® 48 Analyzer(s)	Manual transfer of K-carrier via K-carrier rack(s) onto COBAS® TaqMan® 48 Analyzer	K-tubes on sample racks	F - H
			K-tips in full K-tip racks	M - P
			Input S-tubes containing specimens and controls on sample racks	F - H
			SPUs in SPU racks	J - L
			CS1 on Cassette rack	A
			CS2, CS3, CS4 on Cassette rack	B - E
			Empty barcoded K-carrier on K-carrier rack	M - P
			<u>After specimen processing is finished:</u> K-tubes in K-carrier on K-carrier rack	Same as above (M - P)

Part D. Ordering and Loading of Specimens: EDTA Plasma Samples

- D1. Prepare sample racks as follows: Attach a barcode label clip to each sample rack position where a specimen (S-tube) is to be placed. Attach one of the specific barcode label clips for the controls [**CTM (-) C, HIV-1 L(+)**C, v2.0**** and **HIV-1 H(+)**C, v2.0****] to each sample rack position where the controls (S-tube) are to be placed. The barcode label clips for controls should have the same control lot number as the lot number on the control vials in the kit. Take care in assigning the right control to the position with the appropriate control barcode clip. Place one Input S-tube into each position containing a barcode label clip.
- D2. Using the AMPLILINK software, select the test definition file (HI2CAP96 for COBAS[®] AmpliPrep Instrument plus COBAS[®] TaqMan[®] Analyzer, HI2CAP48 for COBAS[®] AmpliPrep Instrument plus COBAS[®] TaqMan[®] 48 Analyzer), create specimen orders for each specimen and control in the **Orders** window **Sample** folder and complete by saving.
- D3. Assign specimen and control orders to sample rack positions in the **Orders** window **Sample Rack** folder. The sample rack number must be for the rack prepared in Step D1.
- D4. Print the **Sample Rack Order** report to use as a worksheet.
- D5. Prepare specimen and control racks in the designated area for specimen and control addition as follows: Vortex each specimen and control [**CTM (-) C, HIV-1 L(+)**C, v2.0**** and **HIV-1 H(+)**C, v2.0****] for 3 to 5 seconds. Avoid contaminating gloves when manipulating the specimens and controls.
- D6. Transfer 1000 to 1050 µL of each specimen and control [**CTM (-) C, HIV-1 L(+)**C, v2.0**** and **HIV-1 H(+)**C, v2.0****] to the appropriate barcode labeled Input S-tube using a micropipettor with an aerosol barrier or positive displacement RNase-free tip. **Avoid transferring particulates and/or fibrin clots from the original specimen to the Input S-tube.** Specimens and controls should be transferred to tube positions as assigned and recorded on the worksheet in Step D4. The barcode label clips for controls should have the same control lot number as the lot number on the control vials in the kit. Assign the right control to the position with the appropriate control barcode clip. **Avoid contaminating the upper part of the S-tubes with specimens or controls.** If using the **cobas p 630** Instrument for preparation of specimens, refer to the **cobas p 630** Instrument Operator's Manual.
- D7. For workflows 1 and 2, load the sample rack(s) filled with Input S-tubes onto rack positions **F, G** or **H** of the COBAS[®] AmpliPrep Instrument.
- D8. For workflow 3 using the COBAS[®] TaqMan[®] 48 Analyzer, load sample rack(s) with Input S-tubes and K-tubes (one for each Input S-tube, loaded in the right position adjacent to Input S-tubes) onto rack position **F, G** or **H** of the COBAS[®] AmpliPrep Instrument.

Part E. Ordering and Loading of Specimens: PSC Dried Plasma Spot Samples


- E1. Check the integrity of the transport bag before opening. Proceed only if the transport bag is completely sealed.

Open the transport bag and, for each sample bag, proceed only if:

- The laboratory request form is completely filled.
- The barcode of the laboratory request form and the **PSC** match.
- The sample bag is completely closed and each of them contains a **PSC** with 4 grams of desiccant.
- The sample collection date is available, and the sample collection occurred in the past 28 days, and before the expiration date of the **PSC**.
- The **PSC** is not expired.
- The **PSC** dried plasma spot looks homogeneous on the front side and looks completely covered with plasma when observed from the back side (see Figure 8).

Figure 8

PSC Dried Plasma Spots to Be Processed (Left: Front Side; Right: Back Side)




- E2. **PSC** dried plasma spots should be marked and rejected if:
- blood spills are visible (Figure 9a) and/or the membrane is not completely covered with blood (Figure 9b) and therefore the **PSC** spot contains an inhomogeneous front area and/or a back side not completely covered with plasma (visible through the carrier).
 - the membrane is damaged (Figure 9c) and therefore the **PSC** spot contains an inhomogeneous front area and a dark red-brownish back side (visible through the carrier).

Note: The three spots are meant to allow for retesting. A PSC could contain a bad spot, but still provides a good sample for testing. Mark properly the bad spots, in order to be recognizable. Avoid marking them on the spotting layer. Always compare the three spots between each other to evaluate their quality.

Figure 9

Rejection Criteria for PSCs (Left: Front Side; Right: Back Side).

Spots with blood spills (a) or not covered membrane (b), or visibly damaged membranes (c) should not be processed. Spots should be clearly marked when rejected.



- E3. Prepare sample racks as follows: Attach a barcode label clip to each sample rack position where a specimen (S-tube) is to be placed. Attach one of the specific barcode label clips for the controls **[CTM (-) C, HIV-1 L(+)-C, v2.0 and HIV-1 H(+)-C, v2.0]** to each sample rack position where the controls (S-tube) are to be placed. The barcode label clips for controls should have the same control lot number as the lot number on the control vials in the kit. Take care in assigning the right control to the position with the appropriate control barcode clip. Place one Input S-tube into each position containing a barcode label clip.
- E4. Using the AMPLILINK software, select the appropriate test definition file (HI2PSC96 for COBAS® AmpliPrep Instrument plus COBAS® TaqMan® Analyzer, HI2PSC48 for COBAS® AmpliPrep Instrument plus COBAS® TaqMan® 48 Analyzer) and create specimen orders for each specimen and control in the **Orders** window **Sample** folder and complete by saving.
- E5. Assign specimen and control orders to sample rack positions in the **Orders** window **Sample Rack** folder. The sample rack number must be for the rack prepared in Step E1. The patient barcode on the back side of the **PSC** might be scanned through the sample bag (Figure 10). Do not open the sample bag at this point and complete by saving.

Figure 10

Scanning of the Sample Barcode to Order PSC Dried Plasma Spot Samples



- E6. Print the **Sample Rack Order** report to use as a worksheet and verify against the use of the appropriate test file.

- E7. Perform the steps E7 to E11 under a safety hood. For the **PSC** selected, open the sample bag, and remove the spotting layer (Figure 11). Slightly bend the **PSC** and remove one dried plasma spot with sterile forceps or tweezers by pulling it up. Bend the removed dried plasma spot on the **PSC** to facilitate tube insertion (Figure 12).

Figure 11
Removal of the Spotting Layer




Figure 12
Removal of the PSC Dried Plasma Spot and Bending



Note: Use one pair of forceps or tweezers per patient.

- E8. Transfer one pre-bent **PSC** dried plasma spot into the corresponding Input S-tube so that the lowest tip of the **PSC** dried plasma spot reaches the bottom of the tube and is attached to the tube wall to prevent pipetting errors (Figure 13 and Figure 14).

Note: Dried plasma spots may become brittle upon storage. See storage requirements for **PSC**. Handle them carefully while inserting into the tube.

Figure 13
Transfer of the PSC Dried Plasma Spot into the Tube





Figure 14
PSC Dried Plasma Spot Best Positioning in S-Tube



Note: *Ensure that the Input S-tube position number on the Sample Rack Order Report is equal to the SK24 rack position.*

- E9. Place **PSC** with remaining dried plasma spots back to its original Sample Bag containing 4 grams of fresh desiccant for retesting, if required (Figure 15). **PSCs** can be stored for a period of 56 days after transport with and without spotting layer separation (at 18-30°C, or at 2-8°C, or at ≤-10°C).

Figure 15
PSC in Sample Bag for Potential Retesting



- E10. Close the Input S-tube and transfer the tube containing the **PSC** dried plasma spot on the SK24 rack position 4-24 (for further details refer to Table 2). Repeat until all samples are processed.

Note: *Allow SPEX to equilibrate to ambient temperature before use.*

- E11. Open the S-Input tubes individually, save the cap under clean conditions and add 1100 µL of **SPEX** to each S-Input tube (Figure 16). Recap the S-Input tube. Repeat until all samples are processed.

Figure 16
1100 µl SPEX Added to Each PSC Dried Plasma Spot Sample



Note: *Avoid contaminating the upper part of the S-Input tubes with samples and make sure that the tubes are properly capped to prevent evaporation.*

- E12. Transfer the S-Input tubes to the Eppendorf Thermomixer and incubate the prepared S-Input tubes at 56°C and 1000 rpm continuous shaking for 10 minutes to extract the virus from the dried plasma (Figure 17).

Figure 17
Incubation is Performed for 10 Minutes.



Note: *Start the incubation right after the addition of SPEX.*

- E13. During the incubation time prepare the controls as follows.

Note: Do not incubate the controls.

Place one S-Input tube into each position (1 to 3) of the prepared SK24 rack containing a control barcode label clip.

Mix 1 vial of the **CTM (-) C** thoroughly for 20 seconds by vortexing. Transfer 1000 µL of **CTM (-) C** to the S-Input tube 1 of the SK24 rack. Immediately, recap this S-Input tube.

Mix 1 vial of the **HIV-1 L(+)**C**, v2.0** thoroughly for 20 seconds by vortexing. Transfer 1000 µL of **HIV-1 L(+)**C**, v2.0** to the S-Input tube 2 of the SK24 rack. Immediately, recap this S-Input tube.

Mix 1 vial of the **HIV-1 H(+)**C**, v2.0** thoroughly for 20 seconds by vortexing. Transfer 1000 µL of **HIV-1 H(+)**C**, v2.0** to the S-Input tube 3 of the SK24 rack. Immediately, recap this S-Input tube.

Table 2
Ordering and Loading for PSC Dried Plasma Spot Samples

Sample	AMPLILINK Ordering Position	Eppendorf IsoRack Position	SK24 Rack Position
	S-Input tube	S-Input tube	S-Input tube
CTM(-)C	1	Empty	1
HIV-1 L(+)C, v2.0	2	Empty	2
HIV-1 H(+)C, v2.0	3	Empty	3
PSC dried plasma spot + 1100 µL SPEX	4-24	4-24	4-24

E14. After the incubation transfer the samples in the correct order to the SK24 rack containing the controls (Step E13) (for further details refer to Table 2).

Note: Check that the position of the PSC dried plasma spot in the tube is still correct (Figure 14) and adjust with a sterile pipette tip if necessary. In case of bubbles remove them with a sterile pipette tip.

E15. Load the sample rack(s) filled with S-Input tubes onto rack positions **F, G** or **H** of the COBAS® AmpliPrep Instrument.

Part F. Start of COBAS® AmpliPrep Instrument Run

F1. Start the COBAS® AmpliPrep Instrument using the AMPLILINK software.

Part G. End of COBAS® AmpliPrep Instrument Run and Transfer to COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer (for Workflow 2 and 3 in Table 1 Only)

G1. Check for flags or error messages.

G2. Remove processed specimens and controls from the COBAS® AmpliPrep Instrument on either sample racks (for COBAS® TaqMan® Analyzer without Docking Station) or K-carrier racks (for COBAS® TaqMan® 48 Analyzer), depending on the workflow (for further details see Part G).

G3. Remove waste from the COBAS® AmpliPrep Instrument.

Note: All processed specimens and controls should not be exposed to light after completion of specimen and control preparation.

Amplification and Detection

COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer Set-up

The COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer run must be started within 120 minutes following completion of specimen and control preparation.

Note: Do not freeze or store processed specimens and controls at 2-8°C.

Part H. Loading Processed Specimens

- H1. Depending on the workflow, perform the appropriate steps to transfer the K-tubes to the COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer:

Workflow 1: Automated transfer of K-carrier via docking station to COBAS[®] TaqMan[®] Analyzer. Manual intervention is unnecessary.

Workflow 2: Manual transfer of K-tubes in sample rack(s) to COBAS[®] TaqMan[®] Analyzer.

Workflow 3: Manual transfer of K-carrier on K-carrier rack(s) to the COBAS[®] TaqMan[®] 48 Analyzer. Manual transfer of K-carriers into COBAS[®] TaqMan[®] 48 Analyzer using the K-carrier Transporter.

Part I. Start of COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer Run

- I1. Start the COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer by one of the options below depending on the workflow used:

Workflow 1: No intervention necessary.

Workflow 2: Automatic start of the COBAS[®] TaqMan[®] Analyzer after insertion of sample rack(s).

Workflow 3: Fill K-carrier with empty K-tubes if there are fewer than 6 K-tubes on the K-carrier. Filling is guided by the AMPLILINK software. Open thermal cycler cover, load K-carrier into thermal cycler and close lid. Start the COBAS[®] TaqMan[®] 48 Analyzer run.

Part J. End of COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer Run

- J1. At the completion of the COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer run, print Results Report. Check to ensure that the appropriate test definition file was used and look for flags or error messages in the Result report. Specimens with flags and comments are interpreted as described in the Results section. After acceptance, store data in archive.
- J2. Remove used K-tubes from the COBAS[®] TaqMan[®] Analyzer or COBAS[®] TaqMan[®] 48 Analyzer.

RESULTS

The COBAS[®] TaqMan[®] Analyzer or the COBAS[®] TaqMan[®] 48 Analyzer automatically determines the HIV-1 RNA concentration for the specimens and controls. **The HIV-1 RNA concentration is expressed in cp/mL or IU/mL, depending on the used TDF. The conversion factor between HIV-1 RNA cp/mL and HIV-1 IU/mL is 0.6 cp/IU, using WHO 1st International Standard for HIV-1 RNA for Nucleic Acid-Based Techniques (NAT) (NIBSC 97/656)³⁷. This conversion factor was determined using the COBAS[®] AMPLICOR[®] HIV-1 MONITOR Test, v1.5, the COBAS[®] AmpliPrep/COBAS[®] AMPLICOR[®] HIV-1 MONITOR Test, v1.5 and the COBAS[®] TaqMan[®] HIV-1 Test For Use With The High Pure System.**

AMPLILINK Software:

- Determines the Cycle Threshold value (Ct) for the HIV-1 RNA and the HIV-1 QS RNA.
- Determines the HIV-1 RNA concentration based upon the Ct values for the HIV-1 RNA and HIV-1 QS RNA and the lot-specific calibration coefficients provided on the cassette barcodes.
- Determines that the calculated cp/mL for **HIV-1 L(+)**C**, v2.0** and **HIV-1 H(+)**C**, v2.0** fall within the assigned ranges.

Batch Validation – AMPLILINK Version 3.3 and Version 3.4 Series

Check AMPLILINK software results window or printout for flags and comments to ensure that the batch is valid. For control orders, a check is made to determine if the cp/mL or IU/mL value for the control is within its specified range. If the cp/mL or IU/mL value for the control lies outside of its range, a FLAG is generated to show the control has failed.

The batch is valid if no flags appear for any of the controls [**HIV-1 L(+)**C**, v2.0**; **HIV-1 H(+)**C**, v2.0** and **CTM (-) C**].

The batch is not valid if any of the following flags appear for the HIV-1 Controls:

Negative Control

Flag	Result	Interpretation
NC_INVALID	Invalid	An invalid result or a “valid” result that was not negative for HIV-1 target

HIV-1 Low Positive Control, v2.0

Flag	Result	Interpretation
LPCINVALID	Invalid	An invalid result or a control out of range

HIV-1 High Positive Control, v2.0

Flag	Result	Interpretation
HPCINVALID	Invalid	An invalid result or a control out of range

If the batch is invalid, repeat the entire batch including specimen and control preparation, amplification and detection.

Interpretation of Results

For a valid batch, check each individual specimen for flags or comments on the result printout. Interpret the results as follows:

- A valid batch may include both valid and invalid specimen results depending on whether flags and/or comments are obtained for the individual specimens.

Specimen results are interpreted as follows:

	Titer Result	Interpretation
	Target Not Detected	Ct value for HIV-1 above the limit for the assay or no Ct value for HIV-1 obtained. Report results as "HIV-1 RNA not detected".
Copies/mL	< 2.00E+01 cp/mL (Plasma) < 738 cp/ml (PSC)	Calculated cp/mL are below the Limit of Detection of the assay. Report results as "HIV-1 RNA detected, less than 20 HIV-1 RNA cp/mL " for Plasma and "HIV-1 RNA detected, less than 738 HIV-1 RNA cp/mL " for PSC .
	≥ 2.00E+01 cp/mL and ≤ 1.00E+07 cp/mL (Plasma)	Calculated results greater than or equal to 20 cp/mL and less than or equal to 1.00E+07 cp/mL are within the Linear Range of the assay used with Plasma.
	≥ 738 cp/mL and ≤ 1.00E+07 cp/mL (PSC)	Calculated results greater than or equal to 738 cp/mL and less than or equal to 1.00E+07 cp/mL are within the Linear Range of the assay used with the PSC .
	> 1.00E+07 cp/mL (Plasma and PSC)	Calculated cp/mL are above the range of the assay. Report results as "greater than 1.00E+07 HIV-1 RNA cp/mL". If quantitative results are desired, the original specimen should be diluted 1:100 with HIV-1-negative human EDTA-plasma and the test repeated. Multiply the reported result by the dilution factor.
International Units/mL	< 3.34E+01 IU/mL (Plasma)	Calculated IU/mL are below the Limit of Detection of the assay. Report results as "HIV-1 RNA detected, less than 33.4 HIV-1 RNA IU/mL " for Plasma.
	≥ 3.34E+01 IU/mL and ≤ 1.67E+07 IU/mL (Plasma)	Calculated results greater than or equal to 33.4 IU/mL and less than or equal to 1.67E+07 IU/mL are within the Linear Range of the assay used with Plasma.
	> 1.67E+07 IU/mL (Plasma)	Calculated IU/mL are above the range of the assay. Report results as "greater than 1.67E+07 HIV-1 RNA IU/mL". If quantitative results are desired, the original specimen should be diluted 1:100 with HIV-1-negative human EDTA-plasma and the test repeated. Multiply the reported result by the dilution factor.

Note: Specimens above the range of the assay that produce an invalid result with a flag "QS_INVALID" should not be reported as > 1.00E +07 cp/mL or 1.67E +07 IU/mL. The original specimen should be diluted 1:100 with HIV-1-negative human EDTA-plasma and the test repeated. Multiply the reported result by the dilution factor.

Note: Titer Result "Failed". Interpretation: Specimen is not correctly processed during specimen preparation on the COBAS® AmpliPrep Instrument.

Note: Titer Result "Invalid". Interpretation: An Invalid Result.

QUALITY CONTROL

One **CTM (-) C**, one **HIV-1 L(+)**C**, v2.0** and one **HIV-1 H(+)**C**, v2.0** must be included in each test batch. The batch is valid if no flags appear for any of the controls [**HIV-1 L(+)**C**, v2.0**, **HIV-1 H(+)**C**, v2.0** and **CTM (-) C**].

Check the batch printout for flags and comments to ensure that the batch is valid.

Negative Control

The **CTM (-) C** must yield a "Target Not Detected" result. If the **CTM (-) C** is flagged as invalid, then the entire batch is invalid. Repeat the entire process (specimen and control preparation, amplification and detection). If **CTM (-) C** is consistently invalid in multiple batches, contact your local Roche office for technical assistance.

Positive Controls

The assigned titer range for **HIV-1 L(+)**C**, v2.0** and **HIV-1 H(+)**C**, v2.0** is specific for each lot of reagents, and is provided on the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 reagent cassette barcodes.

The HIV-1 RNA cp/mL for **HIV-1 L(+)**C**, v2.0** and **HIV-1 H(+)**C**, v2.0** should fall within their assigned titer ranges. If one or both of the positive controls are flagged as invalid, then the entire batch is invalid. Repeat the entire process (specimen and control preparation, amplification and detection). If the HIV-1 RNA titer of one or both of the positive controls is consistently outside the ranges in multiple batches, contact your local Roche office for technical assistance.

PROCEDURAL PRECAUTIONS

As with any test procedure, good laboratory technique is essential to the proper performance of this assay.

PROCEDURAL LIMITATIONS

1. This test has been validated for use with only human plasma collected in EDTA anticoagulant. Testing of other specimen types may result in inaccurate results.
2. The performance of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 has neither been evaluated with specimens containing HIV-1 group N, nor with specimens containing HIV-2.
3. Reliable results are dependent on adequate specimen collection, transport, storage and processing procedures.
4. The presence of AmpErase enzyme in the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 Master Mix reduces the risk of amplicon contamination. However, contamination from HIV-1 positive controls and clinical specimens can be avoided only by good laboratory practices and careful adherence to the procedures specified in this Package Insert.
5. Use of this product should be limited to personnel trained in the techniques of PCR.
6. This product can only be used with the COBAS® AmpliPrep Instrument and the COBAS® TaqMan® Analyzer or COBAS® TaqMan® 48 Analyzer.
7. Though rare, mutations within the highly conserved regions of the viral genome covered by the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 primers and/or probes may result in the under-quantitation of or failure to detect the virus.
8. Detection of HIV-1 RNA is dependent on the number of virus particles present in the specimen and may be affected by specimen collection methods and patient factors, (i.e., age, presence of symptoms, and/or stage of the infection). While the clinical specificity of the test is 99.3% (95% CI = 98.2% to 99.8%), some low level false positive results in HIV-negative individuals have been noted.
9. Due to inherent differences between technologies, it is recommended that, prior to switching from one technology to the next, users perform method correlation studies in their laboratory to quantify technology differences.

INTERFERING SUBSTANCES

Elevated levels of triglycerides (up to 3500 mg/dL), bilirubin (up to 28 mg/dL), albumin (up to 8900 mg/dL), hemoglobin (up to 900 mg/dL) and human DNA (up to 0.4 mg/dL) in specimens as well as the presence of autoimmune diseases or respective markers such as Systemic Lupus Erythematosus (SLE), Rheumatoid Arthritis (RA) and Antinuclear Antibody (ANA) were shown not to interfere with the quantitation of HIV-1 RNA or impact the specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0. The evaluation was performed according to CLSI Guideline EP7-A2 using one lot of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 reagents.

The following drug compounds tested at 3 times the Peak Plasma Level (Cmax) have been shown not to interfere with the quantitation of HIV-1 RNA or impact the specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0:

HIV drugs:	
Protease Inhibitors Atazanavir Darunavir Fosamprenavir Lopinavir/Ritonavir Nelfinavir mesylate Ritonavir Saquinavir Tipranavir	Nucleoside Analogue, Inhibitor of Reverse Transcriptase Abacavir sulfate Didanosine, ddl Emtricitabine Lamivudine, 3TC Stavudine, 4dT Tenofovir DF Zidovudine
Integrase Inhibitor Raltegravir	Non-nucleoside, Inhibitor of Reverse Transcriptase Efavirenz Nevirapine
Entry Inhibitor Maraviroc	Fusion Inhibitors Enfuvirtide
HBV and / or HCV Drugs:	
Nucleotide analogue Adefovir dipivoxil	Nucleoside analogue Entecavir Telbivudine
Immune Modulator Peginterferon α-2a Peginterferon α-2b Ribavirin	
Compounds for Treatment of Herpes Viruses:	
Nucleotide analogue Acyclovir	Nucleotide derivative Ganciclovir Valganciclovir HCl

NON-CLINICAL PERFORMANCE EVALUATION

Key Performance Characteristics for EDTA Plasma Samples

A. Limit of Detection

The limit of detection of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was determined by testing the 2nd International HIV-1 RNA WHO Standard, NIBSC Code 97/650⁴², HIV-1 subtype B, diluted in HIV-1-negative human EDTA plasma. The limit of detection was determined for three reagent lots. Three dilution series were analyzed for each reagent lot. A total of approximately 126 replicates per concentration level were tested. The evaluation was performed according to CLSI Guideline EP17-A.

The concentration of HIV-1 RNA that can be detected with a positivity rate of greater than 95% as determined by PROBIT Analysis, is 20 cp/mL or 33 IU/mL. The results for the individual lots were 17.7 cp/mL (95% confidence interval: 13.7 – 26.9 cp/mL) for lot 1, 17.0 cp/mL (95% confidence interval: 14.0 – 22.6 cp/mL) for lot 2 and 14.2 cp/mL (95% confidence interval: 11.2 – 22.1 cp/mL) for lot 3. The combined results for all three reagent lots are shown in Table 3. **The conversion factor between IU/mL and cp/mL was determined using the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5, the COBAS® AmpliPrep/COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5 and the COBAS® TaqMan® HIV-1 Test For Use With The High Pure System.**

Table 3
Limit of Detection of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

Nominal Input (HIV-1 RNA IU/mL)	Nominal Input (HIV-1 RNA cp/mL)	No. Replicates	No. Positives	Positivity Rate
100	60	126	126	100%
67	40	186	185	99%
50	30	126	125	99%
33	20	126	124	98%
25	15	59	53	90%
17	10	126	108	86%
8	5	125	66	53%
0	0	126	0	0%
PROBIT 95% Hit Rate		27.5 IU/mL (95% confidence interval: 23.8 – 33.0 IU/mL) 16.5 cp/mL (95% confidence interval: 14.3 – 19.8 cp/mL)		

In addition, dilutions of cell culture supernatants representing HIV-1 group M subtypes A-H in HIV-1-negative human EDTA plasma were analyzed with two reagent lots. For each HIV-1 subtype isolate several concentration levels (nominal titers between 10 and 75 cp/mL) were tested in 24 replicates per reagent lot. The assignment of nominal concentrations to the cell culture stock materials was performed by averaging the titers of the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5, the VERSANT® HIV-1 RNA 3.0 Assay (bDNA) titer and the Abbott RealTime HIV-1 assay. Hit rate analysis shows a positivity rate of greater than 95% for all subtypes at 20 cp/mL or lower. The combined results for the two reagent lots are shown in Table 4.

Table 4
Limit of Detection Verification for the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

Subtype	Isolate Designation	Lowest Concentration Level ≥ 95% Hit Rate (cp/mL)
A	92UG029	10
A	4237A/98	20
B	92TH026	20
B	8E5/LAV	20
C	92BR025	20
C	3777A/97	11
D	92UG021	20
D	92UG035	11
CRF01_AE	92TH022	12
CRF01_AE	92TH009	14
F	93BR020	20
G	ARP173/RU570	13
H	HIV V1557	16

B. Precision

The Precision of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was determined by analysis of serial dilutions of a HIV-1 cell culture supernatant specimen (HIV-1 subtype B) in HIV-1-negative human EDTA plasma. The titer assignment of the cell culture supernatant (stock concentration) was performed by a method that ensures traceability to the 1st International HIV-1 RNA WHO Standard, NIBSC Code 97/656³⁶. Three reagent lots were analyzed and 15 runs per reagent lot were performed, each consisting of 6 dilution levels and 3 replicates at each level. Each specimen was taken through the entire COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 procedure, including specimen preparation, amplification and detection. Therefore, the precision reported here represents all aspects of the test procedure. The results for each reagent lot and for the three reagent lots combined are shown in Table 5.


Table 5
Precision of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

Titer (cp/mL)	Lot 1	Lot 2	Lot 3	All three lots combined	
	Total SD in log	Total SD in log	Total SD in log	Total SD in log	Total Lognormal CV (%)
1.0E+02	0.19	0.16	0.17	0.17	41
1.0E+03	0.07	0.09	0.07	0.08	20
1.0E+04	0.07	0.07	0.06	0.07	16
1.0E+05	0.04	0.05	0.07	0.06	15
1.0E+06	0.10	0.09	0.10	0.10	25
1.0E+07	0.11	0.12	0.14	0.13	33

C. Linear Range

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was found to give a linear response from 20 ($\log_{10} = 1.30$) HIV-1 RNA cp/mL to $1.0E+07$ ($\log_{10} = 7.00$) HIV-1 RNA cp/mL. The evaluation was performed according to CLSI Guideline EP6-A using two lots of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 reagents and serial dilutions of a high titer HIV-1 RNA (+) cell culture supernatant specimen. Two reagent lots were analyzed and 15 runs per reagent lot were performed, each consisting of 12 dilution levels and 3 replicates at each level. The results for one reagent lot are shown in Figure 18.

Figure 18
Linearity for the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0



D. Inclusivity of HIV-1 Group M


Eight subtype categories have been proposed for HIV-1 group M based on nucleotide divergence. These subtypes are designated with capital alphabetical letters from A through H⁴¹.

The performance of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 on HIV-1 subtypes was evaluated by analysis of cell culture stock material of representatives for each HIV-1 group M subtype A through H. The assignment of nominal concentrations to the cell culture stock materials was performed by averaging the titers of the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5, the VERSANT® HIV-1 RNA 3.0 Assay (bDNA) titer and the Abbott RealTime HIV-1 assay. Each cell culture stock material was diluted to nominal concentrations of approximately $2.00E+02$, $2.00E+04$ and $2.00E+06$ cp/mL in EDTA plasma. The concentrations were then tested in 10 replicates by the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 using one reagent lot. The mean \log_{10} titers of all concentrations and subtypes were compared to the respective \log_{10} nominal titers.

The evaluation of the 8 HIV-1 subtype isolates by the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 demonstrates equivalent results for all tested representatives of the HIV-1 group M subtypes (see Figure 19). Mean \log_{10} concentration results for all subtypes were within $\pm 0.3 \log_{10}$ of the assigned input concentration.

Figure 19
COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

□ nominal concentration ■ COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0




E. HIV-1 Group O Detection

Dilutions of a HIV-1 group O cell culture supernatant (isolate MVP5180) in human EDTA plasma were analyzed with two reagent lots. Several concentration levels (nominal titers between 10 and 75 cp/mL) were tested in 24 replicates per reagent lot. Assignment of the nominal concentration to the cell culture stock material was performed by the Abbott RealTime HIV-1 assay. Hit rate analysis shows a positivity rate of greater than 95% at 20 cp/mL.

The HIV-1 group O cell culture stock material was diluted to nominal concentrations of approximately 2.00E+02, 2.00E+03 and 2.00E+05 cp/mL in EDTA plasma. The concentrations were then tested in 10 replicates by the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 using one reagent lot. The mean log₁₀ titers of all concentrations were linear and within ± 0.3 log₁₀ of the respective log₁₀ nominal titer (see Figure 20).

Figure 20
COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

Linearity Subtype O (MVP5180)



In addition, 10 cell culture materials and one diluted patient specimen (11613) representing HIV-1 group O were tested in parallel in the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and in the Abbott RealTime HIV-1 assay. All 11 specimens were found positive with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 (see Table 6). Both tests returned a mean log₁₀ titer for the 11 specimens within 0.1 log₁₀.

Table 6
**Recognition of HIV-1 Group O Isolates by the COBAS® AmpliPrep/
 COBAS® TaqMan® HIV-1 Test, v2.0**

Isolate Designation	Log ₁₀ titer COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0	Log ₁₀ titer Abbott RealTime HIV-1 assay
BBI PRD 301, BV-5050	3.09	2.42
BBI PRD 301, BV-5051	2.86	3.35
BBI PRD 301, BV-5003	3.00	2.71
BBI PRD 301, BV-5024	2.87	2.69
MVP5180	2.78	3.25
HIV-1 CA-9	3.31	3.08
BCF01	5.71	5.61
BCF02	5.16	5.39
BCF07	4.27	4.81
BCF011	5.57	5.26
11613	2.97	2.05
Mean log₁₀ titer	3.78	3.69

F. Specificity

The specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was determined with two reagent lots by analysis of HIV-1-negative EDTA plasma specimens from blood donors. A total of 660 individual EDTA plasma specimens showed valid results and all were negative for HIV-1 RNA in the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0. Based on these results, the specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 is 100% (one-sided lower 95% confidence limit: $\geq 99.6\%$).

G. Analytical Specificity

The analytical specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was evaluated by adding cultured organisms (viruses, bacteria, yeast) or DNA (HTLV-2) at $5E+04$ particles/mL input concentration into HIV-1-negative human EDTA plasma and into HIV-1-positive EDTA plasma at $1.5E+02$ cp/mL HIV-1 (see Table 7).

None of the organisms tested showed cross reaction with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0. HIV-1-positive specimens returned titer results that were within $\pm 0.5 \log_{10}$ from a HIV-1-positive control.

Table 7
Analytical Specificity Specimens

Virus	Bacteria
<i>Adenovirus type 5</i>	<i>Staphylococcus aureus</i>
<i>Cytomegalovirus</i>	<i>Propionibacterium</i>
<i>Epstein-Barr virus</i>	<i>acnes</i>
<i>Human Herpes Virus type 6</i>	
<i>Herpes simplex virus type 1</i>	
<i>Herpes simplex virus type 2</i>	
<i>Human T-Cell Lymphotropic virus type 1</i>	Yeast
<i>Human T-Cell Lymphotropic virus type 2</i>	<i>Candida albicans</i>
<i>Influenza A</i>	
<i>Hepatitis A virus</i>	
<i>Hepatitis B virus</i>	
<i>Hepatitis C virus</i>	

H. Method Correlation

The performance of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was compared to the COBAS® AmpliPrep/COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5, to the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test and to the Abbott RealTime HIV-1 assay by analysis of 92 prospectively collected, undiluted HIV-1 positive clinical specimens and by analysis of 34 diluted cell culture supernatants. The specimens comprised HIV-1 group M subtypes A to H as well as circulating recombinant forms of the virus and were analyzed at two external sites. A total of 126 samples spread over the dynamic range of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 were tested with the four tests. Only valid titer pairs within the linear ranges of both assays compared were considered for Deming regression analysis (see Figure 21 to Figure 23).

Figure 21
Correlation of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0




Figure 22
Correlation of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0





Figure 23
Correlation of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0



I. Whole System Failure

The whole system failure rate for two kit lots of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2 was determined by testing 113 replicates for lot 1 and 114 replicates for lot 2 of EDTA plasma spiked with HIV-1 group M subtype B. These samples were tested at a target concentration of approximately 3 x LoD. The combined results (lot 1 and 2) of this study determined that one replicate returned a negative result for the HIV-1 target, resulting in a whole system failure rate of 0.4%.

Key Performance Characteristics for PSC Dried Plasma Spot Samples

A. Plasma Limit of Detection Using the Plasma Separation Card

The plasma limit of detection of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** was determined by analysis of plasma titers assigned to serial dilutions of HIV-1 group M cell culture supernatant, in HIV-negative human whole blood. Panels of five concentration levels plus a negative were tested over three lots of **PSCs** and three lots of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test reagents, multiple runs, days, operators, and instruments. A part of the highest panel member was centrifuged and the plasma was titer assigned by Calibrator Bracketing Method (CBM) using **cobas**® HIV-1 (on **cobas**® 6800/8800 Systems) with the 3rd HIV-1 WHO International Standard, HIV-1 group M, subtype B for preparation of the high and low calibrator.

The results for **PSC** are shown in Table 8. The study demonstrates that the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** detected HIV-1 RNA at a concentration of 737.9 cp/mL as determined by Probit with a hit rate of 95%.

Table 8
Limit of Detection of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

HIV-1 group M assigned plasma concentration (cp/mL)	Number of Reactives	Number of Valid Replicates	% Reactive
1691.1	61	61	100%
1098.1	51	51	100%
2276.4	61	62	98.39%
845.6	59	60	98.33%
549.1	46	49	93.88%
1138.2	59	60	98.33%
563.7	54	60	90.00%
366	49	53	92.45%
758.8	56	61	91.80%
281.9	46	62	74.19%
183	36	49	73.47%
379.4	50	62	80.65%
140.9	38	63	60.32%
91.5	22	54	40.74%
189.7	37	63	58.73%
0	0	63	0.00%
0	0	54	0.00%
0	0	59	0.00%
PROBIT 95% Hit Rate	737.9 cp/mL (95% confidence interval: 614.3 - 938.5 cp/mL) 1230 IU/mL (1023.8 - 1564.2 IU/mL)		

Estimation of LOD in Whole Blood:

The corresponding titers in whole blood of the same sample cannot be exactly determined, since whole blood is not a sample type for viral load testing. Whole blood titers based on the amount of HIV-1 RNA spiked into the whole blood samples were not used for LoD estimation because the amount of spiked RNA does not necessarily correspond to the amount of RNA in plasma. Even after centrifugation, RNA can remain in the buffy coat or associated with the cellular fraction of whole blood. However, since other technologies like dried blood spots have used spiked whole blood titers to estimate their limit of detection in whole blood, an estimate for the **PSC** whole blood LoD can be provided based on an empirical factor which is assumed to be related to the average hematocrit content (45%) of the samples (Table 9).

Whole blood LOD values are estimates based on the relation:

Whole blood LOD estimate = **PSC** plasma LOD / 1.8

Table 9
Whole Blood LoD Estimate

LoD by PROBIT analysis (95% Hit Rate)	439.0 cp/mL
95% confidence interval	366.1 – 557.6 cp/mL

B. Precision Using the Plasma Separation Card

Precision of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** was determined by analysis of serial dilutions of an HIV-1 high positive sample (high titer HIV-1 RNA positive cell culture supernatant specimen) in HIV negative EDTA whole blood. Five dilution levels were tested in 48 replicates for each level and process volume across two lots of **PSC** and two reagent lots of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 using two CAP/CTM Systems and two operators over 12 days. Each sample was carried through the entire **PSC** workflow and COBAS® AmpliPrep/COBAS®

TaqMan® HIV-1 Test, v2.0 procedure. The precision results reported here represent all aspects of the test procedure. The results are shown in Table 10.

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** showed high precision for two lots of **PSC** and reagents tested across a concentration range of 7.38E+02 cp/mL to 1.00E+07 cp/mL.

Table 10
Within Laboratory Precision of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0
in Combination with the PSC

Measured concentration (cp/mL)	Source material	Pooled SD
6.31E+06	Cell Culture	0.08
1.05E+06	Cell Culture	0.09
1.02E+05	Cell Culture	0.12
1.05E+04	Cell Culture	0.19
2.29E+03	Cell Culture	0.25


C. Linear Range Using the Plasma Separation Card

The linearity study of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** was performed with a dilution series consisting of 9 panel members spanning the linear range for the predominant HIV-1 group M subtype B. Panel members were prepared from a high titer HIV-1 RNA positive cell culture supernatant specimen. The evaluation was performed according to CLSI Guideline EP06-A²⁰. Two **PSC** and two reagent lots were analyzed on two CAP/CTM Systems, three operators and in total 20 replicates per concentration level.

A part of one panel member was centrifuged and the plasma was titer assigned by Calibrator Bracketing Method (CBM) using **cobas**® HIV-1 (on **cobas**® 6800/8800 Systems) with the 3rd HIV-1 WHO International Standard, HIV-1 group M, subtype B for preparation of the high and low calibrator.

In combination with the **PSC**, COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 is linear from 7.38E+02 cp/mL to 1.00E+07 cp/mL and shows an absolute deviation from the better fitting non-linear regression of less than $\pm 0.16 \log_{10}$ with the **PSC** (see Figure 24). Across the linear range, the accuracy of the test was within $\pm 0.3 \log_{10}$.

Figure 24
Linearity for the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0
in Combination with the Plasma Separation Card



D. Subtype Verification Using the Plasma Separation Card

Although HIV subtype should not affect **PSC** performance, cultured HIV-1 samples for common HIV-1M subtypes (A, C and D) were diluted to one concentration level in whole blood. The precision and accuracy determination was performed with 12 replicates for each sample. Testing was conducted with 1 lot of **PSC** and 1 lot of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 reagents.

Each whole blood sample was centrifuged and the plasma was titer assigned by Calibrator Bracketing Method (CBM) using **cobas**® HIV-1 (on **cobas**® 6800/8800 Systems) with the 3rd HIV-1 WHO International Standard, HIV-1 group M, subtype B for preparation of the high and low calibrator.

The results are shown in Table 11. These results verify that COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** detected HIV for HIV-1M (A, C and D).

Table 11
Verification of HIV-1 Group M Subtypes A, C and D

HIV-1 M Subtype	Number of valid replicates	Accuracy	Precision	Plasma vs. PSC-plasma Equivalency
Subtype A	12	-0.02	0.09	0.17
Subtype C	12	0.13	0.16	0.06
Subtype D	12	0.07	0.15	0.05

E. Specificity Using the Plasma Separation Card

The specificity of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** was determined by analyzing HIV negative EDTA whole blood samples from individual donors. 160 individual EDTA whole blood samples were tested with two lots of **PSC** and two lots of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 reagents. One sample tested positive and 159 negative for HIV-1 RNA. In the test panel the specificity of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** was 99.4% (95% confidence limit: $\geq 97.07\%$).

F. Whole System Failure Using the Plasma Separation Card

The whole system failure rate of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in combination with the **PSC** was determined by testing 100 replicates of EDTA whole blood spiked with HIV-1 group M subtype B. These samples were tested at a target concentration of approximately 3 x LoD. The results of this study determined that no replicate returned a negative result for the HIV-1 target, resulting in a whole system failure rate of 0%.

CLINICAL PERFORMANCE EVALUATION

Key Performance Characteristics for EDTA Plasma Samples

Reproducibility

Reproducibility of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 Test was evaluated in EDTA plasma using 2 different workflows (COBAS® AmpliPrep/COBAS® TaqMan® Analyzer System and COBAS® AmpliPrep/COBAS® TaqMan® 48 Analyzer System). The study was performed using panels constructed from well-characterized HIV-1 group M, subtype B cultured virus stock and from EDTA plasma that was negative for HIV-1 RNA and HIV-1/2 antibodies. The panel covered the dynamic range of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 as well as the key medical decision points for the intended use and supported by the 2008 Department of Health and Human Services Guidelines for the Use of Antiretroviral Agents in HIV-1-Infected Adults and Adolescents¹³. The study was designed to evaluate key variables contributing to total precision variance, including lot, site/instrument, operator, day/run, and within-run. Additional analysis were conducted to compare the performance characteristics and comparative precision variability between the two workflows. Two operators at each of 3 sites performed 5 days of testing with each of 3 reagent kit lots using each workflow. Each run consisted of one set of controls (1 high positive, 1 low positive, and 1 negative) and a 7-member panel tested in triplicate (21 sample) on the COBAS® AmpliPrep Instrument. The prepared samples and controls were amplified and detected on the COBAS® TaqMan® Analyzer or on COBAS® TaqMan® 48 Analyzer.

Reproducibility was evaluated by using a random effects model with terms for (a) lot, (b) site/instrument, (c) operator nested within site/instrument, (d) day/run nested within lot, site/instrument, and operator, and (e) aliquots within-run components by using PROC MIXED and log₁₀ transformed results. The percentage of variability due to each component and coefficient of variation of the log₁₀ transformed HIV-1 RNA concentration were calculated. Only the Within Assay Range (2.00E+01 to 1.00E+07 cp/mL) data were investigated.

Table 12 shows the total precision variance and total precision standard deviation obtained from the COBAS® AmpliPrep/COBAS® TaqMan® Analyzer System as determined by analysis of variance. In general, the within-run component contributed more variability than other components.

Table 12
Attributable Percentage of Total Variance, Total Precision Standard Deviation, and Lognormal CV of HIV-1 RNA Concentration (log₁₀ cp/mL) from Tests Within Assay Range

HIV-1 RNA Concentration (log ₁₀ cp/mL)			Contribution to Total Variance (%)					Total Precision
Expected	Observed (Average)	No. of Valid Tests ¹	Lot	Site/Instrument	Operator	Day/Run	Within-Run	Standard Deviation (Lognormal %CV)
1.699	1.832	270	5%	2%	0%	8%	85%	0.20 (48%)
2.602	2.676	275	6%	1%	0%	17%	77%	0.11 (25%)
3.000	3.067	274	16%	0%	4%	12%	69%	0.10 (24%)
3.699	3.822	273	20%	6%	0%	17%	57%	0.10 (23%)
4.699	4.746	273	27%	0%	0%	14%	59%	0.07 (17%)
5.699	5.644	274	33%	10%	0%	19%	38%	0.10 (23%)
6.699	6.751	259	27%	14%	0%	20%	39%	0.12 (27%)

Note: Within assay range results are from 20 cp/mL to 1.00E+07 cp/mL (1.30 log₁₀ cp/mL to 7.00 log₁₀ cp/mL), inclusive.

¹ Number of tests within assay range.

Results obtained from the COBAS® AmpliPrep/COBAS® TaqMan® 48 System Workflow are summarized in Table 13. In general, the within-run component contributed more variability than other components with the exception of the highest titer panel member.

Table 13

Attributable Percentage of Total Variance, Total Precision Standard Deviation, and Lognormal CV of HIV-1 RNA Concentration (\log_{10} cp/mL) from Tests within Assay Range


HIV-1 RNA Concentration (\log_{10} cp/mL)			Contribution to Total Variance (%)					Total Precision
Expected	Observed (Average)	No. of Valid Tests ¹	Lot	Site/Instrument	Operator	Day/Run	Within-Run	Standard Deviation (Lognormal %CV)
1.699	1.804	266	7%	2%	0%	2%	89%	0.21 (52%)
2.602	2.672	273	26%	0%	2%	5%	68%	0.10 (24%)
3.000	3.048	272	17%	0%	0%	6%	77%	0.09 (21%)
3.699	3.814	271	39%	0%	2%	13%	46%	0.08 (19%)
4.699	4.756	272	30%	0%	0%	10%	61%	0.07 (16%)
5.699	5.647	272	35%	0%	6%	16%	43%	0.11 (25%)
6.699	6.727	269	45%	0%	4%	13%	38%	0.11 (26%)

Note: Within assay range results are from 20 cp/mL to 1.00E+07 cp/mL (1.30 \log_{10} cp/mL to 7.00 \log_{10} cp/mL), inclusive.

¹ Number of tests within assay range.

The results shown in Figure 25 display the plot of the total precision standard deviation with the corresponding approximate 95% Confidence Intervals against the mean \log_{10} HIV-1 RNA concentrations. These results indicate a comparable precision performance between the COBAS® AmpliPrep/COBAS® TaqMan® (CAP/CTM) System and the COBAS® AmpliPrep/COBAS® TaqMan® 48 (CAP/CTM48) System configurations.

Figure 25
Total Precision Standard Deviation (approximate 95% CI)



Note: The approximate 95% CI for the total precision standard deviation was calculated by taking the square root of the 95% CI bounds of the total precision variance.

Clinical Sensitivity, Specificity and Method Comparison

Methodology

The primary objective of this study was to evaluate the clinical specificity and sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 in specimens from HIV-negative and HIV-1-positive subjects. Both fresh (never frozen) and frozen EDTA plasma samples were tested in each of the evaluations. The secondary objectives were to compare results and evaluate the positive percent agreement and negative percent agreement of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results to those obtained with the FDA-approved tests, COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test HIV-1 Test and the COBAS® AMPLICOR HIV-1 MONITOR Test, v1.5.

Clinical specificity was evaluated with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 by testing 148 fresh (never frozen) samples and 418 frozen samples collected from blood donors who were negative for HIV-1/2 antibodies. Clinical sensitivity of the test was evaluated with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 by testing 117 fresh samples and 301 frozen samples in EDTA plasma collected from HIV-1-infected subjects (frozen samples were randomly distributed across test sites by CD4 cell count category). Test results from the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 were compared to those obtained with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test and COBAS® AMPLICOR HIV-1 MONITOR Test, v1.5. Testing was conducted at 3 test sites, with 1 COBAS® AmpliPrep/COBAS® TaqMan® System per site. Three COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 reagent lots were used.

Statistical Methods

Fresh and frozen samples from HIV-negative and HIV-1-positive subjects were tested with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0, the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, and the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5. HIV-negative subjects were evaluable for statistical analyses of the specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 if they generated valid COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results. HIV-1-positive subjects were evaluable for statistical analyses of the sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 if they generated valid COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results and had valid COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test results within the linear range of the assay.

The clinical specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was calculated as the percentage of evaluable HIV-negative subjects who had Target Not Detected COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results. The associated 95% exact confidence interval (CI) was also provided. The clinical sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was calculated as the percentage of evaluable HIV-1-positive subjects who had detectable HIV-1 viral load on the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0. The associated 95% exact confidence interval (CI) was also provided. The method comparison evaluated COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results separately with both comparative platforms (COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test and the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5). Positive and negative percent agreements were calculated between the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and each comparative platform. Paired samples from HIV-1-positive subjects contributing within linear range results for both the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test were compared using scatter plots and analyzed using the Deming regression.

Results

A total of 566 evaluable HIV-negative and 418 HIV-1-positive patient specimens were included in clinical specificity and sensitivity analyses. Approximately 75% of the patient specimens were frozen and 25% were fresh. The specific distribution of each platform is summarized in Table 14.

Table 14
Evaluable HIV-1 Negative and Positive Subjects by Sample Type

Sample Type	HIV-Negative Specimens	HIV-1-Positive Specimens
Fresh	148 (26.1%)	117 (28.0%)
Frozen	418 (73.9%)	301 (72.0%)
Total	566	418

The demographic characteristics of the 418 evaluable HIV-1-positive specimens are summarized in Table 15. The CD4 cell counts of the subjects distributed approximately evenly across CD4 cell count categories (<200, 200-500, >500 cells/ μ L). Most of the subjects were male (74.2%) and between 30 to 49 years of age (72.5%). The ethnic distribution is comparable to that observed in the HIV-1 population of the United States³².

Table 15
Demographic Characteristics of Evaluable HIV-1-Positive Subjects

Demographic Characteristic	Category	HIV-1-Positive Subjects
Overall	Total	418
CD4 Cell Count (cells/μL)	< 200	130 (31.1%)
	200 - 500	152 (36.4%)
	> 500	136 (32.5%)
Sample Type	Fresh	117 (28.0%)
	Frozen	301 (72.0%)
Sex	Male	310 (74.2%)
	Female	108 (25.8%)
Age (Years)	18-29	23 (5.5%)
	30-39	100 (23.9%)
	40-49	203 (48.6%)
	50-59	74 (17.7%)
	\geq 60	18 (4.3%)
Ethnicity	Caucasian	129 (30.9%)
	Hispanic	46 (11.0%)
	Black	223 (53.3%)
	Asian / Pacific Islander	3 (0.7%)
	Other	17 (4.1%)
On Antiretroviral Medication	Yes	240 (57.4%)
	No	178 (42.6%)

The clinical specificity of the COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 (Table 16) was 99.3% (562/566; 95% CI = 98.2% to 99.8%), with 4 specimens classified as false positives. Three of these specimens were reported at < 20 cp/mL, below the LLoQ of the assay. The remaining single specimen out of the 566 tested was within the linear range but at a very low titer (28.8 cp/mL). The clinical specificity of

the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was similar for both fresh specimens (99.3% [147/148; 95% CI = 96.3% to 100%]) and frozen specimens (99.3% [415/418; 95% CI = 97.9% to 99.9%]).

Table 16
Clinical Specificity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

Subject Group	CAP/CTM HIV-1 Test, v2.0		Total N	Clinical Specificity (95% exact CI)
	Positive	Negative		
HIV-Negative	4 (0.7%)	562 (99.3%)	566	99.3% (98.2%, 99.8%)

The clinical sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was defined as the percentage of evaluable HIV-1-positive subjects who had a positive COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 result and is summarized in Table 17. The clinical sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 was 100% (418/418; 95% CI = 99.1% to 100%). There were no subjects that had false negative COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results. The clinical sensitivity was tested in an HIV-patient population reflective of that in the United States with regards to gender, age, ethnicity and exposure to antiretroviral therapy²³. The test demonstrated 100% clinical sensitivity independent of the above listed demographics, CD4 cell count, or sample type (fresh versus frozen).

Table 17
Clinical Sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0

Subject Group	CAP/CTM HIV-1 Test, v2.0		Total N	Clinical Sensitivity (95% exact CI)
	Positive	Negative		
HIV-1-Positive	418 (100.0%)	0 (0.0%)	418	100.0% (99.1%, 100.0%)

Clinical Method Comparison

COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 versus the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test

The comparison of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test results for the 950 subjects eligible for the analysis is summarized in Table 18. The positive percent agreement of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 with respect to the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test was 99.5% (427/429; 95% CI = 98.3% to 99.9%). The negative percent agreement of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 with respect to the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test was 98.1% (511/521; 95% CI = 96.5% to 99.1%). There were 10 samples with positive COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 results and negative COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test results. Three samples were at titers below the LLoQ of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, most likely a reflection of the increased sensitivity of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0. Three samples were false positive COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test v2.0 results from HIV-negative subjects identified in the clinical specificity analysis that again were below the LLoQ. Four samples had titers ranging from 24.9 cp/mL to 158 cp/mL and are likely reflective of the known variability associated with low titer quantitation.

Table 18
Comparison of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 versus the
COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test

CAP/CTM HIV-1 Test, v2.0	CAP/CTM HIV-1 Test		Total
	Positive	Negative	
Positive	427	10	437
Negative	2	511	513
Total	429	521	950
Positive Percent Agreement (95% exact CI)	99.5% (98.3%, 99.9%)		
Negative Percent Agreement (95% exact CI)		98.1% (96.5%, 99.1%)	

CI=confidence interval; CAP/CTM HIV-1 Test = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test; CAP/CTM HIV-1 Test, v2.0 = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0.

Note: HIV-negative and HIV-1-positive subjects contributing both valid CAP/CTM HIV-1 Test, v2.0 and CAP/CTM HIV-1 Test results were included in this summary table.

A total of 417 paired HIV-1-positive samples had results within the linear range of both assays and were evaluable for the method comparison analysis. Table 19 shows the mean paired difference and 95% CI for the bias between the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test. The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 returns higher titers than the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, except at both the higher range ($> 5 \log_{10}$ cp/mL) and the lower range ($< 2 \log_{10}$ cp/mL) where it returns titers that are lower (see Table 19). The overall systematic bias is estimated as 0.2591 \log_{10} cp/mL.

Table 19
Mean Paired Difference and 95% CI for the Bias Between the COBAS® AmpliPrep/COBAS®
TaqMan® HIV-1 Test, v2.0 and the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test

Number of Paired HIV-1-Positive Samples Within Linear Range of Both Assays = 417		
Mean Difference (\log_{10} cp/mL)	Standard Error	95% CI
0.2591	0.0122	(0.235, 0.283)

CI = confidence interval; CAP/CTM HIV-1 Test = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test; CAP/CTM HIV-1 Test, v2.0 = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0.

Note: HIV-1-positive subjects contributing both valid CAP/CTM HIV-1 Test and CAP/CTM HIV-1 Test, v2.0 results within the linear range of each assay were included in this summary table.

The results of the Deming regression analysis between COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test results for paired HIV-1-positive specimens within the linear range of both assays are tabulated in Table 20 and displayed graphically in Figure 26 (in this figure, the dashed line indicates perfect agreement between the two test methods, i.e., $y = x$).


Table 20
Parameter Estimates from Deming Regression Analysis Between the COBAS®
AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and the COBAS® AmpliPrep/COBAS®
TaqMan® HIV-1 Test

Number of Paired HIV-1-Positive Samples Within Linear Range of Both Assays = 417				
Parameter	Parameter Estimate log ₁₀ cp/mL	Standard Error	95% CI	r ²
Intercept	0.3871	0.0488	(0.291, 0.483)	0.9375
Slope	0.9669	0.0122	(0.943, 0.991)	

CI = confidence interval; CAP/CTM HIV-1 Test = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test; CAP/CTM HIV-1 Test, v2.0 = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0.

Note: HIV-1-positive subjects contributing both valid CAP/CTM HIV-1 Test and CAP/CTM HIV-1 Test, v2.0 results within the linear range of each assay were included in this summary table.

Figure 26
Deming Regression Analysis Between the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1
Test, v2.0 and the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test



COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 versus COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5

Table 21 shows the comparison of COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 and CA HIV-1 MONITOR Test, v1.5 results for 991 subjects eligible for the analysis. The positive percent agreement of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 with respect to the COBAS® AMPLICOR® (CA) HIV-1 MONITOR Test, v1.5 was 100% (419/419; 95% CI = 99.1% to 100%). The negative percent agreement of the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 with respect to the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5 was 97.4% (557/572; 95% CI = 95.7% to 98.5%). Of the 15 subjects with positive COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test v2.0 results and negative COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5 results, 4 were false positive COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test v2.0 results from HIV-negative subjects identified in the clinical specificity analysis that again were below the LLoQ. Eleven were from HIV-1-positive subjects with COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test v2.0 results ranging from below the LLoQ to 223 cp/mL and negative COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5 results.

Table 21
Comparison of the CAP/CTM HIV-1 Test, v2.0 with the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5

CAP/CTM HIV-1 Test, v2.0	CA HIV-1 MONITOR Test, v1.5		Total
	Positive	Negative	
Positive	419	15	434
Negative	0	557	557
Total	419	572	991
Positive Percent Agreement (95% exact CI)	100.0% (99.1%, 100.0%)		
Negative Percent Agreement (95% exact CI)	97.4% (95.7%, 98.5%)		

CI = confidence interval; CAP/CTM HIV-1 Test = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test; CAP/CTM HIV-1 Test, v2.0 = COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0.

Note: HIV-negative and HIV-1-positive subjects contributing both valid CAP/CTM HIV-1 Test, v2.0 and CA HIV-1 MONITOR Test, v1.5 results were included in this summary table.

Conclusion

EDTA plasma

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 exhibits high levels of agreement with the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test in quantitative analyses ($r^2 = 0.9375$) and in concordance analyses (positive percent agreement = 99.5%; negative percent agreement = 98.1%). It quantifies clinical specimens 0.2591 log₁₀ cp/mL higher overall than the COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, with lower quantitation at the higher range (> 5 log₁₀ cp/mL) and the lower range (< 2 log₁₀ cp/mL).

The COBAS® AmpliPrep/COBAS® TaqMan® HIV-1 Test, v2.0 also shows high levels of agreement with the COBAS® AMPLICOR® HIV-1 MONITOR Test, v1.5 in concordance analyses (positive percent agreement = 100.0%; negative percent agreement = 97.4%).

These test results support the utility of the test for the intended use of assessing disease progression and monitoring antiretroviral therapy in HIV-1 infected patients.

Key Performance Characteristics for PSC Dried Plasma Spot Samples

Performance of PSC Dried Plasma Spot Samples Compared to EDTA Plasma Samples

The performance of **PSC** dried plasma spot samples were compared to centrifuged EDTA plasma samples by testing 325 samples from patients infected with HIV-1 and 2 HIV negative whole blood spiked with cell culture supernatants. Eighty-five specimens that had a measurable titer on both plasma and **PSC** were analyzed. The specimens comprised of HIV-1 M samples (FP= prospective collection by venipuncture and from capillary blood, VL= routine viral load testing, CD=CD4+ cell count leftover samples, SP = spiked samples) and were tested one replicate each (**PSC** and plasma) at an external site. Titers below the quantitation range were excluded from analysis. The Deming regression was performed considering log-transformed titers.

The Deming regression results are shown in Figure 27. The symbols * and • in Figure 27 show single determinations.

Figure 27
Deming Regression




Table 22
Summary of Results for Matrix Equivalency

Matrix (Plasma Type) Equivalency	Number of specimens with valid titer	Bland Altman Analysis		Deming Regression Analysis		
		Mean Log ₁₀ difference	95% CI [lower/upper]	Slope	Intercept	R- squared
PSC vs liquid plasma	85	0.05	[-0.01; 0.11]	1.06	-0.23	0.92

REFERENCES

1. Barre-Sinoussi, F., Chermann, J.C., Rey, F., Nugeyre, M.T., Chamaret, S., Gruest, J., Dauguet, C., Axler-Blin, C., Vezinet-Brun, F., Rouzioux, C., Rozenbaum, W., Montagnier, L. 1983. Isolation of a T-lymphotropic retrovirus from a patient at risk for acquired immune deficiency syndrome (AIDS). *Science* **220**:868-871.
2. Popovic, M., Sarngadharan, M.G., Read, E., Gallo, R.C. 1984. Detection, isolation, and continuous production of cytopathic retroviruses (HTLV-III) from patients with AIDS and pre-AIDS. *Science* **224**:497-500.
3. Gallo, R.C., Salahuddin, S.Z., Popovic, M., Shearer, G.M., Kaplan, M., Haynes, B.F., Palker, T.J., Redfield, R., Oleske, J., Safai, B., White, G., Foster, P., Markham, P.D. 1984. Frequent detection and isolation of cytopathic retroviruses (HTLV-III) from patients with AIDS and at risk for AIDS. *Science* **224**:500-503.
4. Curran, J.W., Jaffe, H.W., Hardy, A.M., Morgan, W.M., Selik, R.M., Dondero, T.J. 1988. Epidemiology of HIV Infection and AIDS in the United States. *Science* **239**:610-616.
5. Gaines, H., von Sydow, M.A., von Stedingk, L.V. 1990. Immunological changes in primary HIV-1 infection. *AIDS* **4**:995-999.
6. Tindall, B., and Cooper, D.A. 1991. Primary HIV-1 infection: host responses and intervention strategies. *AIDS* **5**:1-14.
7. Daar, E.S., Moudgil, T., Meyer, R.D., Ho, D.D. 1991. Transient high levels of viremia in patients with primary human immunodeficiency virus type 1 infection. *New England Journal of Medicine* **324**:961-964.
8. Clark, S.J., Saag, M.S., Decker, W.D. 1991. High titers of cytopathic virus in plasma of patients with symptomatic primary HIV-1 infection. *New England Journal of Medicine* **324**:954-960.
9. Albert J., Abrahamsson B., Nagy K., Aurelius E., Gaines H., Nystrom G., Fenyó E.M. 1990. Rapid development of isolate-specific neutralizing antibodies after primary HIV-1 infection and consequent emergence of virus variants which resist neutralization by autologous sera. *AIDS* **4**:107-112.
10. Horsburgh, C.R. Jr., Ou, C.Y., Jason, J., Holmberg, S.D., Longini, I.M. Jr., Schable, C., Mayer, K.H., Lifson, A.R., Schochetman, G., Ward, J.W., et al. 1989. Duration of human immunodeficiency virus infection before detection of antibody. *Lancet* **16**:637-640.
11. Schnittman, S.M., Psallidopoulos, M.C., Lane, H.C., Thompson, L., Baseler, M., Massari, F., Fox, C.H., Salzman, N.P., Fauci, A.S. 1989. The reservoir for HIV-1 in human peripheral blood is a T cell that maintains expression of CD4. *Science* **245**:305-308. Erratum in: *Science* 1989 245, preceding 694.
12. Schnittman, S.M., Greenhouse, J.J., Psallidopoulos, M.C., Baseler, M., Salzman, N.P., Fauci, A.S., Lane, H.C. 1990. Increasing viral burden in CD4⁺ T cells from patients with human immunodeficiency virus (HIV) infection reflects rapidly progressive immunosuppression and clinical disease. *Annals of Internal Medicine* **113**:438-443.
13. Pantaleo, G., Graziosi, C., Fauci, A.S. 1993. New concepts in the immunopathogenesis of human immunodeficiency virus (HIV) infection. *New England Journal of Medicine* **328**:327-335.
14. Piatak, M. Jr., Saag, M.S., Yang, L.C., Clark, S.J., Kappes, J.C., Luk, K.C., Hahn, B.H., Shaw, G.M., Lifson, J.D. 1993. High levels of HIV-1 in plasma during all stages of infection determined by competitive PCR. *Science* **259**:1749-1754.
15. Fauci, A.S., Schnittman, S.M., Poli, G., Koenig, S., Pantaleo, G. 1991. NIH conference: immunopathogenic mechanisms in human immunodeficiency virus (HIV) infection. *Annals of Internal Medicine* **114**:678-693.
16. Coffin, J.M. 1995. HIV-1 population dynamics in vivo: Implications for genetic variation, pathogenesis, and therapy. *Science* **267**:483-489.
17. Ho, D.D., Neumann, A.U., Perelson, A.S., Chen, W., Leonard, J.M., Markowitz, M. 1995. Rapid turnover of plasma virions and CD4 lymphocytes in HIV-1 infection. *Nature* **373**:123-126.

18. Wei, X., Ghosh, S.K., Taylor, M.E., Johnson, V.A., Emini, E.A., Deutsch, P., Lifson, J.D., Bonhoeffer, S., Nowak, M.A., Hahn, B.H., et al. 1995. Viral dynamics in human immunodeficiency virus type 1 infection. *Nature* **373**:117-122.
19. O'Brien, W.A., Hartigan, P.M., Martin, D., Esinhart, J., Hill, A., Benoit, S., Rubin, M., Simberkoff, M.S., Hamilton, J.D. 1996. Changes in plasma HIV-1 RNA and CD4 lymphocyte counts and the risk of progression to AIDS. Veterans Affairs Cooperative Study Group on AIDS. *New England Journal of Medicine* **334**:426-431.
20. Welles, S.L., Jackson, J.B., Yen-Lieberman, B., Demeter, L., Japour, A.J., Smeaton, L.M., Johnson, V.A., Kuritzkes, D.R., D'Aquila, R.T., Reichelderfer, P.A., Richman, D.D., Reichman, R., Fischl, M., Dolin, R., Coombs, R.W., Kahn, J.O., McLaren, C., Todd, J., Kwok, S., Crumpacker, C.S. 1996. Prognostic value of plasma Human Immunodeficiency Virus Type I (HIV-1) RNA levels in patients with advanced HIV-1 disease and with little or no zidovudine therapy. AIDS Clinical Trials Group Protocol 116A/116B/117 Team. *Journal of Infectious Diseases* **174**:696-703.
21. Coombs, R.W., Welles, S.L., Hooper, C., Reichelderfer, P.S., D'Aquila, R.T., Japour, A.J., Johnson, V.A., Kuritzkes, D.R., Richman, D.D., Kwok, S., Todd, J., Jackson, J.B., DeGruttola, V., Crumpacker, C.S., Kahn, J. 1996. Association of plasma Human Immunodeficiency Virus Type I RNA level with risk of clinical progression in patients with advanced infection. AIDS Clinical Trials Group (ACTG) 116B/117 Study Team. ACTG Virology Committee Resistance and HIV-1 RNA Working Groups. *Journal of Infectious Diseases* **174**:704-712.
22. Hammer, S., Crumpacker, C., D'Aquila, R., Jackson, B., Lathey, J., Livnat, D., Reichelderfer, P. 1993. Use of virologic assays for detection of human immunodeficiency virus in clinical trials: Recommendations of the AIDS Clinical Trials Group Virology Committee. *Journal of Clinical Microbiology* **31**:2557-2564.
23. Schochetman, G., George, J.R., ed. AIDS testing: a comprehensive guide to technical, medical, social, legal and management issues. 2nd ed. New York: Springer-Verlag, 1994.
24. Mulder, J., McKinney, N., Christopherson, C., Sninsky, J., Greenfield, L., Kwok, S. 1994. Rapid and simple PCR assay for quantitation of human immunodeficiency virus type 1 RNA in plasma: Application to acute retroviral infection. *Journal of Clinical Microbiology* **32**:292-300.
25. Dewar, R.L., Highbarger, H.C., Sarmiento, M.D., Todd, J.A., Vasudevachari, M.B., Davey, R.T. Jr., Kovacs, J.A., Salzman, N.P., Lane, H.C., Urdea, M.S. 1994. Application of branched DNA signal amplification to monitor human immunodeficiency virus type 1 burden in human plasma. *Journal of Infectious Diseases* **170**:1172-1179.
26. van Gemen, B., Kievits, T., Schukkink, R., van Strijp, D., Malek, L.T., Sooknanan, R., Huisman, H.G., Lens, P. 1993. Quantification of HIV-1 RNA in plasma using NASBA during HIV-1 primary infection. *Journal of Virological Methods* **43**:177-187.
27. Consolidated guidelines on the use of antiretroviral drugs for treating and preventing HIV infection – Recommendations for a public health approach. World Health Organization HIV/AIDS Programme, Geneva. 2016.
28. Saiki, R.K., Scharf, S., Faloona, F., Mullis, K.B., Horn, G.T., Erlich, H.A., Arnheim, N. 1985. Enzymatic amplification of b-globin genomic sequences and restriction site analysis for diagnosis of Sickle Cell Anemia. *Science* **230**:1350-1354.
29. Saiki, R.K., Gelfand, D.H., Stoffel, S., Scharf, S.J., Higuchi, R., Horn, G.T., Mullis, K.B., Erlich, H.A. 1988. Primer-directed enzymatic amplification of DNA with a thermostable DNA polymerase. *Science* **239**:487-491.
30. Mullis, K.B., Faloona, F.A. 1987. Specific synthesis of DNA in vitro via a polymerase-catalyzed chain reaction. *Methods in Enzymology* **155**:335-350.
31. Q. Meng, C. Wong, A. Rangachari, S. Tamatsukuri, M. Sasaki, E. Fiss, L. Cheng, T. Ramankutty, D. Clarke, H. Yawata, Y. Sakakura, T. Hirose, and C. Impraim. 2001. Automated Multiplex Assay System for Simultaneous Detection of Hepatitis B Virus DNA, Hepatitis C Virus RNA, and Human Immunodeficiency Virus Type 1 RNA. *Journal of Clinical Microbiology* **39** (8):2937-2945.

31. Smith, E.S., Li, A.K., Wang, A.M., Gelfand, D.H., Myers, T.M. 2003. Amplification of RNA: High-Temperature Reverse Transcription and DNA Amplification with a Magnesium-Activated Thermostable DNA Polymerase. In *PCR Primer: A Laboratory Manual*, 2nd Edition, Dieffenbach C.W. and Dveksler G.S., Eds. Cold Spring Harbor Laboratory Press, Cold Spring Harbor, New York, pp. 211-219.
33. Kwok, S., Sninsky, J.J. 1993. PCR detection of human immunodeficiency virus type 1 proviral DNA sequences. In: *Diagnostic Molecular Biology: Principles and Applications*. eds. Persing, D.H., Smith, T.F., Smith, F.C., et al. ASM, Washington, D.C.
34. Longo, M.C., Berninger, M.S., Hartley, J.L. 1990. Use of uracil DNA glycosylase to control carry-over contamination in polymerase chain reactions. *Gene* **93**:125-128.
35. Higuchi, R., Dollinger, G., Walsh, P.S., Griffith, R. 1992. Simultaneous amplification and detection of specific DNA sequences. *Biotechnology (N Y)*. **10**:413-417.
36. Heid, C.A., Stevens, J., Livak, J.K., Williams, P.M. 1996. Real time quantitative PCR. *Genome Research* **6**:986-994.
37. Holmes, H., Davis, C., Heath, A., Hewlett, I. and Lelie, N. 2001. An international collaborative study to establish the 1st international standard for HIV-1 RNA for use in nucleic acid-based techniques. *Journal of Virological Methods* **92**:141-150.
38. Richmond, J.Y. and McKinney, R.W. eds. 1999. *Biosafety in Microbiological and Biomedical Laboratories*. HHS Publication Number (CDC) 93-8395.
39. Clinical and Laboratory Standards Institute (CLSI). *Protection of Laboratory Workers from Occupationally Acquired Infections*. Approved Guideline-Third Edition. CLSI Document M29-A3 Wayne, PA:CLSI, 2005.
40. International Air Transport Association. *Dangerous Goods Regulations*, 41st Edition. 2000. 704 pp.
41. Robertson DL, Anderson JP, Bradac JA, Carr JK, Foley B, Funkhouser RK, Gao F, Hahn BH, Kalish ML, Kuiken C, Learn GH, Leitner T, McCutchan F, Osmanov S, Peeters M, Pieniasek D, Salminen M, Sharp PM, Wolinsky S, and Korber B: HIV-1 Nomenclature proposal: A reference guide to HIV-1 classification. In: *Human Retroviruses and AIDS 1999: A Compilation and Analysis of Nucleic Acid and Amino Acid Sequences* (Kuiken C, Foley B, Hahn B, Korber B, McCutchan F, Marx PA, Mellors JW, Mullins JI, Sodroski J, and Wolinsky S, eds.). Theoretical Biology and Biophysics Group, Los Alamos National Laboratory, Los Alamos, New Mexico, 1999, pp. 492-505.
42. Davis, C., Berry, N., Heath, A. and Holmes, H. 2008. An international collaborative study to establish a replacement World Health Organization International Standard for human immunodeficiency virus 1 RNA nucleic acid assays. *Vox Sanguinis* **95**: 218-225.

Document Revision Information	
Doc Rev. 17.0 02/2019	Updated the Notified Body Number under the CE mark. Please contact your local Roche Representative if you have any questions.
Doc Rev. 18.0 09/2019	Updated STORAGE AND HANDLING REQUIREMENTS section to instruct the user to visually inspect the product for signs of leakage before use and to not use the product if there is any evidence of leakage. Updated the harmonized symbol page. Please contact your local Roche Representative if you have any questions.



Roche Molecular Systems, Inc.
1080 US Highway 202 South
Branchburg, NJ 08876 USA
www.roche.com



Distributed by

Roche Diagnostics (Schweiz) AG
Industriestrasse 7
6343 Rotkreuz, Switzerland

Roche Diagnostics GmbH
Sandhofer Strasse 116
68305 Mannheim, Germany

Roche Diagnostics, SL
Avda. Generalitat, 171-173
E-08174 Sant Cugat del Vallès
Barcelona, Spain

Roche Diagnostica Brasil Ltda.
Av. Engenheiro Billings, 1729
Jaguará, Building 10
05321-010 São Paulo, SP Brazil

Roche Diagnostics
201, boulevard Armand-Frappier
H7V 4A2 Laval, Québec, Canada
(For Technical Assistance call:
Pour toute assistance technique,
appeler le: 1-877-273-3433)

Roche Diagnostics
2, Avenue du Vercors
38240 Meylan, France

Distributore in Italia:
Roche Diagnostics S.p.A.
Viale G. B. Stucchi 110
20052 Monza, Milano, Italy

Distribuidor em Portugal:
Roche Sistemas de Diagnósticos Lda.
Estrada Nacional, 249-1
2720-413 Amadora, Portugal

Trademarks and Patents

See <http://www.roche-diagnostics.us/patents>

©2019 Roche Molecular Systems, Inc.

09/2019
Doc Rev. 18.0

05328268001-18



Roche Diagnostics GmbH
Sandhofer Str. 116
68305 Mannheim
Germany



The following symbols are now used in labeling for Roche PCR diagnostic products.



Ancillary Software



In Vitro diagnostic medical device



Authorized representative
in the European community



Lower Limit of Assigned Range



Barcode Data Sheet



Manufacturer



Batch code



Store in the dark



Biological risks



Contains sufficient for $<n>$ tests



Catalogue number



Temperature limit



Consult instructions for use



Test Definition File



Contents of kit



Upper Limit of Assigned Range



Distributed by



Use-by date



For IVD performance evaluation only



Global Trade Item Number

Rx Only

US Only: Federal law restricts this
device to sale by or on the order of a
physician.



Date of manufacture



This product fulfills the requirements of the European Directive 98/79 EC for
in vitro diagnostic medical devices.